



## ANALYTICAL ENVIRONMENTAL SERVICES, INC.

June 25, 2010

Jessica Vickers  
Tetra Tech EM Inc.  
1955 Evergreen Blvd.  
Duluth GA 30096

TEL: (678) 775-3104  
FAX: (678) 775-3138

RE: Powder Springs Plating ER

Dear Jessica Vickers:

Order No: 1006J67

Analytical Environmental Services, Inc. received 2 samples on 6/22/2010 6:34:00 PM for the analyses presented in following report.

No problems were encountered during the analyses. Additionally, all results for the associated Quality Control samples were within EPA and/or AES established limits. Any discrepancies associated with the analyses contained herein will be noted and submitted in the form of a project Case Narrative.

AES' certifications are as follows:

- NELAC/Florida Certification number E87582 for analysis of Environmental Water, soil/hazardous waste, and Drinking Water Microbiology, effective 07/01/09-06/30/10.
- AIHA Certification ID #100671 for Industrial Hygiene samples (Organics, Inorganics), Environmental Lead (Paint, Soil, Dust Wipes, Air), and Environmental Microbiology (Fungal) effective until 09/01/11.

These results relate only to the items tested. This report may only be reproduced in full.

If you have any questions regarding these test results, please feel free to call.

James Forrest  
Project Manager



**ANALYTICAL ENVIRONMENTAL SERVICES, INC**  
 3785 Presidential Parkway, Atlanta GA 30340-3704  
**AES** TEL: (770) 457-8177 / TOLL-FREE (800) 972-4889 / FAX: (770) 457-8188

**CHAIN OF CUSTODY**

Work Order: **1006367**

Date: **6/22/10** Page **1** of **1**

<b>COMPANY:</b> Tetra Tech EM, Inc 1955 Evergreen Blvd. Bldg 200, Suite 300 Duluth GA 30096 PHONE: 678-775-3104 FAX:		<b>ADDRESS:</b> 1955 Evergreen Blvd. Bldg 200, Suite 300 Duluth GA 30096 FAX:		<b>ANALYSIS REQUESTED</b> Visit our website <a href="http://www.aesatlanta.com">www.aesatlanta.com</a> to check on the status of your results, place bottle orders, etc.		No # of Containers		
<b>SAMPLED BY:</b> Linda Nyland		<b>SIGNATURE:</b> <i>Linda Nyland</i>		<b>PRESERVATION (See codes)</b>		<b>REMARKS</b>		
#	SAMPLE ID	SAMPLED		Grab	Composite	Matrix (See codes)	I	I
		DATE	TIME					
1	PSP-111A-CON	6/22/10	1615	X	So		X	X
2	PSP-111B-CON	6/22/10	1620	X	So		X	X
3	PSP-111A-CON	6/22/10	1620	X	So		X	X
4	PSP-111B-CON	6/22/10	1620	X	So		X	X
5								
6								
7								
8								
9								
10								
11								
12								
13								
14								

<b>PROJECT INFORMATION</b> PROJECT NAME: Powder Springs Plotting ER PROJECT #: 103DX9017000.P012.3002 SITE ADDRESS: Austell GA SEND REPORT TO: Jessica.Vickie@ktrtech.com INVOICE TO: (IF DIFFERENT FROM ABOVE) SHIPMENT METHOD: OUT / / VIA: IN <input checked="" type="radio"/> CLIENT FedEx UPS MAIL COURIER SPECIAL INSTRUCTIONS/COMMENTS:		<b>RECEIPT</b> Total # of Containers: 2 Turnaround Time Request: <input type="radio"/> Standard 5 Business Days <input type="radio"/> 2 Business Day Rush <input type="radio"/> Next Business Day Rush <input checked="" type="radio"/> Same Day Rush (auth req.) <input type="radio"/> Other	
STATE PROGRAM (if any): E-mail: <input checked="" type="radio"/> / N; Fax? Y / N DATA PACKAGE: I II III IV		QUOTE #: PO#:	

White Copy - Original; Yellow Copy - Client

**Analytical Environmental Services, Inc**
**Date:** 25-Jun-10

**Client:** Tetra Tech EM Inc.  
**Project Name:** Powder Springs Plating ER  
**Lab ID:** 1006J67-001

**Client Sample ID:** PSP-111A-CON  
**Collection Date:** 6/22/2010 4:15:00 PM  
**Matrix:** Soil

Analyses	Result	Qual	MDL	Reporting Limit	Units	BatchID	DF	Date Analyzed	Analyst
<b>TOTAL MERCURY SW7471B</b>									
Mercury	BRL		0.0388	0.125	mg/Kg-dry	131280	1	06/23/2010 13:20	MP
<b>Hexavalent Chromium in Soil SW7196</b>									
Chromium, Hexavalent	BRL		0.740	1.24	mg/Kg-dry	131329	1	06/23/2010 15:40	CG
<b>METALS, TOTAL SW6010C</b>									
Arsenic	1.93	J	0.0881	5.84	mg/Kg-dry	131271	1	06/23/2010 12:13	MW
Barium	254		0.0360	5.84	mg/Kg-dry	131271	1	06/23/2010 12:13	MW
Cadmium	BRL		0.0996	2.92	mg/Kg-dry	131271	1	06/23/2010 12:13	MW
Chromium	20.6		0.0951	2.92	mg/Kg-dry	131271	1	06/23/2010 12:13	MW
Lead	42.1		0.191	5.84	mg/Kg-dry	131271	1	06/23/2010 12:13	MW
Selenium	0.300	J	0.209	5.84	mg/Kg-dry	131271	1	06/23/2010 12:13	MW
Silver	BRL		0.0231	2.92	mg/Kg-dry	131271	1	06/23/2010 12:13	MW
<b>PERCENT MOISTURE D2216</b>									
Percent Moisture	20.9		0	0	wt%	R174608	1	06/22/2010 19:00	AS

**Qualifiers:** \* Value exceeds maximum contaminant level  
 BRL Not detected at MDL  
 H Holding times for preparation or analysis exceeded  
 N Analyte not NELAC certified  
 B Analyte detected in the associated method blank

E Estimated value above quantitation range  
 S Spike Recovery outside limits due to matrix  
 J Estimated value detected below Reporting Limit  
 > Greater than Result value  
 < Less than Result value

## Analytical Environmental Services, Inc

Date: 25-Jun-10

<b>Client:</b> Tetra Tech EM Inc.	<b>Client Sample ID:</b> PSP-111B-CON
<b>Project Name:</b> Powder Springs Plating ER	<b>Collection Date:</b> 6/22/2010 4:20:00 PM
<b>Lab ID:</b> 1006J67-002	<b>Matrix:</b> Soil

Analyses	Result	Qual	MDL	Reporting Limit	Units	BatchID	DF	Date Analyzed	Analyst
<b>TOTAL MERCURY SW7471B</b>									
				<b>(SW7471)</b>					
Mercury	BRL		0.0365	0.117	mg/Kg-dry	131280	1	06/23/2010 13:11	MP
<b>Hexavalent Chromium in Soil SW7196</b>									
				<b>(SW3060A)</b>					
Chromium, Hexavalent	BRL		0.703	1.18	mg/Kg-dry	131329	1	06/23/2010 15:40	CG
<b>METALS, TOTAL SW6010C</b>									
				<b>(SW3050B)</b>					
Arsenic	1.33	J	0.0822	5.45	mg/Kg-dry	131271	1	06/23/2010 11:52	MW
Barium	48.4		0.0336	5.45	mg/Kg-dry	131271	1	06/23/2010 11:52	MW
Cadmium	BRL		0.0928	2.72	mg/Kg-dry	131271	1	06/23/2010 11:52	MW
Chromium	38.0		0.0887	2.72	mg/Kg-dry	131271	1	06/23/2010 11:52	MW
Lead	50.2		0.178	5.45	mg/Kg-dry	131271	1	06/23/2010 11:52	MW
Selenium	BRL		0.195	5.45	mg/Kg-dry	131271	1	06/23/2010 11:52	MW
Silver	0.0567	J	0.0216	2.72	mg/Kg-dry	131271	1	06/23/2010 11:52	MW
<b>PERCENT MOISTURE D2216</b>									
Percent Moisture	15.0		0	0	wt%	R174608	1	06/22/2010 19:00	AS

**Qualifiers:**

- \* Value exceeds maximum contaminant level
- BRL Not detected at MDL
- H Holding times for preparation or analysis exceeded
- N Analyte not NELAC certified
- B Analyte detected in the associated method blank

- E Estimated value above quantitation range
- S Spike Recovery outside limits due to matrix
- J Estimated value detected below Reporting Limit
- > Greater than Result value
- < Less than Result value

**Analytical Environmental Services, Inc.**

**Sample/Cooler Receipt Checklist**

Client Telre Ind

Work Order Number 1006567

Checklist completed by mafe 6/22/10  
Signature Date

Carrier name: FedEx ☐ UPS ☐ Courier ☐ Client ☒ US Mail ☐ Other ☐

Shipping container/cooler in good condition? Yes ☒ No ☐ Not Present ☐

Custody seals intact on shipping container/cooler? Yes ☐ No ☐ Not Present ☒

Custody seals intact on sample bottles? Yes ☐ No ☐ Not Present ☒

Container/Temp Blank temperature in compliance? (4°C±2)\* Yes ☒ No ☐

Cooler #1 3.6 Cooler #2 ☐ Cooler #3 ☐ Cooler #4 ☐ Cooler#5 ☐ Cooler #6 ☐

Chain of custody present? Yes ☒ No ☐

Chain of custody signed when relinquished and received? Yes ☒ No ☐

Chain of custody agrees with sample labels? Yes ☒ No ☐

Samples in proper container/bottle? Yes ☒ No ☐

Sample containers intact? Yes ☒ No ☐

Sufficient sample volume for indicated test? Yes ☒ No ☐

All samples received within holding time? Yes ☒ No ☐

Was TAT marked on the COC? Yes ☒ No ☐

Proceed with Standard TAT as per project history? Yes ☐ No ☐ Not Applicable ☒

Water - VOA vials have zero headspace? No VOA vials submitted ☒ Yes ☐ No ☐

Water - pH acceptable upon receipt? Yes ☐ No ☐ Not Applicable ☒

Adjusted? ☐ Checked by ☐

Sample Condition: Good ☒ Other(Explain) ☐

(For diffusive samples or AIHA lead) Is a known blank included? Yes ☐ No ☒

**See Case Narrative for resolution of the Non-Conformance.**

\* Samples do not have to comply with the given range for certain parameters.

\\Quality Assurance\Checklists Procedures Sign-Off Templates\Checklists\Sample Receipt Checklists\Sample\_Cooler\_Receipt\_Checklist

**Client:** Tetra Tech EM Inc.  
**Project:** Powder Springs Plating ER  
**Lab Order:** 1006J67

**Dates Report**

Lab Sample ID	Client Sample ID	Collection Date	Matrix	Test Name	TCLP Date	Prep Date	Analysis Date
1006J67-001A	PSP-111A-CON	6/22/2010 4:15:00PM	Soil	TOTAL METALS BY ICP		06/23/2010	06/23/2010
1006J67-001A	PSP-111A-CON	6/22/2010 4:15:00PM	Soil	Hexavalent Chromium-Soils		06/23/2010	06/23/2010
1006J67-001A	PSP-111A-CON	6/22/2010 4:15:00PM	Soil	MERCURY		06/23/2010	06/23/2010
1006J67-001A	PSP-111A-CON	6/22/2010 4:15:00PM	Soil	PERCENT MOISTURE			06/22/2010
1006J67-002A	PSP-111B-CON	6/22/2010 4:20:00PM	Soil	TOTAL METALS BY ICP		06/23/2010	06/23/2010
1006J67-002A	PSP-111B-CON	6/22/2010 4:20:00PM	Soil	Hexavalent Chromium-Soils		06/23/2010	06/23/2010
1006J67-002A	PSP-111B-CON	6/22/2010 4:20:00PM	Soil	MERCURY		06/23/2010	06/23/2010
1006J67-002A	PSP-111B-CON	6/22/2010 4:20:00PM	Soil	PERCENT MOISTURE			06/22/2010

Client: Tetra Tech EM Inc.  
 Project Name: Powder Springs Plating ER  
 Workorder: 1006J67

## ANALYTICAL QC SUMMARY REPORT

BatchID: 131271

Sample ID: <b>MB-131271</b>	Client ID:					Units: <b>mg/Kg</b>	Prep Date: <b>06/23/2010</b>	Run No: <b>174571</b>			
SampleType: <b>MBLK</b>	TestCode: <b>METALS, TOTAL</b>	<b>SW6010C</b>				BatchID: <b>131271</b>	Analysis Date: <b>06/23/2010</b>	Seq No: <b>3630097</b>			
Analyte	Result	RPT Limit	SPK value	SPK Ref Val	%REC	Low Limit	High Limit	RPD Ref Val	%RPD	RPD Limit	Qual

Arsenic	0.08254	5.00	0	0	0	0	0	0	0	0	J
Barium	0.03585	5.00	0	0	0	0	0	0	0	0	J
Cadmium	BRL	2.50	0	0	0	0	0	0	0	0	
Chromium	BRL	2.50	0	0	0	0	0	0	0	0	
Lead	BRL	5.00	0	0	0	0	0	0	0	0	
Selenium	BRL	5.00	0	0	0	0	0	0	0	0	
Silver	BRL	2.50	0	0	0	0	0	0	0	0	

Sample ID: <b>LCS-131271</b>	Client ID:					Units: <b>mg/Kg</b>	Prep Date: <b>06/23/2010</b>	Run No: <b>174571</b>			
SampleType: <b>LCS</b>	TestCode: <b>METALS, TOTAL</b>	<b>SW6010C</b>				BatchID: <b>131271</b>	Analysis Date: <b>06/23/2010</b>	Seq No: <b>3630095</b>			
Analyte	Result	RPT Limit	SPK value	SPK Ref Val	%REC	Low Limit	High Limit	RPD Ref Val	%RPD	RPD Limit	Qual

Arsenic	50.09	5.00	50	0.08254	100	80	120	0	0	0	
Barium	50.65	5.00	50	0.03585	101	80	120	0	0	0	
Cadmium	50.10	2.50	50	0	100	80	120	0	0	0	
Chromium	51.85	2.50	50	0	104	80	120	0	0	0	
Lead	49.43	5.00	50	0	98.9	80	120	0	0	0	
Selenium	49.89	5.00	50	0	99.8	80	120	0	0	0	
Silver	5.005	2.50	5	0	100	80	120	0	0	0	

Sample ID: <b>1006J67-002AMS</b>	Client ID: <b>PSP-111B-CON</b>					Units: <b>mg/Kg-dry</b>	Prep Date: <b>06/23/2010</b>	Run No: <b>174571</b>			
SampleType: <b>MS</b>	TestCode: <b>METALS, TOTAL</b>	<b>SW6010C</b>					BatchID: <b>131271</b>	Analysis Date: <b>06/23/2010</b>	Seq No: <b>3630100</b>		
Analyte	Result	RPT Limit	SPK value	SPK Ref Val	%REC	Low Limit	High Limit	RPD Ref Val	%RPD	RPD Limit	Qual

Arsenic	51.41	5.45	54.47	1.334	91.9	75	125	0	0	0	
Barium	112.1	5.45	54.47	48.40	117	75	125	0	0	0	
Cadmium	53.59	2.72	54.47	0	98.4	75	125	0	0	0	
Chromium	97.50	2.72	54.47	38.04	109	75	125	0	0	0	

<b>Qualifiers:</b>	>	Greater than Result value	<	Less than Result value	B	Analyte detected in the associated method blank
	BRL	Below reporting limit	E	Estimated (value above quantitation range)	H	Holding times for preparation or analysis exceeded
	J	Estimated value detected below Reporting Limit	N	Analyte not NELAC certified	R	RPD outside limits due to matrix
	Rpt Lim	Reporting Limit	S	Spike Recovery outside limits due to matrix		

Client: Tetra Tech EM Inc.  
Project Name: Powder Springs Plating ER  
Workorder: 1006J67

ANALYTICAL QC SUMMARY REPORT

BatchID: 131271

Sample ID: <b>1006J67-002AMS</b>	Client ID: <b>PSP-111B-CON</b>	Units: <b>mg/Kg-dry</b>			Prep Date: <b>06/23/2010</b>	Run No: <b>174571</b>					
SampleType: <b>MS</b>	TestCode: <b>METALS, TOTAL SW6010C</b>	BatchID: <b>131271</b>			Analysis Date: <b>06/23/2010</b>	Seq No: <b>3630100</b>					
Analyte	Result	RPT Limit	SPK value	SPK Ref Val	%REC	Low Limit	High Limit	RPD Ref Val	%RPD	RPD Limit	Qual

Lead	96.92	5.45	54.47	50.15	85.9	75	125	0	0	0	
Selenium	46.16	5.45	54.47	0	84.7	75	125	0	0	0	
Silver	5.405	2.72	5.447	0.05669	98.2	75	125	0	0	0	

Sample ID: <b>1006J67-002AMSD</b>	Client ID: <b>PSP-111B-CON</b>	Units: <b>mg/Kg-dry</b>			Prep Date: <b>06/23/2010</b>	Run No: <b>174571</b>					
SampleType: <b>MSD</b>	TestCode: <b>METALS, TOTAL SW6010C</b>	BatchID: <b>131271</b>			Analysis Date: <b>06/23/2010</b>	Seq No: <b>3630101</b>					
Analyte	Result	RPT Limit	SPK value	SPK Ref Val	%REC	Low Limit	High Limit	RPD Ref Val	%RPD	RPD Limit	Qual

Arsenic	50.69	5.44	54.38	1.334	90.8	75	125	51.41	1.41	20	
Barium	101.1	5.44	54.38	48.40	96.9	75	125	112.1	10.3	20	
Cadmium	52.96	2.72	54.38	0	97.4	75	125	53.59	1.19	20	
Chromium	89.46	2.72	54.38	38.04	94.6	75	125	97.50	8.61	20	
Lead	111.6	5.44	54.38	50.15	113	75	125	96.92	14.1	20	
Selenium	46.56	5.44	54.38	0	85.6	75	125	46.16	0.858	20	
Silver	5.332	2.72	5.438	0.05669	97	75	125	5.405	1.36	20	

Qualifiers:	>	Greater than Result value	<	Less than Result value	B	Analyte detected in the associated method blank
	BRL	Below reporting limit	E	Estimated (value above quantitation range)	H	Holding times for preparation or analysis exceeded
	J	Estimated value detected below Reporting Limit	N	Analyte not NELAC certified	R	RPD outside limits due to matrix
	Rpt Lim	Reporting Limit	S	Spike Recovery outside limits due to matrix		



Client: Tetra Tech EM Inc.  
Project Name: Powder Springs Plating ER  
Workorder: 1006J67

ANALYTICAL QC SUMMARY REPORT

BatchID: 131280

Sample ID: <b>MB-131280</b>	Client ID:	Units: <b>mg/Kg</b>				Prep Date: <b>06/23/2010</b>	Run No: <b>174642</b>				
SampleType: <b>MBLK</b>	TestCode: <b>TOTAL MERCURY SW7471B</b>	BatchID: <b>131280</b>				Analysis Date: <b>06/23/2010</b>	Seq No: <b>3630910</b>				
Analyte	Result	RPT Limit	SPK value	SPK Ref Val	%REC	Low Limit	High Limit	RPD Ref Val	%RPD	RPD Limit	Qual

Mercury	BRL	0.100	0	0	0	0	0	0	0	0	
---------	-----	-------	---	---	---	---	---	---	---	---	--

Sample ID: <b>LCS-131280</b>	Client ID:	Units: <b>mg/Kg</b>				Prep Date: <b>06/23/2010</b>	Run No: <b>174642</b>				
SampleType: <b>LCS</b>	TestCode: <b>TOTAL MERCURY SW7471B</b>	BatchID: <b>131280</b>				Analysis Date: <b>06/23/2010</b>	Seq No: <b>3630911</b>				
Analyte	Result	RPT Limit	SPK value	SPK Ref Val	%REC	Low Limit	High Limit	RPD Ref Val	%RPD	RPD Limit	Qual

Mercury	0.4055	0.100	0.4	0	101	80	120	0	0	0	
---------	--------	-------	-----	---	-----	----	-----	---	---	---	--

Sample ID: <b>1006J67-002AMS</b>	Client ID: <b>PSP-111B-CON</b>	Units: <b>mg/Kg-dry</b>			Prep Date: <b>06/23/2010</b>	Run No: <b>174642</b>					
SampleType: <b>MS</b>	TestCode: <b>TOTAL MERCURY SW7471B</b>	BatchID: <b>131280</b>			Analysis Date: <b>06/23/2010</b>	Seq No: <b>3630913</b>					
Analyte	Result	RPT Limit	SPK value	SPK Ref Val	%REC	Low Limit	High Limit	RPD Ref Val	%RPD	RPD Limit	Qual

Mercury	0.5040	0.118	0.4708	0	107	70	130	0	0	0	
---------	--------	-------	--------	---	-----	----	-----	---	---	---	--

Sample ID: <b>1006J67-002AMSD</b>	Client ID: <b>PSP-111B-CON</b>	Units: <b>mg/Kg-dry</b>			Prep Date: <b>06/23/2010</b>	Run No: <b>174642</b>					
SampleType: <b>MSD</b>	TestCode: <b>TOTAL MERCURY SW7471B</b>	BatchID: <b>131280</b>			Analysis Date: <b>06/23/2010</b>	Seq No: <b>3630915</b>					
Analyte	Result	RPT Limit	SPK value	SPK Ref Val	%REC	Low Limit	High Limit	RPD Ref Val	%RPD	RPD Limit	Qual

Mercury	0.5023	0.118	0.4708	0	107	70	130	0.5040	0.348	30	
---------	--------	-------	--------	---	-----	----	-----	--------	-------	----	--

Qualifiers:	>	Greater than Result value	<	Less than Result value	B	Analyte detected in the associated method blank
	BRL	Below reporting limit	E	Estimated (value above quantitation range)	H	Holding times for preparation or analysis exceeded
	J	Estimated value detected below Reporting Limit	N	Analyte not NELAC certified	R	RPD outside limits due to matrix
	Rpt Lim	Reporting Limit	S	Spike Recovery outside limits due to matrix		

Client: Tetra Tech EM Inc.  
 Project Name: Powder Springs Plating ER  
 Workorder: 1006J67

## ANALYTICAL QC SUMMARY REPORT

BatchID: 131329

Sample ID: <b>MB-131329</b>	Client ID:					Units: <b>mg/Kg</b>	Prep Date: <b>06/23/2010</b>	Run No: <b>174638</b>			
SampleType: <b>MBLK</b>	TestCode: <b>Hexavalent Chromium in Soil</b>	<b>SW7196</b>	BatchID: <b>131329</b>				Analysis Date: <b>06/23/2010</b>	Seq No: <b>3630841</b>			
Analyte	Result	RPT Limit	SPK value	SPK Ref Val	%REC	Low Limit	High Limit	RPD Ref Val	%RPD	RPD Limit	Qual

Chromium, Hexavalent BRL 2.00 0 0 0 0 0 0 0 0 0

Sample ID: <b>LCS-131329</b>	Client ID:					Units: <b>mg/Kg</b>	Prep Date: <b>06/23/2010</b>	Run No: <b>174638</b>			
SampleType: <b>LCS</b>	TestCode: <b>Hexavalent Chromium in Soil</b>	<b>SW7196</b>	BatchID: <b>131329</b>				Analysis Date: <b>06/23/2010</b>	Seq No: <b>3630842</b>			
Analyte	Result	RPT Limit	SPK value	SPK Ref Val	%REC	Low Limit	High Limit	RPD Ref Val	%RPD	RPD Limit	Qual

Chromium, Hexavalent 96.44 2.00 100 0 96.4 80 120 0 0 0

Sample ID: <b>1006J67-001AMS</b>	Client ID: <b>PSP-111A-CON</b>	Units: <b>mg/Kg-dry</b>				Prep Date: <b>06/23/2010</b>	Run No: <b>174638</b>				
SampleType: <b>MS</b>	TestCode: <b>Hexavalent Chromium in Soil SW7196</b>	BatchID: <b>131329</b>				Analysis Date: <b>06/23/2010</b>	Seq No: <b>3630848</b>				
Analyte	Result	RPT Limit	SPK value	SPK Ref Val	%REC	Low Limit	High Limit	RPD Ref Val	%RPD	RPD Limit	Qual

Chromium, Hexavalent 115.0 2.53 126.4 0 91 75 125 0 0 0

Sample ID: <b>1006J67-001AMSD</b>	Client ID: <b>PSP-111A-CON</b>	Units: <b>mg/Kg-dry</b>				Prep Date: <b>06/23/2010</b>	Run No: <b>174638</b>				
SampleType: <b>MSD</b>	TestCode: <b>Hexavalent Chromium in Soil SW7196</b>	BatchID: <b>131329</b>				Analysis Date: <b>06/23/2010</b>	Seq No: <b>3630852</b>				
Analyte	Result	RPT Limit	SPK value	SPK Ref Val	%REC	Low Limit	High Limit	RPD Ref Val	%RPD	RPD Limit	Qual

Chromium, Hexavalent 114.3 2.53 126.4 0 90.4 75 125 115.0 0.662 20

<b>Qualifiers:</b>	> Greater than Result value	< Less than Result value	B Analyte detected in the associated method blank
	BRL Below reporting limit	E Estimated (value above quantitation range)	H Holding times for preparation or analysis exceeded
	J Estimated value detected below Reporting Limit	N Analyte not NELAC certified	R RPD outside limits due to matrix
	Rpt Lim Reporting Limit	S Spike Recovery outside limits due to matrix	