

August 04, 2005

Client: WESTON SOLUTIONS
20 N. Wacker Drive Suite 1210
Chicago, IL 60606

Attn: Heidi Gorrill

Work Order: WOH0145
Project Name: Watertown Tire Fire E. R.
Project Number: [none]
Site/Location ID: Yes
Date Received: 08/04/05

An executed copy of the chain of custody is also included as an addendum to this report.

If you have any questions relating to this analytical report, please contact your Laboratory Project Manager at 1-800-833-7036

SAMPLE IDENTIFICATION	LAB NUMBER	COLLECTION DATE AND TIME
WTF080305 Raw	WOH0145-01	08/03/05 17:00
WTF080405 OCF	WOH0145-02	08/04/05 08:30
WTF080405 GAC	WOH0145-03	08/04/05 08:40
WTF080405 PIX	WOH0145-04	08/04/05 08:45
Trip Blank	WOH0145-05	08/03/05

Samples were received into laboratory on ice.

Wisconsin Certification Number: 128053530, DATCP #266

Unless subcontracted, volatiles analyses (including VOC, PVOC, GRO, BTEX, and TPH gasoline) performed by TestAmerica Watertown at 1101 Industrial Drive, Units 9&10. All other analyses performed at the address shown in the heading of this report.

Approved By:



TestAmerica Analytical - Watertown
Brian DeJong For Dan F. Milewsky
Project Manager

WESTON SOLUTIONS
20 N. Wacker Drive Suite 1210
Chicago, IL 60606
Heidi Gorrell

Work Order: WOH0145
Project: Watertown Tire Fire E. R.
Project Number: [none]

Received: 08/04/05
Reported: 08/04/05 15:11

ANALYTICAL REPORT

Analyte	Sample Result	Data Qualifiers	Units	MDL	MRL	Dilution Factor	Date Analyzed	Analyst	Seq/ Batch	Method
Sample ID: WOH0145-01 (WTF080305 Raw - Ground Water)						Sampled: 08/03/05 17:00				
VOCs by SW8260B										
Benzene	0.84		ug/L	0.20	0.67	1	08/04/05 14:24	LCG	5080134	SW 8260B
Bromobenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 14:24	LCG	5080134	SW 8260B
Bromochloromethane	<0.50		ug/L	0.50	1.7	1	08/04/05 14:24	LCG	5080134	SW 8260B
Bromodichloromethane	<0.20		ug/L	0.20	0.67	1	08/04/05 14:24	LCG	5080134	SW 8260B
Bromoform	<0.20		ug/L	0.20	0.67	1	08/04/05 14:24	LCG	5080134	SW 8260B
Bromomethane	<0.20		ug/L	0.20	0.67	1	08/04/05 14:24	LCG	5080134	SW 8260B
n-Butylbenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 14:24	LCG	5080134	SW 8260B
sec-Butylbenzene	<0.25		ug/L	0.25	0.83	1	08/04/05 14:24	LCG	5080134	SW 8260B
tert-Butylbenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 14:24	LCG	5080134	SW 8260B
Carbon Tetrachloride	<0.50		ug/L	0.50	1.7	1	08/04/05 14:24	LCG	5080134	SW 8260B
Chlorobenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 14:24	LCG	5080134	SW 8260B
Chlorodibromomethane	<0.20		ug/L	0.20	0.67	1	08/04/05 14:24	LCG	5080134	SW 8260B
Chloroethane	<1.0		ug/L	1.0	3.3	1	08/04/05 14:24	LCG	5080134	SW 8260B
Chloroform	<0.20		ug/L	0.20	0.67	1	08/04/05 14:24	LCG	5080134	SW 8260B
Chloromethane	<0.20		ug/L	0.20	0.67	1	08/04/05 14:24	LCG	5080134	SW 8260B
2-Chlorotoluene	<0.50		ug/L	0.50	1.7	1	08/04/05 14:24	LCG	5080134	SW 8260B
4-Chlorotoluene	<0.20		ug/L	0.20	0.67	1	08/04/05 14:24	LCG	5080134	SW 8260B
1,2-Dibromo-3-chloropropane	<0.50		ug/L	0.50	1.7	1	08/04/05 14:24	LCG	5080134	SW 8260B
1,2-Dibromoethane (EDB)	<0.20		ug/L	0.20	0.67	1	08/04/05 14:24	LCG	5080134	SW 8260B
Dibromomethane	<0.20		ug/L	0.20	0.67	1	08/04/05 14:24	LCG	5080134	SW 8260B
1,2-Dichlorobenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 14:24	LCG	5080134	SW 8260B
1,3-Dichlorobenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 14:24	LCG	5080134	SW 8260B
1,4-Dichlorobenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 14:24	LCG	5080134	SW 8260B
Dichlorodifluoromethane	<0.50		ug/L	0.50	1.7	1	08/04/05 14:24	LCG	5080134	SW 8260B
1,1-Dichloroethane	<0.50		ug/L	0.50	1.7	1	08/04/05 14:24	LCG	5080134	SW 8260B
1,2-Dichloroethane	<0.50		ug/L	0.50	1.7	1	08/04/05 14:24	LCG	5080134	SW 8260B
1,1-Dichloroethene	<0.50		ug/L	0.50	1.7	1	08/04/05 14:24	LCG	5080134	SW 8260B
cis-1,2-Dichloroethene	<0.50		ug/L	0.50	1.7	1	08/04/05 14:24	LCG	5080134	SW 8260B
trans-1,2-Dichloroethene	<0.50		ug/L	0.50	1.7	1	08/04/05 14:24	LCG	5080134	SW 8260B
1,2-Dichloropropane	<0.50		ug/L	0.50	1.7	1	08/04/05 14:24	LCG	5080134	SW 8260B
1,3-Dichloropropane	<0.25		ug/L	0.25	0.83	1	08/04/05 14:24	LCG	5080134	SW 8260B
2,2-Dichloropropane	<0.50		ug/L	0.50	1.7	1	08/04/05 14:24	LCG	5080134	SW 8260B
1,1-Dichloropropene	<0.50		ug/L	0.50	1.7	1	08/04/05 14:24	LCG	5080134	SW 8260B
cis-1,3-Dichloropropene	<0.20		ug/L	0.20	0.67	1	08/04/05 14:24	LCG	5080134	SW 8260B
trans-1,3-Dichloropropene	<0.20		ug/L	0.20	0.67	1	08/04/05 14:24	LCG	5080134	SW 8260B
Isopropyl Ether	<0.50		ug/L	0.50	1.7	1	08/04/05 14:24	LCG	5080134	SW 8260B
Ethylbenzene	1.0	J	ug/L	0.50	1.7	1	08/04/05 14:24	LCG	5080134	SW 8260B
Hexachlorobutadiene	<0.50		ug/L	0.50	1.7	1	08/04/05 14:24	LCG	5080134	SW 8260B
Isopropylbenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 14:24	LCG	5080134	SW 8260B
p-Isopropyltoluene	0.83		ug/L	0.20	0.67	1	08/04/05 14:24	LCG	5080134	SW 8260B
Methylene Chloride	<1.0		ug/L	1.0	3.3	1	08/04/05 14:24	LCG	5080134	SW 8260B
Methyl tert-Butyl Ether	<0.50		ug/L	0.50	1.7	1	08/04/05 14:24	LCG	5080134	SW 8260B
Naphthalene	0.92		ug/L	0.25	0.83	1	08/04/05 14:24	LCG	5080134	SW 8260B
n-Propylbenzene	<0.50		ug/L	0.50	1.7	1	08/04/05 14:24	LCG	5080134	SW 8260B
Styrene	<0.20		ug/L	0.20	0.67	1	08/04/05 14:24	LCG	5080134	SW 8260B
1,1,1,2-Tetrachloroethane	<0.25		ug/L	0.25	0.83	1	08/04/05 14:24	LCG	5080134	SW 8260B
1,1,2,2-Tetrachloroethane	<0.20		ug/L	0.20	0.67	1	08/04/05 14:24	LCG	5080134	SW 8260B
Tetrachloroethene	<0.50		ug/L	0.50	1.7	1	08/04/05 14:24	LCG	5080134	SW 8260B
Toluene	0.99		ug/L	0.20	0.67	1	08/04/05 14:24	LCG	5080134	SW 8260B

WESTON SOLUTIONS
20 N. Wacker Drive Suite 1210
Chicago, IL 60606
Heidi Gorrell

Work Order: WOH0145
Project: Watertown Tire Fire E. R.
Project Number: [none]

Received: 08/04/05
Reported: 08/04/05 15:11

Analyte	Sample Result	Data Qualifiers	Units	MDL	MRL	Dilution Factor	Date Analyzed	Analyst	Seq/ Batch	Method
Sample ID: WOH0145-01 (WTF080305 Raw - Ground Water) - cont.							Sampled: 08/03/05 17:00			
VOCs by SW8260B - cont.										
1,2,3-Trichlorobenzene	<0.25		ug/L	0.25	0.83	1	08/04/05 14:24	LCG	5080134	SW 8260B
1,2,4-Trichlorobenzene	<0.25		ug/L	0.25	0.83	1	08/04/05 14:24	LCG	5080134	SW 8260B
1,1,1-Trichloroethane	<0.50		ug/L	0.50	1.7	1	08/04/05 14:24	LCG	5080134	SW 8260B
1,1,2-Trichloroethane	<0.25		ug/L	0.25	0.83	1	08/04/05 14:24	LCG	5080134	SW 8260B
Trichloroethene	<0.20		ug/L	0.20	0.67	1	08/04/05 14:24	LCG	5080134	SW 8260B
Trichlorofluoromethane	<0.50		ug/L	0.50	1.7	1	08/04/05 14:24	LCG	5080134	SW 8260B
1,2,3-Trichloropropane	<0.50		ug/L	0.50	1.7	1	08/04/05 14:24	LCG	5080134	SW 8260B
1,2,4-Trimethylbenzene	0.74		ug/L	0.20	0.67	1	08/04/05 14:24	LCG	5080134	SW 8260B
1,3,5-Trimethylbenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 14:24	LCG	5080134	SW 8260B
Vinyl chloride	<0.20		ug/L	0.20	0.67	1	08/04/05 14:24	LCG	5080134	SW 8260B
Xylenes, Total	5.1		ug/L	0.50	1.7	1	08/04/05 14:24	LCG	5080134	SW 8260B
Surr: Dibromofluoromethane (89-119%)	99 %									
Surr: Toluene-d8 (91-109%)	100 %									
Surr: 4-Bromofluorobenzene (89-114%)	100 %									

Sample ID: WOH0145-02 (WTF080405 OCF - Ground Water)

Sampled: 08/04/05 08:30

VOCs by SW8260B

Benzene	1.1		ug/L	0.20	0.67	1	08/04/05 13:59	lcg	5080133	SW 8260B
Bromobenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:59	lcg	5080133	SW 8260B
Bromochloromethane	<0.50		ug/L	0.50	1.7	1	08/04/05 13:59	lcg	5080133	SW 8260B
Bromodichloromethane	<0.20		ug/L	0.20	0.67	1	08/04/05 13:59	lcg	5080133	SW 8260B
Bromoform	<0.20		ug/L	0.20	0.67	1	08/04/05 13:59	lcg	5080133	SW 8260B
Bromomethane	<0.20	C	ug/L	0.20	0.67	1	08/04/05 13:59	lcg	5080133	SW 8260B
n-Butylbenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:59	lcg	5080133	SW 8260B
sec-Butylbenzene	<0.25		ug/L	0.25	0.83	1	08/04/05 13:59	lcg	5080133	SW 8260B
tert-Butylbenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:59	lcg	5080133	SW 8260B
Carbon Tetrachloride	<0.50		ug/L	0.50	1.7	1	08/04/05 13:59	lcg	5080133	SW 8260B
Chlorobenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:59	lcg	5080133	SW 8260B
Chlorodibromomethane	<0.20		ug/L	0.20	0.67	1	08/04/05 13:59	lcg	5080133	SW 8260B
Chloroethane	<1.0		ug/L	1.0	3.3	1	08/04/05 13:59	lcg	5080133	SW 8260B
Chloroform	<0.20		ug/L	0.20	0.67	1	08/04/05 13:59	lcg	5080133	SW 8260B
Chloromethane	<0.20		ug/L	0.20	0.67	1	08/04/05 13:59	lcg	5080133	SW 8260B
2-Chlorotoluene	<0.50		ug/L	0.50	1.7	1	08/04/05 13:59	lcg	5080133	SW 8260B
4-Chlorotoluene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:59	lcg	5080133	SW 8260B
1,2-Dibromo-3-chloropropane	<0.50		ug/L	0.50	1.7	1	08/04/05 13:59	lcg	5080133	SW 8260B
1,2-Dibromoethane (EDB)	<0.20		ug/L	0.20	0.67	1	08/04/05 13:59	lcg	5080133	SW 8260B
Dibromomethane	<0.20		ug/L	0.20	0.67	1	08/04/05 13:59	lcg	5080133	SW 8260B
1,2-Dichlorobenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:59	lcg	5080133	SW 8260B
1,3-Dichlorobenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:59	lcg	5080133	SW 8260B
1,4-Dichlorobenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:59	lcg	5080133	SW 8260B
Dichlorodifluoromethane	<0.50		ug/L	0.50	1.7	1	08/04/05 13:59	lcg	5080133	SW 8260B
1,1-Dichloroethane	<0.50		ug/L	0.50	1.7	1	08/04/05 13:59	lcg	5080133	SW 8260B
1,2-Dichloroethane	7.9		ug/L	0.50	1.7	1	08/04/05 13:59	lcg	5080133	SW 8260B
1,1-Dichloroethene	<0.50		ug/L	0.50	1.7	1	08/04/05 13:59	lcg	5080133	SW 8260B
cis-1,2-Dichloroethene	<0.50		ug/L	0.50	1.7	1	08/04/05 13:59	lcg	5080133	SW 8260B
trans-1,2-Dichloroethene	<0.50		ug/L	0.50	1.7	1	08/04/05 13:59	lcg	5080133	SW 8260B
1,2-Dichloropropane	<0.50		ug/L	0.50	1.7	1	08/04/05 13:59	lcg	5080133	SW 8260B
1,3-Dichloropropane	<0.25		ug/L	0.25	0.83	1	08/04/05 13:59	lcg	5080133	SW 8260B
2,2-Dichloropropane	<0.50		ug/L	0.50	1.7	1	08/04/05 13:59	lcg	5080133	SW 8260B
1,1-Dichloropropene	<0.50		ug/L	0.50	1.7	1	08/04/05 13:59	lcg	5080133	SW 8260B
cis-1,3-Dichloropropene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:59	lcg	5080133	SW 8260B
trans-1,3-Dichloropropene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:59	lcg	5080133	SW 8260B

TestAmerica Analytical - Watertown

Brian DeJong For Dan F. Milewsky

Project Manager

WESTON SOLUTIONS
20 N. Wacker Drive Suite 1210
Chicago, IL 60606
Heidi Gorrell

Work Order: WOH0145
Project: Watertown Tire Fire E. R.
Project Number: [none]

Received: 08/04/05
Reported: 08/04/05 15:11

Analyte	Sample Result	Data Qualifiers	Units	MDL	MRL	Dilution Factor	Date Analyzed	Analyst	Seq/ Batch	Method
Sample ID: WOH0145-02 (WTF080405 OCF - Ground Water) - cont.							Sampled: 08/04/05 08:30			
VOCs by SW8260B - cont.										
Isopropyl Ether	<0.50		ug/L	0.50	1.7	1	08/04/05 13:59	lcg	5080133	SW 8260B
Ethylbenzene	1.0	J	ug/L	0.50	1.7	1	08/04/05 13:59	lcg	5080133	SW 8260B
Hexachlorobutadiene	<0.50		ug/L	0.50	1.7	1	08/04/05 13:59	lcg	5080133	SW 8260B
Isopropylbenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:59	lcg	5080133	SW 8260B
p-Isopropyltoluene	0.57	J	ug/L	0.20	0.67	1	08/04/05 13:59	lcg	5080133	SW 8260B
Methylene Chloride	<1.0		ug/L	1.0	3.3	1	08/04/05 13:59	lcg	5080133	SW 8260B
Methyl tert-Butyl Ether	<0.50		ug/L	0.50	1.7	1	08/04/05 13:59	lcg	5080133	SW 8260B
Naphthalene	0.81	J	ug/L	0.25	0.83	1	08/04/05 13:59	lcg	5080133	SW 8260B
n-Propylbenzene	<0.50		ug/L	0.50	1.7	1	08/04/05 13:59	lcg	5080133	SW 8260B
Styrene	0.21	J	ug/L	0.20	0.67	1	08/04/05 13:59	lcg	5080133	SW 8260B
1,1,1,2-Tetrachloroethane	<0.25		ug/L	0.25	0.83	1	08/04/05 13:59	lcg	5080133	SW 8260B
1,1,2,2-Tetrachloroethane	<0.20		ug/L	0.20	0.67	1	08/04/05 13:59	lcg	5080133	SW 8260B
Tetrachloroethene	<0.50		ug/L	0.50	1.7	1	08/04/05 13:59	lcg	5080133	SW 8260B
Toluene	1.5		ug/L	0.20	0.67	1	08/04/05 13:59	lcg	5080133	SW 8260B
1,2,3-Trichlorobenzene	<0.25		ug/L	0.25	0.83	1	08/04/05 13:59	lcg	5080133	SW 8260B
1,2,4-Trichlorobenzene	<0.25		ug/L	0.25	0.83	1	08/04/05 13:59	lcg	5080133	SW 8260B
1,1,1-Trichloroethane	<0.50		ug/L	0.50	1.7	1	08/04/05 13:59	lcg	5080133	SW 8260B
1,1,2-Trichloroethane	<0.25		ug/L	0.25	0.83	1	08/04/05 13:59	lcg	5080133	SW 8260B
Trichloroethene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:59	lcg	5080133	SW 8260B
Trichlorofluoromethane	<0.50		ug/L	0.50	1.7	1	08/04/05 13:59	lcg	5080133	SW 8260B
1,2,3-Trichloropropane	<0.50		ug/L	0.50	1.7	1	08/04/05 13:59	lcg	5080133	SW 8260B
1,2,4-Trimethylbenzene	0.58	J	ug/L	0.20	0.67	1	08/04/05 13:59	lcg	5080133	SW 8260B
1,3,5-Trimethylbenzene	0.30	J	ug/L	0.20	0.67	1	08/04/05 13:59	lcg	5080133	SW 8260B
Vinyl chloride	<0.20		ug/L	0.20	0.67	1	08/04/05 13:59	lcg	5080133	SW 8260B
Xylenes, Total	4.5		ug/L	0.50	1.7	1	08/04/05 13:59	lcg	5080133	SW 8260B
Surr: Dibromofluoromethane (89-119%)	102 %									
Surr: Toluene-d8 (91-109%)	96 %									
Surr: 4-Bromofluorobenzene (89-114%)	98 %									

WESTON SOLUTIONS
20 N. Wacker Drive Suite 1210
Chicago, IL 60606
Heidi Gorrell

Work Order: WOH0145
Project: Watertown Tire Fire E. R.
Project Number: [none]

Received: 08/04/05
Reported: 08/04/05 15:11

Analyte	Sample Result	Data Qualifiers	Units	MDL	MRL	Dilution Factor	Date Analyzed	Analyst	Seq/ Batch	Method
Sample ID: WOH0145-03 (WTF080405 GAC - Ground Water)							Sampled: 08/04/05 08:40			
VOCs by SW8260B										
Benzene	<0.20	C	ug/L	0.20	0.67	1	08/04/05 12:55	lcg	5080133	SW 8260B
Bromobenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 12:55	lcg	5080133	SW 8260B
Bromochloromethane	<0.50		ug/L	0.50	1.7	1	08/04/05 12:55	lcg	5080133	SW 8260B
Bromodichloromethane	<0.20		ug/L	0.20	0.67	1	08/04/05 12:55	lcg	5080133	SW 8260B
Bromoform	<0.20		ug/L	0.20	0.67	1	08/04/05 12:55	lcg	5080133	SW 8260B
Bromomethane	<0.20		ug/L	0.20	0.67	1	08/04/05 12:55	lcg	5080133	SW 8260B
n-Butylbenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 12:55	lcg	5080133	SW 8260B
sec-Butylbenzene	<0.25		ug/L	0.25	0.83	1	08/04/05 12:55	lcg	5080133	SW 8260B
tert-Butylbenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 12:55	lcg	5080133	SW 8260B
Carbon Tetrachloride	<0.50		ug/L	0.50	1.7	1	08/04/05 12:55	lcg	5080133	SW 8260B
Chlorobenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 12:55	lcg	5080133	SW 8260B
Chlorodibromomethane	<0.20		ug/L	0.20	0.67	1	08/04/05 12:55	lcg	5080133	SW 8260B
Chloroethane	<1.0		ug/L	1.0	3.3	1	08/04/05 12:55	lcg	5080133	SW 8260B
Chloroform	<0.20		ug/L	0.20	0.67	1	08/04/05 12:55	lcg	5080133	SW 8260B
Chloromethane	<0.20		ug/L	0.20	0.67	1	08/04/05 12:55	lcg	5080133	SW 8260B
2-Chlorotoluene	<0.50		ug/L	0.50	1.7	1	08/04/05 12:55	lcg	5080133	SW 8260B
4-Chlorotoluene	<0.20		ug/L	0.20	0.67	1	08/04/05 12:55	lcg	5080133	SW 8260B
1,2-Dibromo-3-chloropropane	<0.50		ug/L	0.50	1.7	1	08/04/05 12:55	lcg	5080133	SW 8260B
1,2-Dibromoethane (EDB)	<0.20		ug/L	0.20	0.67	1	08/04/05 12:55	lcg	5080133	SW 8260B
Dibromomethane	<0.20		ug/L	0.20	0.67	1	08/04/05 12:55	lcg	5080133	SW 8260B
1,2-Dichlorobenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 12:55	lcg	5080133	SW 8260B
1,3-Dichlorobenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 12:55	lcg	5080133	SW 8260B
1,4-Dichlorobenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 12:55	lcg	5080133	SW 8260B
Dichlorodifluoromethane	<0.50		ug/L	0.50	1.7	1	08/04/05 12:55	lcg	5080133	SW 8260B
1,1-Dichloroethane	<0.50		ug/L	0.50	1.7	1	08/04/05 12:55	lcg	5080133	SW 8260B
1,2-Dichloroethane	<0.50		ug/L	0.50	1.7	1	08/04/05 12:55	lcg	5080133	SW 8260B
1,1-Dichloroethene	<0.50		ug/L	0.50	1.7	1	08/04/05 12:55	lcg	5080133	SW 8260B
cis-1,2-Dichloroethene	<0.50		ug/L	0.50	1.7	1	08/04/05 12:55	lcg	5080133	SW 8260B
trans-1,2-Dichloroethene	<0.50		ug/L	0.50	1.7	1	08/04/05 12:55	lcg	5080133	SW 8260B
1,2-Dichloropropane	<0.50		ug/L	0.50	1.7	1	08/04/05 12:55	lcg	5080133	SW 8260B
1,3-Dichloropropane	<0.25		ug/L	0.25	0.83	1	08/04/05 12:55	lcg	5080133	SW 8260B
2,2-Dichloropropane	<0.50		ug/L	0.50	1.7	1	08/04/05 12:55	lcg	5080133	SW 8260B
1,1-Dichloropropene	<0.50		ug/L	0.50	1.7	1	08/04/05 12:55	lcg	5080133	SW 8260B
cis-1,3-Dichloropropene	<0.20		ug/L	0.20	0.67	1	08/04/05 12:55	lcg	5080133	SW 8260B
trans-1,3-Dichloropropene	<0.20		ug/L	0.20	0.67	1	08/04/05 12:55	lcg	5080133	SW 8260B
Isopropyl Ether	<0.50		ug/L	0.50	1.7	1	08/04/05 12:55	lcg	5080133	SW 8260B
Ethylbenzene	<0.50		ug/L	0.50	1.7	1	08/04/05 12:55	lcg	5080133	SW 8260B
Hexachlorobutadiene	<0.50		ug/L	0.50	1.7	1	08/04/05 12:55	lcg	5080133	SW 8260B
Isopropylbenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 12:55	lcg	5080133	SW 8260B
p-Isopropyltoluene	<0.20		ug/L	0.20	0.67	1	08/04/05 12:55	lcg	5080133	SW 8260B
Methylene Chloride	<1.0		ug/L	1.0	3.3	1	08/04/05 12:55	lcg	5080133	SW 8260B
Methyl tert-Butyl Ether	<0.50		ug/L	0.50	1.7	1	08/04/05 12:55	lcg	5080133	SW 8260B
Naphthalene	<0.25		ug/L	0.25	0.83	1	08/04/05 12:55	lcg	5080133	SW 8260B
n-Propylbenzene	<0.50		ug/L	0.50	1.7	1	08/04/05 12:55	lcg	5080133	SW 8260B
Styrene	<0.20		ug/L	0.20	0.67	1	08/04/05 12:55	lcg	5080133	SW 8260B
1,1,1,2-Tetrachloroethane	<0.25		ug/L	0.25	0.83	1	08/04/05 12:55	lcg	5080133	SW 8260B
1,1,2,2-Tetrachloroethane	<0.20		ug/L	0.20	0.67	1	08/04/05 12:55	lcg	5080133	SW 8260B
Tetrachloroethene	<0.50		ug/L	0.50	1.7	1	08/04/05 12:55	lcg	5080133	SW 8260B
Toluene	<0.20		ug/L	0.20	0.67	1	08/04/05 12:55	lcg	5080133	SW 8260B
1,2,3-Trichlorobenzene	<0.25		ug/L	0.25	0.83	1	08/04/05 12:55	lcg	5080133	SW 8260B
1,2,4-Trichlorobenzene	<0.25		ug/L	0.25	0.83	1	08/04/05 12:55	lcg	5080133	SW 8260B

WESTON SOLUTIONS
20 N. Wacker Drive Suite 1210
Chicago, IL 60606
Heidi Gorrell

Work Order: WOH0145
Project: Watertown Tire Fire E. R.
Project Number: [none]

Received: 08/04/05
Reported: 08/04/05 15:11

Analyte	Sample Result	Data Qualifiers	Units	MDL	MRL	Dilution Factor	Date Analyzed	Analyst	Seq/ Batch	Method
Sample ID: WOH0145-03 (WTF080405 GAC - Ground Water) - cont.							Sampled: 08/04/05 08:40			
VOCs by SW8260B - cont.										
1,1,1-Trichloroethane	<0.50		ug/L	0.50	1.7	1	08/04/05 12:55	lcg	5080133	SW 8260B
1,1,2-Trichloroethane	<0.25		ug/L	0.25	0.83	1	08/04/05 12:55	lcg	5080133	SW 8260B
Trichloroethene	<0.20		ug/L	0.20	0.67	1	08/04/05 12:55	lcg	5080133	SW 8260B
Trichlorofluoromethane	<0.50		ug/L	0.50	1.7	1	08/04/05 12:55	lcg	5080133	SW 8260B
1,2,3-Trichloropropane	<0.50		ug/L	0.50	1.7	1	08/04/05 12:55	lcg	5080133	SW 8260B
1,2,4-Trimethylbenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 12:55	lcg	5080133	SW 8260B
1,3,5-Trimethylbenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 12:55	lcg	5080133	SW 8260B
Vinyl chloride	<0.20		ug/L	0.20	0.67	1	08/04/05 12:55	lcg	5080133	SW 8260B
Xylenes, Total	<0.50		ug/L	0.50	1.7	1	08/04/05 12:55	lcg	5080133	SW 8260B
Surr: Dibromofluoromethane (89-119%)	102 %									
Surr: Toluene-d8 (91-109%)	96 %									
Surr: 4-Bromofluorobenzene (89-114%)	98 %									

Sample ID: WOH0145-04 (WTF080405 PIX - Ground Water)						Sampled: 08/04/05 08:45				
VOCs by SW8260B										
Benzene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:27	lcg	5080133	SW 8260B
Bromobenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:27	lcg	5080133	SW 8260B
Bromochloromethane	<0.50		ug/L	0.50	1.7	1	08/04/05 13:27	lcg	5080133	SW 8260B
Bromodichloromethane	<0.20		ug/L	0.20	0.67	1	08/04/05 13:27	lcg	5080133	SW 8260B
Bromoform	<0.20		ug/L	0.20	0.67	1	08/04/05 13:27	lcg	5080133	SW 8260B
Bromomethane	<0.20	C	ug/L	0.20	0.67	1	08/04/05 13:27	lcg	5080133	SW 8260B
n-Butylbenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:27	lcg	5080133	SW 8260B
sec-Butylbenzene	<0.25		ug/L	0.25	0.83	1	08/04/05 13:27	lcg	5080133	SW 8260B
tert-Butylbenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:27	lcg	5080133	SW 8260B
Carbon Tetrachloride	<0.50		ug/L	0.50	1.7	1	08/04/05 13:27	lcg	5080133	SW 8260B
Chlorobenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:27	lcg	5080133	SW 8260B
Chlorodibromomethane	<0.20		ug/L	0.20	0.67	1	08/04/05 13:27	lcg	5080133	SW 8260B
Chloroethane	<1.0		ug/L	1.0	3.3	1	08/04/05 13:27	lcg	5080133	SW 8260B
Chloroform	<0.20		ug/L	0.20	0.67	1	08/04/05 13:27	lcg	5080133	SW 8260B
Chloromethane	<0.20		ug/L	0.20	0.67	1	08/04/05 13:27	lcg	5080133	SW 8260B
2-Chlorotoluene	<0.50		ug/L	0.50	1.7	1	08/04/05 13:27	lcg	5080133	SW 8260B
4-Chlorotoluene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:27	lcg	5080133	SW 8260B
1,2-Dibromo-3-chloropropane	<0.50		ug/L	0.50	1.7	1	08/04/05 13:27	lcg	5080133	SW 8260B
1,2-Dibromoethane (EDB)	<0.20		ug/L	0.20	0.67	1	08/04/05 13:27	lcg	5080133	SW 8260B
Dibromomethane	<0.20		ug/L	0.20	0.67	1	08/04/05 13:27	lcg	5080133	SW 8260B
1,2-Dichlorobenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:27	lcg	5080133	SW 8260B
1,3-Dichlorobenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:27	lcg	5080133	SW 8260B
1,4-Dichlorobenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:27	lcg	5080133	SW 8260B
Dichlorodifluoromethane	<0.50		ug/L	0.50	1.7	1	08/04/05 13:27	lcg	5080133	SW 8260B
1,1-Dichloroethane	<0.50		ug/L	0.50	1.7	1	08/04/05 13:27	lcg	5080133	SW 8260B
1,2-Dichloroethane	27		ug/L	0.50	1.7	1	08/04/05 13:27	lcg	5080133	SW 8260B
1,1-Dichloroethene	<0.50		ug/L	0.50	1.7	1	08/04/05 13:27	lcg	5080133	SW 8260B
cis-1,2-Dichloroethene	<0.50		ug/L	0.50	1.7	1	08/04/05 13:27	lcg	5080133	SW 8260B
trans-1,2-Dichloroethene	<0.50		ug/L	0.50	1.7	1	08/04/05 13:27	lcg	5080133	SW 8260B
1,2-Dichloropropane	<0.50		ug/L	0.50	1.7	1	08/04/05 13:27	lcg	5080133	SW 8260B
1,3-Dichloropropane	<0.25		ug/L	0.25	0.83	1	08/04/05 13:27	lcg	5080133	SW 8260B
2,2-Dichloropropane	<0.50		ug/L	0.50	1.7	1	08/04/05 13:27	lcg	5080133	SW 8260B
1,1-Dichloropropene	<0.50		ug/L	0.50	1.7	1	08/04/05 13:27	lcg	5080133	SW 8260B
cis-1,3-Dichloropropene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:27	lcg	5080133	SW 8260B
trans-1,3-Dichloropropene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:27	lcg	5080133	SW 8260B
Isopropyl Ether	<0.50		ug/L	0.50	1.7	1	08/04/05 13:27	lcg	5080133	SW 8260B
Ethylbenzene	<0.50		ug/L	0.50	1.7	1	08/04/05 13:27	lcg	5080133	SW 8260B

WESTON SOLUTIONS
20 N. Wacker Drive Suite 1210
Chicago, IL 60606
Heidi Gorrell

Work Order: WOH0145
Project: Watertown Tire Fire E. R.
Project Number: [none]

Received: 08/04/05
Reported: 08/04/05 15:11

Analyte	Sample Result	Data Qualifiers	Units	MDL	MRL	Dilution Factor	Date Analyzed	Analyst	Seq/ Batch	Method
Sample ID: WOH0145-04 (WTF080405 PIX - Ground Water) - cont.							Sampled: 08/04/05 08:45			
VOCs by SW8260B - cont.										
Hexachlorobutadiene	<0.50		ug/L	0.50	1.7	1	08/04/05 13:27	lcg	5080133	SW 8260B
Isopropylbenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:27	lcg	5080133	SW 8260B
p-Isopropyltoluene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:27	lcg	5080133	SW 8260B
Methylene Chloride	<1.0		ug/L	1.0	3.3	1	08/04/05 13:27	lcg	5080133	SW 8260B
Methyl tert-Butyl Ether	<0.50		ug/L	0.50	1.7	1	08/04/05 13:27	lcg	5080133	SW 8260B
Naphthalene	<0.25		ug/L	0.25	0.83	1	08/04/05 13:27	lcg	5080133	SW 8260B
n-Propylbenzene	<0.50		ug/L	0.50	1.7	1	08/04/05 13:27	lcg	5080133	SW 8260B
Styrene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:27	lcg	5080133	SW 8260B
1,1,1,2-Tetrachloroethane	<0.25		ug/L	0.25	0.83	1	08/04/05 13:27	lcg	5080133	SW 8260B
1,1,2,2-Tetrachloroethane	<0.20		ug/L	0.20	0.67	1	08/04/05 13:27	lcg	5080133	SW 8260B
Tetrachloroethene	<0.50		ug/L	0.50	1.7	1	08/04/05 13:27	lcg	5080133	SW 8260B
Toluene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:27	lcg	5080133	SW 8260B
1,2,3-Trichlorobenzene	<0.25		ug/L	0.25	0.83	1	08/04/05 13:27	lcg	5080133	SW 8260B
1,2,4-Trichlorobenzene	<0.25		ug/L	0.25	0.83	1	08/04/05 13:27	lcg	5080133	SW 8260B
1,1,1-Trichloroethane	<0.50		ug/L	0.50	1.7	1	08/04/05 13:27	lcg	5080133	SW 8260B
1,1,2-Trichloroethane	<0.25		ug/L	0.25	0.83	1	08/04/05 13:27	lcg	5080133	SW 8260B
Trichloroethene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:27	lcg	5080133	SW 8260B
Trichlorofluoromethane	<0.50		ug/L	0.50	1.7	1	08/04/05 13:27	lcg	5080133	SW 8260B
1,2,3-Trichloropropane	<0.50		ug/L	0.50	1.7	1	08/04/05 13:27	lcg	5080133	SW 8260B
1,2,4-Trimethylbenzene	0.25	J	ug/L	0.20	0.67	1	08/04/05 13:27	lcg	5080133	SW 8260B
1,3,5-Trimethylbenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:27	lcg	5080133	SW 8260B
Vinyl chloride	<0.20		ug/L	0.20	0.67	1	08/04/05 13:27	lcg	5080133	SW 8260B
Xylenes, Total	<0.50		ug/L	0.50	1.7	1	08/04/05 13:27	lcg	5080133	SW 8260B
Surr: Dibromofluoromethane (89-119%)	102 %									
Surr: Toluene-d8 (91-109%)	96 %									
Surr: 4-Bromofluorobenzene (89-114%)	99 %									

Sample ID: WOH0145-05 (Trip Blank - Ground Water)

Sampled: 08/03/05

VOCs by SW8260B

Benzene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:54	LCG	5080134	SW 8260B
Bromobenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:54	LCG	5080134	SW 8260B
Bromochloromethane	<0.50		ug/L	0.50	1.7	1	08/04/05 13:54	LCG	5080134	SW 8260B
Bromodichloromethane	<0.20		ug/L	0.20	0.67	1	08/04/05 13:54	LCG	5080134	SW 8260B
Bromoform	<0.20		ug/L	0.20	0.67	1	08/04/05 13:54	LCG	5080134	SW 8260B
Bromomethane	<0.20		ug/L	0.20	0.67	1	08/04/05 13:54	LCG	5080134	SW 8260B
n-Butylbenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:54	LCG	5080134	SW 8260B
sec-Butylbenzene	<0.25		ug/L	0.25	0.83	1	08/04/05 13:54	LCG	5080134	SW 8260B
tert-Butylbenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:54	LCG	5080134	SW 8260B
Carbon Tetrachloride	<0.50		ug/L	0.50	1.7	1	08/04/05 13:54	LCG	5080134	SW 8260B
Chlorobenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:54	LCG	5080134	SW 8260B
Chlorodibromomethane	<0.20		ug/L	0.20	0.67	1	08/04/05 13:54	LCG	5080134	SW 8260B
Chloroethane	<1.0		ug/L	1.0	3.3	1	08/04/05 13:54	LCG	5080134	SW 8260B
Chloroform	<0.20		ug/L	0.20	0.67	1	08/04/05 13:54	LCG	5080134	SW 8260B
Chloromethane	<0.20		ug/L	0.20	0.67	1	08/04/05 13:54	LCG	5080134	SW 8260B
2-Chlorotoluene	<0.50		ug/L	0.50	1.7	1	08/04/05 13:54	LCG	5080134	SW 8260B
4-Chlorotoluene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:54	LCG	5080134	SW 8260B
1,2-Dibromo-3-chloropropane	<0.50		ug/L	0.50	1.7	1	08/04/05 13:54	LCG	5080134	SW 8260B
1,2-Dibromoethane (EDB)	<0.20		ug/L	0.20	0.67	1	08/04/05 13:54	LCG	5080134	SW 8260B
Dibromomethane	<0.20		ug/L	0.20	0.67	1	08/04/05 13:54	LCG	5080134	SW 8260B
1,2-Dichlorobenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:54	LCG	5080134	SW 8260B
1,3-Dichlorobenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:54	LCG	5080134	SW 8260B
1,4-Dichlorobenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:54	LCG	5080134	SW 8260B

TestAmerica Analytical - Watertown

Brian DeJong For Dan F. Milewsky

Project Manager

WESTON SOLUTIONS
20 N. Wacker Drive Suite 1210
Chicago, IL 60606
Heidi Gorrell

Work Order: WOH0145
Project: Watertown Tire Fire E. R.
Project Number: [none]

Received: 08/04/05
Reported: 08/04/05 15:11

Analyte	Sample Result	Data Qualifiers	Units	MDL	MRL	Dilution Factor	Date Analyzed	Analyst	Seq/ Batch	Method
Sample ID: WOH0145-05 (Trip Blank - Ground Water) - cont.							Sampled: 08/03/05			
VOCs by SW8260B - cont.										
Dichlorodifluoromethane	<0.50		ug/L	0.50	1.7	1	08/04/05 13:54	LCG	5080134	SW 8260B
1,1-Dichloroethane	<0.50		ug/L	0.50	1.7	1	08/04/05 13:54	LCG	5080134	SW 8260B
1,2-Dichloroethane	<0.50		ug/L	0.50	1.7	1	08/04/05 13:54	LCG	5080134	SW 8260B
1,1-Dichloroethene	<0.50		ug/L	0.50	1.7	1	08/04/05 13:54	LCG	5080134	SW 8260B
cis-1,2-Dichloroethene	<0.50		ug/L	0.50	1.7	1	08/04/05 13:54	LCG	5080134	SW 8260B
trans-1,2-Dichloroethene	<0.50		ug/L	0.50	1.7	1	08/04/05 13:54	LCG	5080134	SW 8260B
1,2-Dichloropropane	<0.50		ug/L	0.50	1.7	1	08/04/05 13:54	LCG	5080134	SW 8260B
1,3-Dichloropropane	<0.25		ug/L	0.25	0.83	1	08/04/05 13:54	LCG	5080134	SW 8260B
2,2-Dichloropropane	<0.50		ug/L	0.50	1.7	1	08/04/05 13:54	LCG	5080134	SW 8260B
1,1-Dichloropropene	<0.50		ug/L	0.50	1.7	1	08/04/05 13:54	LCG	5080134	SW 8260B
cis-1,3-Dichloropropene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:54	LCG	5080134	SW 8260B
trans-1,3-Dichloropropene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:54	LCG	5080134	SW 8260B
Isopropyl Ether	<0.50		ug/L	0.50	1.7	1	08/04/05 13:54	LCG	5080134	SW 8260B
Ethylbenzene	<0.50		ug/L	0.50	1.7	1	08/04/05 13:54	LCG	5080134	SW 8260B
Hexachlorobutadiene	<0.50		ug/L	0.50	1.7	1	08/04/05 13:54	LCG	5080134	SW 8260B
Isopropylbenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:54	LCG	5080134	SW 8260B
p-Isopropyltoluene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:54	LCG	5080134	SW 8260B
Methylene Chloride	<1.0		ug/L	1.0	3.3	1	08/04/05 13:54	LCG	5080134	SW 8260B
Methyl tert-Butyl Ether	<0.50		ug/L	0.50	1.7	1	08/04/05 13:54	LCG	5080134	SW 8260B
Naphthalene	<0.25		ug/L	0.25	0.83	1	08/04/05 13:54	LCG	5080134	SW 8260B
n-Propylbenzene	<0.50		ug/L	0.50	1.7	1	08/04/05 13:54	LCG	5080134	SW 8260B
Styrene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:54	LCG	5080134	SW 8260B
1,1,1,2-Tetrachloroethane	<0.25		ug/L	0.25	0.83	1	08/04/05 13:54	LCG	5080134	SW 8260B
1,1,2,2-Tetrachloroethane	<0.20		ug/L	0.20	0.67	1	08/04/05 13:54	LCG	5080134	SW 8260B
Tetrachloroethene	<0.50		ug/L	0.50	1.7	1	08/04/05 13:54	LCG	5080134	SW 8260B
Toluene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:54	LCG	5080134	SW 8260B
1,2,3-Trichlorobenzene	<0.25		ug/L	0.25	0.83	1	08/04/05 13:54	LCG	5080134	SW 8260B
1,2,4-Trichlorobenzene	<0.25		ug/L	0.25	0.83	1	08/04/05 13:54	LCG	5080134	SW 8260B
1,1,1-Trichloroethane	<0.50		ug/L	0.50	1.7	1	08/04/05 13:54	LCG	5080134	SW 8260B
1,1,2-Trichloroethane	<0.25		ug/L	0.25	0.83	1	08/04/05 13:54	LCG	5080134	SW 8260B
Trichloroethene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:54	LCG	5080134	SW 8260B
Trichlorofluoromethane	<0.50		ug/L	0.50	1.7	1	08/04/05 13:54	LCG	5080134	SW 8260B
1,2,3-Trichloropropane	<0.50		ug/L	0.50	1.7	1	08/04/05 13:54	LCG	5080134	SW 8260B
1,2,4-Trimethylbenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:54	LCG	5080134	SW 8260B
1,3,5-Trimethylbenzene	<0.20		ug/L	0.20	0.67	1	08/04/05 13:54	LCG	5080134	SW 8260B
Vinyl chloride	<0.20		ug/L	0.20	0.67	1	08/04/05 13:54	LCG	5080134	SW 8260B
Xylenes, Total	<0.50		ug/L	0.50	1.7	1	08/04/05 13:54	LCG	5080134	SW 8260B
Surr: Dibromofluoromethane (89-119%)	100 %									
Surr: Toluene-d8 (91-109%)	100 %									
Surr: 4-Bromofluorobenzene (89-114%)	101 %									

WESTON SOLUTIONS
20 N. Wacker Drive Suite 1210
Chicago, IL 60606
Heidi Gorriall

Work Order: WOH0145
Project: Watertown Tire Fire E. R.
Project Number: [none]

Received: 08/04/05
Reported: 08/04/05 15:11

LABORATORY BLANK QC DATA

Analyte	Seq/ Batch	Source Result	Spike Level	Units	MDL	MRL	Result	Dup Result	% REC	Dup %REC	% REC Limits	RPD RPD	RPD Limit	Q
VOCs by SW8260B														
Benzene	5080133			ug/L	0.20	0.67	<0.20							
Bromobenzene	5080133			ug/L	0.20	0.67	<0.20							
Bromochloromethane	5080133			ug/L	0.50	1.7	<0.50							
Bromodichloromethane	5080133			ug/L	0.20	0.67	<0.20							
Bromoform	5080133			ug/L	0.20	0.67	<0.20							
Bromomethane	5080133			ug/L	0.20	0.67	<0.20							C
n-Butylbenzene	5080133			ug/L	0.20	0.67	<0.20							
sec-Butylbenzene	5080133			ug/L	0.25	0.83	<0.25							
tert-Butylbenzene	5080133			ug/L	0.20	0.67	<0.20							
Carbon Tetrachloride	5080133			ug/L	0.50	1.7	<0.50							
Chlorobenzene	5080133			ug/L	0.20	0.67	<0.20							
Chlorodibromomethane	5080133			ug/L	0.20	0.67	<0.20							
Chloroethane	5080133			ug/L	1.0	3.3	<1.0							
Chloroform	5080133			ug/L	0.20	0.67	<0.20							
Chloromethane	5080133			ug/L	0.20	0.67	<0.20							
2-Chlorotoluene	5080133			ug/L	0.50	1.7	<0.50							
4-Chlorotoluene	5080133			ug/L	0.20	0.67	<0.20							
1,2-Dibromo-3-chloropropane	5080133			ug/L	0.50	1.7	<0.50							
1,2-Dibromoethane (EDB)	5080133			ug/L	0.20	0.67	<0.20							
Dibromomethane	5080133			ug/L	0.20	0.67	<0.20							
1,2-Dichlorobenzene	5080133			ug/L	0.20	0.67	<0.20							
1,3-Dichlorobenzene	5080133			ug/L	0.20	0.67	<0.20							
1,4-Dichlorobenzene	5080133			ug/L	0.20	0.67	<0.20							
Dichlorodifluoromethane	5080133			ug/L	0.50	1.7	<0.50							
1,1-Dichloroethane	5080133			ug/L	0.50	1.7	<0.50							
1,2-Dichloroethane	5080133			ug/L	0.50	1.7	<0.50							
1,1-Dichloroethene	5080133			ug/L	0.50	1.7	<0.50							
cis-1,2-Dichloroethene	5080133			ug/L	0.50	1.7	<0.50							
trans-1,2-Dichloroethene	5080133			ug/L	0.50	1.7	<0.50							
1,2-Dichloropropane	5080133			ug/L	0.50	1.7	<0.50							
1,3-Dichloropropane	5080133			ug/L	0.25	0.83	<0.25							
2,2-Dichloropropane	5080133			ug/L	0.50	1.7	<0.50							
1,1-Dichloropropene	5080133			ug/L	0.50	1.7	<0.50							
cis-1,3-Dichloropropene	5080133			ug/L	0.20	0.67	<0.20							
trans-1,3-Dichloropropene	5080133			ug/L	0.20	0.67	<0.20							
Isopropyl Ether	5080133			ug/L	0.50	1.7	<0.50							
Ethylbenzene	5080133			ug/L	0.50	1.7	<0.50							
Hexachlorobutadiene	5080133			ug/L	0.50	1.7	<0.50							
Isopropylbenzene	5080133			ug/L	0.20	0.67	<0.20							
p-Isopropyltoluene	5080133			ug/L	0.20	0.67	<0.20							
Methylene Chloride	5080133			ug/L	1.0	3.3	<1.0							
Methyl tert-Butyl Ether	5080133			ug/L	0.50	1.7	<0.50							
Naphthalene	5080133			ug/L	0.25	0.83	<0.25							
n-Propylbenzene	5080133			ug/L	0.50	1.7	<0.50							
Styrene	5080133			ug/L	0.20	0.67	<0.20							

WESTON SOLUTIONS
20 N. Wacker Drive Suite 1210
Chicago, IL 60606
Heidi Gorrell

Work Order: WOH0145
Project: Watertown Tire Fire E. R.
Project Number: [none]

Received: 08/04/05
Reported: 08/04/05 15:11

LABORATORY BLANK QC DATA

Analyte	Seq/ Batch	Source Result	Spike Level	Units	MDL	MRL	Result	Dup Result	% REC	Dup %REC	% REC Limits	RPD RPD	RPD Limit	Q
VOCs by SW8260B														
1,1,1,2-Tetrachloroethane	5080133			ug/L	0.25	0.83	<0.25							
1,1,2,2-Tetrachloroethane	5080133			ug/L	0.20	0.67	<0.20							
Tetrachloroethene	5080133			ug/L	0.50	1.7	<0.50							
Toluene	5080133			ug/L	0.20	0.67	<0.20							
1,2,3-Trichlorobenzene	5080133			ug/L	0.25	0.83	<0.25							
1,2,4-Trichlorobenzene	5080133			ug/L	0.25	0.83	<0.25							
1,1,1-Trichloroethane	5080133			ug/L	0.50	1.7	<0.50							
1,1,2-Trichloroethane	5080133			ug/L	0.25	0.83	<0.25							
Trichloroethene	5080133			ug/L	0.20	0.67	<0.20							
Trichlorofluoromethane	5080133			ug/L	0.50	1.7	<0.50							
1,2,3-Trichloropropane	5080133			ug/L	0.50	1.7	<0.50							
1,2,4-Trimethylbenzene	5080133			ug/L	0.20	0.67	<0.20							
1,3,5-Trimethylbenzene	5080133			ug/L	0.20	0.67	<0.20							
Vinyl chloride	5080133			ug/L	0.20	0.67	<0.20							
Xylenes, Total	5080133			ug/L	0.50	1.7	<0.50							
<i>Surrogate: Dibromofluoromethane</i>	<i>5080133</i>			ug/L					<i>101</i>		<i>89-119</i>			
<i>Surrogate: Toluene-d8</i>	<i>5080133</i>			ug/L					<i>96</i>		<i>91-109</i>			
<i>Surrogate: 4-Bromofluorobenzene</i>	<i>5080133</i>			ug/L					<i>98</i>		<i>89-114</i>			
Benzene	5080134			ug/L	0.20	0.67	<0.20							
Bromobenzene	5080134			ug/L	0.20	0.67	<0.20							
Bromochloromethane	5080134			ug/L	0.50	1.7	<0.50							
Bromodichloromethane	5080134			ug/L	0.20	0.67	<0.20							
Bromoform	5080134			ug/L	0.20	0.67	<0.20							
Bromomethane	5080134			ug/L	0.20	0.67	<0.20							
n-Butylbenzene	5080134			ug/L	0.20	0.67	<0.20							
sec-Butylbenzene	5080134			ug/L	0.25	0.83	<0.25							
tert-Butylbenzene	5080134			ug/L	0.20	0.67	<0.20							
Carbon Tetrachloride	5080134			ug/L	0.50	1.7	<0.50							
Chlorobenzene	5080134			ug/L	0.20	0.67	<0.20							
Chlorodibromomethane	5080134			ug/L	0.20	0.67	<0.20							
Chloroethane	5080134			ug/L	1.0	3.3	<1.0							
Chloroform	5080134			ug/L	0.20	0.67	<0.20							
Chloromethane	5080134			ug/L	0.20	0.67	<0.20							
2-Chlorotoluene	5080134			ug/L	0.50	1.7	<0.50							
4-Chlorotoluene	5080134			ug/L	0.20	0.67	<0.20							
1,2-Dibromo-3-chloropropane	5080134			ug/L	0.50	1.7	<0.50							
1,2-Dibromoethane (EDB)	5080134			ug/L	0.20	0.67	<0.20							
Dibromomethane	5080134			ug/L	0.20	0.67	<0.20							
1,2-Dichlorobenzene	5080134			ug/L	0.20	0.67	<0.20							
1,3-Dichlorobenzene	5080134			ug/L	0.20	0.67	<0.20							
1,4-Dichlorobenzene	5080134			ug/L	0.20	0.67	<0.20							
Dichlorodifluoromethane	5080134			ug/L	0.50	1.7	<0.50							
1,1-Dichloroethane	5080134			ug/L	0.50	1.7	<0.50							
1,2-Dichloroethane	5080134			ug/L	0.50	1.7	<0.50							

WESTON SOLUTIONS
20 N. Wacker Drive Suite 1210
Chicago, IL 60606
Heidi Gorrell

Work Order: WOH0145
Project: Watertown Tire Fire E. R.
Project Number: [none]

Received: 08/04/05
Reported: 08/04/05 15:11

LABORATORY BLANK QC DATA

Analyte	Seq/ Batch	Source Result	Spike Level	Units	MDL	MRL	Result	Dup Result	% REC	Dup %REC	% REC Limits	RPD RPD	RPD Limit	Q
VOCs by SW8260B														
1,1-Dichloroethene	5080134			ug/L	0.50	1.7	<0.50							
cis-1,2-Dichloroethene	5080134			ug/L	0.50	1.7	<0.50							
trans-1,2-Dichloroethene	5080134			ug/L	0.50	1.7	<0.50							
1,2-Dichloropropane	5080134			ug/L	0.50	1.7	<0.50							
1,3-Dichloropropane	5080134			ug/L	0.25	0.83	<0.25							
2,2-Dichloropropane	5080134			ug/L	0.50	1.7	<0.50							
1,1-Dichloropropene	5080134			ug/L	0.50	1.7	<0.50							
cis-1,3-Dichloropropene	5080134			ug/L	0.20	0.67	<0.20							
trans-1,3-Dichloropropene	5080134			ug/L	0.20	0.67	<0.20							
Isopropyl Ether	5080134			ug/L	0.50	1.7	<0.50							
Ethylbenzene	5080134			ug/L	0.50	1.7	<0.50							
Hexachlorobutadiene	5080134			ug/L	0.50	1.7	<0.50							
Isopropylbenzene	5080134			ug/L	0.20	0.67	<0.20							
p-Isopropyltoluene	5080134			ug/L	0.20	0.67	<0.20							
Methylene Chloride	5080134			ug/L	1.0	3.3	<1.0							
Methyl tert-Butyl Ether	5080134			ug/L	0.50	1.7	<0.50							
Naphthalene	5080134			ug/L	0.25	0.83	<0.25							
n-Propylbenzene	5080134			ug/L	0.50	1.7	<0.50							
Styrene	5080134			ug/L	0.20	0.67	<0.20							
1,1,1,2-Tetrachloroethane	5080134			ug/L	0.25	0.83	<0.25							
1,1,2,2-Tetrachloroethane	5080134			ug/L	0.20	0.67	<0.20							
Tetrachloroethene	5080134			ug/L	0.50	1.7	<0.50							
Toluene	5080134			ug/L	0.20	0.67	<0.20							
1,2,3-Trichlorobenzene	5080134			ug/L	0.25	0.83	<0.25							
1,2,4-Trichlorobenzene	5080134			ug/L	0.25	0.83	<0.25							
1,1,1-Trichloroethane	5080134			ug/L	0.50	1.7	<0.50							
1,1,2-Trichloroethane	5080134			ug/L	0.25	0.83	<0.25							
Trichloroethene	5080134			ug/L	0.20	0.67	<0.20							
Trichlorofluoromethane	5080134			ug/L	0.50	1.7	<0.50							
1,2,3-Trichloropropane	5080134			ug/L	0.50	1.7	<0.50							
1,2,4-Trimethylbenzene	5080134			ug/L	0.20	0.67	<0.20							
1,3,5-Trimethylbenzene	5080134			ug/L	0.20	0.67	<0.20							
Vinyl chloride	5080134			ug/L	0.20	0.67	<0.20							
Xylenes, Total	5080134			ug/L	0.50	1.7	<0.50							
Surrogate: Dibromofluoromethane	5080134			ug/L					98		89-119			
Surrogate: Toluene-d8	5080134			ug/L					99		91-109			
Surrogate: 4-Bromofluorobenzene	5080134			ug/L					99		89-114			

WESTON SOLUTIONS
20 N. Wacker Drive Suite 1210
Chicago, IL 60606
Heidi Gorrill

Work Order: WOH0145
Project: Watertown Tire Fire E. R.
Project Number: [none]

Received: 08/04/05
Reported: 08/04/05 15:11

CCV QC DATA

Analyte	Seq/ Batch	Source Result	Spike Level	Units	MDL	MRL	Result	Dup Result	% REC	Dup %REC	% REC Limits	RPD RPD	RPD Limit	Q
VOCs by SW8260B														
Benzene	5H04002		50.0	ug/L	N/A	N/A	49.7		99		80-120			
Bromobenzene	5H04002		50.0	ug/L	N/A	N/A	51.5		103		80-120			
Bromochloromethane	5H04002		50.0	ug/L	N/A	N/A	51.3		103		80-120			
Bromodichloromethane	5H04002		50.0	ug/L	N/A	N/A	52.1		104		80-120			
Bromoform	5H04002		50.0	ug/L	N/A	N/A	52.2		104		80-120			
Bromomethane	5H04002		50.0	ug/L	N/A	N/A	64.3		129		80-120			C
n-Butylbenzene	5H04002		50.0	ug/L	N/A	N/A	45.1		90		80-120			
sec-Butylbenzene	5H04002		50.0	ug/L	N/A	N/A	50.8		102		80-120			
tert-Butylbenzene	5H04002		50.0	ug/L	N/A	N/A	50.1		100		80-120			
Carbon Tetrachloride	5H04002		50.0	ug/L	N/A	N/A	54.3		109		80-120			
Chlorobenzene	5H04002		50.0	ug/L	N/A	N/A	51.1		102		80-120			
Chlorodibromomethane	5H04002		50.0	ug/L	N/A	N/A	52.2		104		80-120			
Chloroethane	5H04002		50.0	ug/L	N/A	N/A	48.8		98		80-120			
Chloroform	5H04002		50.0	ug/L	N/A	N/A	50.1		100		80-120			
Chloromethane	5H04002		50.0	ug/L	N/A	N/A	45.1		90		80-120			
2-Chlorotoluene	5H04002		50.0	ug/L	N/A	N/A	48.0		96		80-120			
4-Chlorotoluene	5H04002		50.0	ug/L	N/A	N/A	59.0		118		80-120			
1,2-Dibromo-3-chloropropane	5H04002		50.0	ug/L	N/A	N/A	50.9		102		80-120			
1,2-Dibromoethane (EDB)	5H04002		50.0	ug/L	N/A	N/A	50.8		102		80-120			
Dibromomethane	5H04002		50.0	ug/L	N/A	N/A	53.8		108		80-120			
1,2-Dichlorobenzene	5H04002		50.0	ug/L	N/A	N/A	50.7		101		80-120			
1,3-Dichlorobenzene	5H04002		50.0	ug/L	N/A	N/A	51.0		102		80-120			
1,4-Dichlorobenzene	5H04002		50.0	ug/L	N/A	N/A	50.8		102		80-120			
Dichlorodifluoromethane	5H04002		50.0	ug/L	N/A	N/A	53.4		107		80-120			
1,1-Dichloroethane	5H04002		50.0	ug/L	N/A	N/A	47.0		94		80-120			
1,2-Dichloroethane	5H04002		50.0	ug/L	N/A	N/A	50.9		102		80-120			
1,1-Dichloroethene	5H04002		50.0	ug/L	N/A	N/A	50.4		101		80-120			
cis-1,2-Dichloroethene	5H04002		50.0	ug/L	N/A	N/A	51.1		102		80-120			
trans-1,2-Dichloroethene	5H04002		50.0	ug/L	N/A	N/A	48.4		97		80-120			
1,2-Dichloropropane	5H04002		50.0	ug/L	N/A	N/A	50.2		100		80-120			
1,3-Dichloropropane	5H04002		50.0	ug/L	N/A	N/A	50.6		101		80-120			
2,2-Dichloropropane	5H04002		50.0	ug/L	N/A	N/A	57.1		114		80-120			
1,1-Dichloropropene	5H04002		50.0	ug/L	N/A	N/A	50.7		101		80-120			
cis-1,3-Dichloropropene	5H04002		50.0	ug/L	N/A	N/A	51.2		102		80-120			
trans-1,3-Dichloropropene	5H04002		50.0	ug/L	N/A	N/A	51.9		104		80-120			
Isopropyl Ether	5H04002		50.0	ug/L	N/A	N/A	46.7		93		80-120			
Ethylbenzene	5H04002		50.0	ug/L	N/A	N/A	50.6		101		80-120			
Hexachlorobutadiene	5H04002		50.0	ug/L	N/A	N/A	51.3		103		80-120			
Isopropylbenzene	5H04002		50.0	ug/L	N/A	N/A	51.3		103		80-120			
p-Isopropyltoluene	5H04002		50.0	ug/L	N/A	N/A	49.1		98		80-120			
Methylene Chloride	5H04002		50.0	ug/L	N/A	N/A	44.9		90		80-120			
Methyl tert-Butyl Ether	5H04002		50.0	ug/L	N/A	N/A	46.5		93		80-120			
Naphthalene	5H04002		50.0	ug/L	N/A	N/A	40.3		81		80-120			
n-Propylbenzene	5H04002		50.0	ug/L	N/A	N/A	51.4		103		80-120			
Styrene	5H04002		50.0	ug/L	N/A	N/A	51.8		104		80-120			

WESTON SOLUTIONS
20 N. Wacker Drive Suite 1210
Chicago, IL 60606
Heidi Gorrell

Work Order: WOH0145
Project: Watertown Tire Fire E. R.
Project Number: [none]

Received: 08/04/05
Reported: 08/04/05 15:11

CCV QC DATA

Analyte	Seq/ Batch	Source Result	Spike Level	Units	MDL	MRL	Result	Dup Result	% REC	Dup %REC	% REC Limits	RPD RPD	RPD Limit	Q
VOCs by SW8260B														
1,1,1,2-Tetrachloroethane	5H04002		50.0	ug/L	N/A	N/A	51.4		103		80-120			
1,1,2,2-Tetrachloroethane	5H04002		50.0	ug/L	N/A	N/A	50.5		101		80-120			
Tetrachloroethene	5H04002		50.0	ug/L	N/A	N/A	52.8		106		80-120			
Toluene	5H04002		50.0	ug/L	N/A	N/A	50.8		102		80-120			
1,2,3-Trichlorobenzene	5H04002		50.0	ug/L	N/A	N/A	42.5		85		80-120			
1,2,4-Trichlorobenzene	5H04002		50.0	ug/L	N/A	N/A	42.7		85		80-120			
1,1,1-Trichloroethane	5H04002		50.0	ug/L	N/A	N/A	51.7		103		80-120			
1,1,2-Trichloroethane	5H04002		50.0	ug/L	N/A	N/A	50.7		101		80-120			
Trichloroethene	5H04002		50.0	ug/L	N/A	N/A	52.0		104		80-120			
Trichlorofluoromethane	5H04002		50.0	ug/L	N/A	N/A	52.1		104		80-120			
1,2,3-Trichloropropane	5H04002		50.0	ug/L	N/A	N/A	50.3		101		80-120			
1,2,4-Trimethylbenzene	5H04002		50.0	ug/L	N/A	N/A	45.2		90		80-120			
1,3,5-Trimethylbenzene	5H04002		50.0	ug/L	N/A	N/A	47.4		95		80-120			
Vinyl chloride	5H04002		50.0	ug/L	N/A	N/A	48.9		98		80-120			
Xylenes, Total	5H04002		150	ug/L	N/A	N/A	155		103		80-120			
<i>Surrogate: Dibromofluoromethane</i>	<i>5H04002</i>			ug/L					<i>102</i>		<i>80-120</i>			
<i>Surrogate: Toluene-d8</i>	<i>5H04002</i>			ug/L					<i>100</i>		<i>80-120</i>			
<i>Surrogate: 4-Bromofluorobenzene</i>	<i>5H04002</i>			ug/L					<i>102</i>		<i>80-120</i>			
Benzene	5H04003		50.0	ug/kg wet	N/A	N/A	47.0		94		80-120			
Bromobenzene	5H04003		50.0	ug/kg wet	N/A	N/A	50.3		101		80-120			
Bromochloromethane	5H04003		50.0	ug/kg wet	N/A	N/A	46.8		94		80-120			
Bromodichloromethane	5H04003		50.0	ug/kg wet	N/A	N/A	46.8		94		80-120			
Bromoform	5H04003		50.0	ug/kg wet	N/A	N/A	48.9		98		80-120			
Bromomethane	5H04003		50.0	ug/kg wet	N/A	N/A	56.3		113		80-120			
n-Butylbenzene	5H04003		50.0	ug/kg wet	N/A	N/A	48.4		97		80-120			
sec-Butylbenzene	5H04003		50.0	ug/kg wet	N/A	N/A	49.2		98		80-120			
tert-Butylbenzene	5H04003		50.0	ug/kg wet	N/A	N/A	49.6		99		80-120			
Carbon Tetrachloride	5H04003		50.0	ug/kg wet	N/A	N/A	49.7		99		80-120			
Chlorobenzene	5H04003		50.0	ug/kg wet	N/A	N/A	49.6		99		80-120			
Chlorodibromomethane	5H04003		50.0	ug/kg wet	N/A	N/A	50.5		101		80-120			
Chloroethane	5H04003		50.0	ug/kg wet	N/A	N/A	46.5		93		80-120			
Chloroform	5H04003		50.0	ug/kg wet	N/A	N/A	47.5		95		80-120			
Chloromethane	5H04003		50.0	ug/kg wet	N/A	N/A	41.2		82		80-120			
2-Chlorotoluene	5H04003		50.0	ug/kg wet	N/A	N/A	47.1		94		80-120			
4-Chlorotoluene	5H04003		50.0	ug/kg wet	N/A	N/A	46.6		93		80-120			
1,2-Dibromo-3-chloropropane	5H04003		50.0	ug/kg wet	N/A	N/A	48.6		97		80-120			
1,2-Dibromoethane (EDB)	5H04003		50.0	ug/kg wet	N/A	N/A	49.4		99		80-120			
Dibromomethane	5H04003		50.0	ug/kg wet	N/A	N/A	49.4		99		80-120			
1,2-Dichlorobenzene	5H04003		50.0	ug/kg wet	N/A	N/A	49.3		99		80-120			
1,3-Dichlorobenzene	5H04003		50.0	ug/kg wet	N/A	N/A	49.6		99		80-120			
1,4-Dichlorobenzene	5H04003		50.0	ug/kg wet	N/A	N/A	49.3		99		80-120			
Dichlorodifluoromethane	5H04003		50.0	ug/kg wet	N/A	N/A	54.4		109		80-120			
1,1-Dichloroethane	5H04003		50.0	ug/kg wet	N/A	N/A	45.8		92		80-120			
1,2-Dichloroethane	5H04003		50.0	ug/kg wet	N/A	N/A	49.0		98		80-120			

WESTON SOLUTIONS
20 N. Wacker Drive Suite 1210
Chicago, IL 60606
Heidi Gorrell

Work Order: WOH0145
Project: Watertown Tire Fire E. R.
Project Number: [none]

Received: 08/04/05
Reported: 08/04/05 15:11

CCV QC DATA

Analyte	Seq/ Batch	Source Result	Spike Level	Units	MDL	MRL	Result	Dup Result	% REC	Dup %REC	% REC Limits	RPD RPD	RPD Limit	Q
VOCs by SW8260B														
1,1-Dichloroethene	5H04003		50.0	ug/kg wet	N/A	N/A	48.6		97		80-120			
cis-1,2-Dichloroethene	5H04003		50.0	ug/kg wet	N/A	N/A	48.7		97		80-120			
trans-1,2-Dichloroethene	5H04003		50.0	ug/kg wet	N/A	N/A	48.1		96		80-120			
1,2-Dichloropropane	5H04003		50.0	ug/kg wet	N/A	N/A	45.0		90		80-120			
1,3-Dichloropropane	5H04003		50.0	ug/kg wet	N/A	N/A	46.6		93		80-120			
2,2-Dichloropropane	5H04003		50.0	ug/kg wet	N/A	N/A	50.6		101		80-120			
1,1-Dichloropropene	5H04003		50.0	ug/kg wet	N/A	N/A	46.6		93		80-120			
cis-1,3-Dichloropropene	5H04003		50.0	ug/kg wet	N/A	N/A	46.9		94		80-120			
trans-1,3-Dichloropropene	5H04003		50.0	ug/kg wet	N/A	N/A	49.1		98		80-120			
2,3-Dichloropropene	5H04003		50.0	ug/kg wet	N/A	N/A	50.4		101		80-120			
Isopropyl Ether	5H04003		50.0	ug/kg wet	N/A	N/A	47.6		95		80-120			
Ethylbenzene	5H04003		50.0	ug/kg wet	N/A	N/A	49.2		98		80-120			
Hexachlorobutadiene	5H04003		50.0	ug/kg wet	N/A	N/A	47.1		94		80-120			
Isopropylbenzene	5H04003		50.0	ug/kg wet	N/A	N/A	49.8		100		80-120			
p-Isopropyltoluene	5H04003		50.0	ug/kg wet	N/A	N/A	50.4		101		80-120			
Methylene Chloride	5H04003		50.0	ug/kg wet	N/A	N/A	47.4		95		80-120			
Methyl tert-Butyl Ether	5H04003		50.0	ug/kg wet	N/A	N/A	50.2		100		80-120			
Naphthalene	5H04003		50.0	ug/kg wet	N/A	N/A	48.0		96		80-120			
n-Propylbenzene	5H04003		50.0	ug/kg wet	N/A	N/A	51.0		102		80-120			
Styrene	5H04003		50.0	ug/kg wet	N/A	N/A	51.7		103		80-120			
1,1,1,2-Tetrachloroethane	5H04003		50.0	ug/kg wet	N/A	N/A	50.2		100		80-120			
1,1,2,2-Tetrachloroethane	5H04003		50.0	ug/kg wet	N/A	N/A	46.9		94		80-120			
Tetrachloroethene	5H04003		50.0	ug/kg wet	N/A	N/A	49.9		100		80-120			
Toluene	5H04003		50.0	ug/kg wet	N/A	N/A	49.4		99		80-120			
1,2,3-Trichlorobenzene	5H04003		50.0	ug/kg wet	N/A	N/A	48.9		98		80-120			
1,2,4-Trichlorobenzene	5H04003		50.0	ug/kg wet	N/A	N/A	50.0		100		80-120			
1,1,1-Trichloroethane	5H04003		50.0	ug/kg wet	N/A	N/A	49.5		99		80-120			
1,1,2-Trichloroethane	5H04003		50.0	ug/kg wet	N/A	N/A	48.4		97		80-120			
Trichloroethene	5H04003		50.0	ug/kg wet	N/A	N/A	50.0		100		80-120			
Trichlorofluoromethane	5H04003		50.0	ug/kg wet	N/A	N/A	49.0		98		80-120			
1,2,3-Trichloropropane	5H04003		50.0	ug/kg wet	N/A	N/A	49.4		99		80-120			
1,2,4-Trimethylbenzene	5H04003		50.0	ug/kg wet	N/A	N/A	50.8		102		80-120			
1,3,5-Trimethylbenzene	5H04003		50.0	ug/kg wet	N/A	N/A	50.5		101		80-120			
Vinyl chloride	5H04003		50.0	ug/kg wet	N/A	N/A	43.1		86		80-120			
Xylenes, total	5H04003		150	ug/kg wet	N/A	N/A	150		100		80-120			
Surrogate: Dibromofluoromethane	5H04003			ug/kg wet					98		80-120			
Surrogate: Toluene-d8	5H04003			ug/kg wet					100		80-120			
Surrogate: 4-Bromofluorobenzene	5H04003			ug/kg wet					98		80-120			

WESTON SOLUTIONS
20 N. Wacker Drive Suite 1210
Chicago, IL 60606
Heidi Gorrell

Work Order: WOH0145
Project: Watertown Tire Fire E. R.
Project Number: [none]

Received: 08/04/05
Reported: 08/04/05 15:11

CERTIFICATION SUMMARY

TestAmerica Analytical - Watertown

Method	Matrix	Nelac	Wisconsin
SW 8260B	Water - NonPotable	X	X

DATA QUALIFIERS AND DEFINITIONS

C Calibration Verification recovery was above the method control limit for this analyte. Analyte not detected, data not impacted.
J Results reported between the Method Detection Limit (MDL) and Limit of Quantitation (LOQ) are less certain than results at or above the LOQ.

ADDITIONAL COMMENTS

Results are reported on a wet weight basis unless otherwise noted.

