

# Region III Equipment Condition Codes- Current as of 3/15/2013

Air Monitoring Equipment					
Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
MD H&S	Alnor Electronic Balometer APM 150 w/Probe	3771 (Base S/N 9315)	S96029	4	Next Cal Due: 10/18/2013
MD H&S	Alnor Electronic Velometer AVM 430	AVM431010006	S96021	4	Next Cal Due: 7/18/2013
MD H&S	Alnor Electronic Compu-Flow Thermo-Anemometer CF 8570	56120224	S96026	4	Next Cal Due: 11/26/2013
BTHWYN	Draeger Accuro Bellows Pump	ARTE-F005	940347	4	W/ C. Wagner
BTHWYN	Drager Model 31 Bellows Pump	0005771488	720178	4	Repaired and restored to service.
WV	Draeger Accuro Bellows Pump	940346	940346	4	
WV	Draeger Model 31 Pump	720179	720179	4	
WV	Draeger Model 31 Pump	609594	609594	4	
BTHWYN	Gastec GV-100 Detector Tube Pump	940749	940749	4	
BTHWYN	MIE DataRAM 4 Particle Sizing Monitor	D262	A41650	4	Next Cal Due: 2/27/2014
WV	MIE DataRAM 4 Particle Sizing Monitor	D264	A41651	4	Next Cal Due: 8/9/2013
BTHWYN	MIE DataRAM 4 Particle Sizing Monitor	D265	A41649	4	Next Cal Due: 6/12/2013
BTHWYN	MIE DataRAM 4 Particle Sizing Monitor	D266	A41648	4	Next Cal Due: 9/24/2013 OOS for repair
BTHWYN	MIE DataRAM 4 Particle Sizing Monitor	D267	A41647 new battery 4/15/10	7	Next Cal Due: 1/4/2014
WV	MIE DataRAM 4 Particle Sizing Monitor	D687	940745	4	Next Cal Due: 11/8/2013
BTHWYN	MIE DataRAM 4 Particle Sizing Monitor	D691	940829	4	Next Cal Due: 11/9/2013
BTHWYN	MIE DataRAM 4 Particle Sizing Monitor	D695	940828	4	Next Cal Due: 11/8/2013
BTHWYN	MIE pDR-1500 Active Personal Particulate Monitor	0919036747	940838	4	Next Cal Due: 3/8/2014.
BTHWYN	MIE pDR-1500 Active Personal Particulate Monitor	0929437673	940839	4	Next Cal Due: 1/12/2013
BTHWYN	MIE pDR-1500 Active Personal Particulate Monitor	0918736916	940840	4	Next Cal Due: 10/14/2012
WV	MIE pDR-1500 Active Personal Particulate Monitor	0919036748	940841	4	Next Cal Due: 12/7/2013
PHL H&S	Shortridge Inst ADM-870C Micromanometer	M93771	940792	4	Next Cal Due: 11/10/2013
PHL H&S	Haz Dust, 10uM Particulate Monitor	696384	963434	4	Next Cal Due: 2/12/2014
PHL H&S	SperScientific Anemometer a/flow,H,T,Lux	C3871 7	C3871 7	7	Cannot locate
PHL H&S	TSI IAQ-Calc 8762 CO/CO2/RH/Temp	00080251	940686	4	Next Cal Due: 8/16/2012
MD H&S	TSI IAQ-Calc 7545 CO/CO2/RH/Temp	T75451012006	S96028	4	Next Cal Due: 6/21/2013
WV H&S	TSI IAQ-Calc 7545 CO/CO2/RH/Temp	T75451012007	940960	4	Next Cal Due: 6/5/2013
PHL H&S	TSI IAQ-Calc 7545 CO/CO2/RH/Temp	T75451012009	940961	4	Next Cal Due: 3/21/2013 OOS for cal
PHL H&S	TSI Model 8525 Aerosol Monitor P-Trak	8525-08011021	940426 @ PH&S	4	Cal Due 7/13/10 need A/C adapter PN2613194
PHL H&S	TSI Q-Trak 8551 Air Quality Monitor	50065	940763 (Cal. Only )	4	Cal Due 10/8/2011
PHL H&S	TSI Q-Trak 8551 Air Quality Monitor	51942	940673 (Cal. Only )	4	Next Cal Due: 7/27/2012

# Region III Equipment Condition Codes- Current as of 3/15/2013

Air Sampling Pumps					
Owning Location	Item Description	Serial Number	BTHWYN	Condition Code	Notes
PHL H&S	Allegro A100 High Vol Pump	61746	940808	4	
PHL H&S	Allegro A100 High Vol Pump	61747	940809	4	
PHL H&S	Allegro A100 High Vol Pump	61748	940806	4	
BTHWYN	Bios DryCal DC-Lite M (10 ml to 12 lpm)	101238	A73057	4	Next Cal Due: 6/13/2013
WV	Bios DryCal DC-Lite M (10 ml to 12 lpm)	101236	A73056	4	Next Calibration Due: 3/5/2014
BTHWYN	Bios Nexus DCNL Module Datalogger	101232	A73058	4	
MD H&S	CDS PAS 3000 Personal Air Sampler	TBD	na	4	
PHL H&S	Gast Model 1 532-V106-G557X High Vol Pump	110120281	940807	4	
PHL H&S	Gast Model 1 532-V106-G557X High Vol Pump	110120365	940810	4	
PHL H&S	Gilian 3500 Pump	101352	940620 Keep @ COOP	4	
PHL H&S	Gilian 3500 Pump	101366	940622 Keep @ COOP	4	OOS - Battery
PHL H&S	Gilian 3500 Pump	101384	940621 Keep @ COOP	4	
PHL H&S	Gilian 3500 Pump	101392	940623 Keep @ COOP	4	
PHL H&S	Gilian 3500 Pump	101396	940624 Keep @ COOP	4	
PHL H&S	Gilian 3500 Pump (5pk) Charger	013-007	940625 Keep @ COOP	4	
BTHWYN	Gilian Aircon-2 High Vol Sampler	0311647	940277	4	
BTHWYN	Gilian Aircon-2 High Vol Sampler	0311648	940278	4	OOS for repair
BTHWYN	Gilian Aircon-2 High Vol Sampler	0311650	940279	4	
BTHWYN	Gilian Aircon-2 High Vol Sampler	0311651	940280	4	
BTHWYN	Gilian Aircon-2 High Vol Sampler	0311653	940281	4	
BTHWYN	Gilian Aircon-2 High Vol Sampler	0311654	940282	4	
BTHWYN	Gilian GilAir 3 Battery Maintenance Station(Charger)	D0904	939907	4	
BTHWYN	Gilian GilAir 3 Low Volume Pump	15256	940722	4	
BTHWYN	Gilian GilAir 3 Low Volume Pump	15257	940723	4	
BTHWYN	Gilian GilAir 3 Low Volume Pump	15258	940724	4	
BTHWYN	Gilian GilAir 3 Low Volume Pump	15259	940725	4	
BTHWYN	Gilian GilAir 3 Low Volume Pump	15260	940726	4	
BTHWYN	Gilian GilAir 5 Constant Flow Pump	20030914017	A73044	4	
BTHWYN	Gilian GilAir 5 Constant Flow Pump	20030914018	A73045	4	
BTHWYN	Gilian GilAir 5 Constant Flow Pump	20030914019	A73046	4	

# Region III Equipment Condition Codes- Current as of 3/15/2013

Air Sampling Pumps					
Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
BTHWYN	Gilian GilAir 5 Constant Flow Pump	20030914020	A73047	4	
BTHWYN	Gilian GilAir 5 Constant Flow Pump	20030914021	A73048	4	
BTHWYN	Gilian GilAir 5 Constant Flow Pump	20030916035	940230	4	
BTHWYN	Gilian GilAir 5 Constant Flow Pump	20030916036	940231	4	
BTHWYN	Gilian GilAir 5 Constant Flow Pump	20030916037	940232	4	
BTHWYN	Gilian GilAir 5 Constant Flow Pump	20030916038	940901	4	
BTHWYN	Gilian GilAir 5 Constant Flow Pump	20030916039	940234	4	
BTHWYN	Gilian GilAir 5 Constant Flow Pump	20030919001	940228	4	
BTHWYN	Gilian GilAir 5 Constant Flow Pump	20030919002	940227	4	
BTHWYN	Gilian GilAir 5 Constant Flow Pump	20030919003	940226	4	
BTHWYN	Gilian GilAir 5 Constant Flow Pump	20030919004	940225	4	
BTHWYN	Gilian GilAir 5 Constant Flow Pump	20030919005	940223	4	
WV	Gilian GilAir 5 Constant Flow Pump	20030914002	A73042	4	
WV	Gilian GilAir 5 Constant Flow Pump	20030914003	A73041	4	
WV	Gilian GilAir 5 Constant Flow Pump	20030914004	A73040	4	
WV	Gilian GilAir 5 Constant Flow Pump	20030914005	A73038	4	
WV	Gilian GilAir 5 Constant Flow Pump	20030914006	A73039	4	
WV	Gilian GilAir 5 Constant Flow Pump	20030914012	A73050	4	
WV	Gilian GilAir 5 Constant Flow Pump	20030914013	A73051	4	
WV	Gilian GilAir 5 Constant Flow Pump	20030914014	A73052	4	
WV	Gilian GilAir 5 Constant Flow Pump	20030914015	A73053	4	
WV	Gilian GilAir 5 Constant Flow Pump	20030914016	A73054	4	
BTHWYN	Gilian GilAir 5 Constant Flow Pump (5pk) Charger	0310165	940222	4	
BTHWYN	Gilian GilAir 5 Constant Flow Pump (5pk) Charger	0307098	A73049	4	
BTHWYN	Gilian GilAir 5 Constant Flow Pump (5pk) Charger	0307141	940229	4	
WV	Gilian GilAir 5 Constant Flow Pump (5pk) Charger	0307121	A73043	4	
WV	Gilian GilAir 5 Constant Flow Pump (5pk) Charger	0307129	A73055	4	
PHL H&S	Gilian Gilibrator Flow Cell Calibrator	6859-B	939779	4	Next Cal Due: 7/27/2012
PHL H&S	Gilian Bubble Generator Standard Flow	10992-S	940797 P/N 800286	4	Next Cal Due: 7/27/2012
PHL H&S	Gilian Bubble Generator - High	15220-H	940796 P/N D-800285	4	Next Cal Due: 7/27/2012

# Region III Equipment Condition Codes- Current as of 3/15/2013

Air Sampling Pumps					
Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
BTHWYN	SKC 320-5-245 Rotameter Low Flow	A41646	A41646	4	
BTHWYN	SKC 320-5-245 Rotameter Low Flow	A73091	A73091	4	
BTHWYN	SKC 320-5-4600 Rotameter High Flow	A73092	A73092	4	
BTHWYN	SKC 320-5-4600 Rotameter High Flow	A73093	A73093	4	
BTHWYN	SKC 320-235-4600 Rotameter Dual Range	A41644	A41644	4	
BTHWYN	SKC 320-235-4600 Rotameter Dual Range	A41645	A41645	4	
BTHWYN	Thermo GMW GPS-1 PUF	0304-333	940377	4	
BTHWYN	Thermo GMW GPS-1 PUF	0304-335	940386	4	
BTHWYN	Thermo GMW GPS-1 PUF	0304-337	940378	4	
BTHWYN	Thermo GMW GPS-1 PUF	0304-338	940380	4	
BTHWYN	Thermo GMW GPS-1 PUF	0304-339	940379	4	
BTHWYN	Thermo GMW GPS-1 PUF	0104-363	940390	4	
WV	Thermo GMW GS2313 TSP	0304-161	940385	4	
WV	Thermo GMW GS2313 TSP	0204-150	940389	4	
WV	Thermo GMW GS2313 TSP	0204-151	940387	4	
WV	Thermo GMW GS2313 TSP	0204-153	940400	4	
WV	Thermo GMW GS2313 TSP	0204-152	940403	4	
WV	Thermo GMW GS2313 TSP	0204-154	940391	4	
BTHWYN	Thermo GMW GUV-16H PM10	7090/41043	940392/940382	4	
BTHWYN	Thermo GMW GUV-16H PM10	7091/41042	940402/940374	4	
BTHWYN	Thermo GMW GUV-16H PM10	7092/41045	940369/940372	4	
BTHWYN	Thermo GMW GUV-16H PM10	7093/41041	940370/940371	4	
BTHWYN	Thermo GMW GUV-16H PM10	7094/41044	940368/940373	4	
BTHWYN	Thermo GMW GUV-16H PM10	7098/41046	940384/940381	4	
BTHWYN	Thermo GMW GV2360 TSP	7065	940376	4	
BTHWYN	Thermo GMW GV2360 TSP	7066	940383	4	
BTHWYN	Thermo GMW GV2360 TSP	7067	940375	4	
BTHWYN	Thermo GMW GV2360 TSP	7068	940388	4	
BTHWYN	Thermo GMW GV2360 TSP	7069	940393	4	
BTHWYN	Thermo GMW GV2360 TSP	7070	940401	4	

# Region III Equipment Condition Codes- Current as of 3/15/2013

Automated External Defibrillators					
Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
BTHWYN	Philips HeartStart FR2+ Defibrillator Model M3861A	0309299603	940842	4	New 11/09 Phila. MCP
BTHWYN	Philips HeartStart FR2+ Defibrillator Model M3861A	0309299596	940843	4	New 11/09 Wheeling MCP
BTHWYN	Philips HeartStart FR2+ Defibrillator Model M3861A	0309299663	940844	4	New 11/09 CSV
BTHWYN	Philips HeartStart FR2+ Defibrillator Model M3861A	0309299308	940845	4	New 11/09 @ COOP
BTHWYN	Philips HeartStart FR2+ Defibrillator Model M3861A	0309299564	940846	4	New 11/09 @ COOP

Audio – Visual					
Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
BTHWYN	Small Satellite Dish	DT960A4306D0008	SA5253	4	New 11/2008
BTHWYN	"I" Direct Controller Model 3100 (for Satellite Dish)	048238	940522	4	New 11/2008
BTHWYN	GPS Map 76	00415912	A41619	4	Returned by M. Aquino
BTHWYN	Garmin Nuvi 360	10368526	S37841	4	Returned by M. Aquino w/ CSV 12/2009
PHL H&S	Extech Model 407026 Light LUXFc Meter	F042667	940681	4	Cal Due 6/2/10
BTHWYN	Iridium 9505 Satellite Phone Unassg	300003000322130 520LCO	940470	4	@ COOP No SIM Card
BTHWYN	Iridium 9505 Satellite Phone Unassg.	300003000632620 530986	940471	4	@ COOP No SIM Card
BTHWYN	Iridium 9505 Satellite Phone F. Cruz.	300003000327100 520L95	940469 8816 2249 1176	4	@ COOP New Sim 10/9/08
BTHWYN	Iridium 9505 Satellite Phone Unassg.	300003000300470 500M7X	A73069	4	@ COOP NO SIM CARD
BTHWYN	Iridium 9505 Satellite Phone Unassg.	300003000310380	A73072 8816 2249 1171	4	@ COOP New Sim 10/9/08
BTHWYN	Iridium 9505 Satellite Phone L. Casillas	300003000312360	A73076 8816 2249 1173	4	New SIM 1/26/2009 @ COOP
BTHWYN	Iridium 9505 Satellite Phone R. Fetzer	300003000301460	A73077 8816 2249 1175	4	New Sim 10/9/08
BTHWYN	Iridium 9505 Satellite Phone J. Kelly	300003000315360	A73078 8816 2249 1165	4	New Sim 10/2011
BTHWYN	Iridium 9505 Satellite Phone S. Jarvela	300003000309460	A73079 8816 2249 1182	4	New Sim 11/7/08
MD	Iridium 9505 Satellite Phone G. Ham	300003000303460	A73080 8816 2249 1172	4	New Sim 10/9/08
BTHWYN	Iridium 9505 Satellite Phone M. Bartos	300003000311360	A73081 8816 2249 1164	4	New Sim 10/2011
BTHWYN	Iridium 9505 Satellite Phone A. DiDonato	300003000303450 500MH8	A73082 8816 2249 1169	4	@ COOP New SIM 3/16/09
BTHWYN	Iridium 9505 Satellite Phone R. Kelly	300003000304450	A73083 8816 2249 1170	4	New Sim 10/2011
BTHWYN	Iridium 9505 Satellite Phone M. Towle	300003000309450 500LTR	A73087 8816 2249 1166	4	New Sim 10/2011
WV	Iridium 9505 Satellite Phone D. Matlock	300003000313360 500LSJ	A73088 8816 2249 1161	4	New Sim 10/2011 inWV Acc in WV
BTHWYN	Iridium 9505 Satellite Phone D. McLaughlin	300003000302460 500GNX	A73089 8816 2249 1168	4	New Sim 10/2011
BTHWYN	Iridium 9505 Satellite Phone Unassigned	300003000328100	940473 8816 2249 1159	4	New Sim 10/2011 in WV Acc in WV

# Region III Equipment Condition Codes- Current as of 3/15/2013

Audio – Visual							
Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes		
BTHWYN	Iridium 9505 Satellite Phone C. Wagner.	300003000637620 5304SH	940472 8816 2249 1157	4	New Sim 10/2011		
BTHWYN	Iridium 9505 Satellite Phone D. Ventura	300003000300460	A73070 8816 5146 3148	4	New Sim 10/2011		
MD	Iridium 9505 Satellite Phone C. Santiago	300003000319370 500M6Z	A73074 8816 2249 1162	4	New SIM 10/2011		
WV	Iridium 9505 Satellite Phone D. Lindsey	300003000308440 500MDN	A73073 8816 2249 1160	4	New Sim 10/2011	in WV	Acc in WV
BTHWYN	Iridium 9505 Satellite Phone Unassigned	300003000301450 500LWY	A73075 8816 2249 1158	4	New Sim 10/2011	in WV	Acc in WV
MD	Iridium 9505 Satellite Phone C. Fitzsimmons	300003000300450	A73068 8816 2249 1167	4	New Sim 10/2011		
BTHWYN	Iridium 9505 Satellite Phone V. Zenone	300003000308450	A73071 8816 2249 1163	4	New Sim 10/2011		
MD	Iridium 9505 Satellite Phone	500GNX	A73090 ( lost @ Katrina ? )	L	NI cannot locate or ? assigned to		
BTHWYN	Iridium 9505A Satellite Phone w/ Helen Duteau	300114040006620	940766 8816 2249 1151	4			
BTHWYN	Iridium 9505A Satellite Phone w/ Helen Duteau	300114040001620	940769 8816 2249 1153	4			
BTHWYN	JBL EON Music Mix 10 System	P1006-20232C	A04282	5			
BTHWYN	Motorola Noise Canceling Headset	940700	940700	4			
BTHWYN	Motorola Noise Canceling Headset	940701	940701	4			
BTHWYN	Motorola Noise Canceling Headset	940702	940702	4			
BTHWYN	Motorola Noise Canceling Headset	940703	940703	4			
BTHWYN	Motorola Noise Canceling Headset	940704	940704	4			
BTHWYN	Motorola Noise Canceling Headset	940705	940705	4			
BTHWYN	Motorola Noise Canceling Headset	940706	940706	4			
BTHWYN	Motorola Noise Canceling Headset	940707	940707	4			
BTHWYN	Motorola Noise Canceling Headset	940708	940708	4			
BTHWYN	Motorola Noise Canceling Headset	940709	940709	4			
BTHWYN	Motorola Noise Canceling Headset	940710	940710	4			
BTHWYN	Motorola Noise Canceling Headset	940711	940711	4			
BTHWYN	Motorola Noise Canceling Headset	940712	940712	4			
BTHWYN	Motorola Noise Canceling Headset	940713	940713	4			
BTHWYN	Motorola Noise Canceling Headset	940714	940714	4			
BTHWYN	Motorola Noise Canceling Headset	940715	940715	4			
BTHWYN	Motorola Noise Canceling Headset	940716	940716	4			
BTHWYN	Motorola Noise Canceling Headset	940717	940717	4			
BTHWYN	Motorola Noise Canceling Headset	940718	940718	4			

# Region III Equipment Condition Codes- Current as of 3/15/2013

BTHWYN	Motorola Noise Canceling Headset	940719	940719	4	
<b>Audio – Visual</b>					
Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
BTHWYN	Headset	940852	940852	4	
BTHWYN	Headset	940853	940853	4	
BTHWYN	Headset	940854	940854	4	
BTHWYN	Headset	940855	940855	4	
BTHWYN	Headset	940856	940856	4	
BTHWYN	Headset	940857	940857	4	
BTHWYN	Headset	940858	940858	4	
BTHWYN	Headset	940859	940859	4	
BTHWYN	Headset	940860	940860	4	
BTHWYN	Headset	940861	940861	4	
BTHWYN	Motorola HT1250LS VHF 2 Way Radio	749HEAA047	940252	L	Lost. Not returned from LRE 4/2010
BTHWYN	Motorola HT1250LS VHF 2 Way Radio	749HEAA048 0357	940253	4	
BTHWYN	Motorola HT1250LS VHF 2 Way Radio	749HEAA052	940254	4	
BTHWYN	Motorola HT1250LS VHF 2 Way Radio	749HEAA057 0374	940255	4	OOS – Inoperable per OSC Rovira
BTHWYN	Motorola HT1250LS VHF 2 Way Radio	749HEAA060 0367	940256	4	McLaughlin
BTHWYN	Motorola HT1250LS VHF 2 Way Radio	749HEAA076 0362	940242	9	
BTHWYN	Motorola HT1250LS VHF 2 Way Radio	749HEAA077 0359	940257	4	
BTHWYN	Motorola HT1250LS VHF 2 Way Radio	749HEAA087	940258	4	
BTHWYN	Motorola HT1250LS VHF 2 Way Radio	749HEAA115	940235	4	
BTHWYN	Motorola HT1250LS VHF 2 Way Radio	749HEAA117 0371	940243	9	
BTHWYN	Motorola HT1250LS VHF 2 Way Radio	749HEAA151 0355	940236	4	
BTHWYN	Motorola HT1250LS VHF 2 Way Radio	749HEAA152 0356	940244	4	
BTHWYN	Motorola HT1250LS VHF 2 Way Radio	749HEAA153 0369	940245	4	
BTHWYN	Motorola HT1250LS VHF 2 Way Radio	749HEAA154	940237	4	Zenone
BTHWYN	Motorola HT1250LS VHF 2 Way Radio	749HEAA155 0360	940246	4	Richardson 1-6-10
BTHWYN	Motorola HT1250LS VHF 2 Way Radio	749HEAA158 0358	940238	4	Assg. To D. Ventura
MD	Motorola HT1250LS VHF 2 Way Radio	749HEAA160 0368	940259	4	Rupert
BTHWYN	Motorola HT1250LS VHF 2 Way Radio	749HEAA162 0361	940247	4	Bartos/Rovira
BTHWYN	Motorola HT1250LS VHF 2 Way Radio	749HEAA163	940248	4	
BTHWYN	Motorola HT1250LS VHF 2 Way Radio	749HEAA174	940239	4	

# Region III Equipment Condition Codes- Current as of 3/15/2013

BTHWYN	Motorola HT1250LS VHF 2 Way Radio	749HEAA179 0366	940240	4	Ham
<b>Audio – Visual</b>					
Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
BTHWYN	Motorola HT1250LS VHF 2 Way Radio	749HEAA183	940241	4	w/ E. Rovira
WV	Motorola HT1250LS VHF 2 Way Radio	749HEAA165	940250	4	WHERE? Acc in WV
WV	Motorola HT1250LS VHF 2 Way Radio	749HEAA168 0373	940260	4	In WV
WV	Motorola HT1250LS VHF 2 Way Radio	749HEAA172 0372	940251 Needs cover and screw	4	In WV Acc in WV
WV	Motorola HT1250LS VHF 2 Way Radio	749HEAA173	940261	4	WHERE? Acc in WV
MD	Motorola HT1250LS VHF 2 Way Radio	749HEAA164	940249	4	C. Fitzsimmons
BTHWYN	Motorola HT1250LS VHF 2 Way Radio	749THW3842	940649	4	New MCP Radios
BTHWYN	Motorola HT1250LS VHF 2 Way Radio	749THW3849	940650 0353	S	Water Damage. Repairir too costly
BTHWYN	Motorola HT1250LS VHF 2 Way Radio	749THW4322	940651 0346	4	Out to BoRit Site 8/10/10
BTHWYN	Motorola HT1250LS VHF 2 Way Radio	749THW3845	940652 0344	4	New MCP Radios @ COOP
BTHWYN	Motorola HT1250LS VHF 2 Way Radio	749THW3843	940653 0348	4	New MCP Radios w/ E.Rovira
BTHWYN	Motorola HT1250LS VHF 2 Way Radio	749THW4330	940654 0341	4	New MCP Radios w/ E.Rovira
BTHWYN	Motorola HT1250LS VHF 2 Way Radio	749THW3836	940655 0345	L	Lost. Not returned from LRE 4/2010
BTHWYN	Motorola HT1250LS VHF 2 Way Radio	749THW3844	940656 0350	4	New MCP Radios
BTHWYN	Motorola HT1250LS VHF 2 Way Radio	749THW4331	940657 0340	4	New MCP Radios @ COOP
BTHWYN	Motorola HT1250LS VHF 2 Way Radio	749THW4333	940658 0349	4	New MCP Radios w/ E. Rovira
BTHWYN	Motorola HT1250LS VHF 2 Way Radio	749THW4346	940659 0343	9	Unit removed for repair
BTHWYN	Motorola HT1250LS VHF 2 Way Radio	749THW4329	940660 0351	4	New MCP Radios @ COOP
BTHWYN	Motorola HT1250LS VHF 2 Way Radio	749THW4341	940661 0342	4	
BTHWYN	Motorola HT1250LS VHF 2 Way Radio	749THW3850	940662 0347	4	New MCP Radios w/ E.Rovira
BTHWYN	Motorola HT1250LS VHF 2 Way Radio	749THW4340	940663 0375	4	New MCP Radios K.Boyd 10/16/08
BTHWYN	Motorola HT1250 UHF 2 Way Radio	749TKU1825	940862	4	New 11/09
BTHWYN	Motorola HT1250 UHF 2 Way Radio	749TKU1822	940863	4	New 11/09
BTHWYN	Motorola HT1250 UHF 2 Way Radio	749TKU1865	940864	4	New 11/09
BTHWYN	Motorola HT1250 UHF 2 Way Radio	749TKU1819	940865	4	New 11/09 @ COOP E.R. 3/29/10
BTHWYN	Motorola HT1250 UHF 2 Way Radio	749TKU1826	940866	4	New 11/09
BTHWYN	Motorola HT1250 UHF 2 Way Radio	749TKU1818	940867	4	New 11/09
BTHWYN	Motorola HT1250 UHF 2 Way Radio	749TKU1823	940868	4	New 11/09
BTHWYN	Motorola HT1250 UHF 2 Way Radio	749TKU1820	940869	4	New 11/09
BTHWYN	Motorola HT1250 UHF 2 Way Radio	749TKU1866	940870	4	New 11/09



## Region III Equipment Condition Codes- Current as of 3/15/2013

BTHWYN	Motorola HT1250 UHF 2 Way Radio	749TKU1888	940871	4	New 11/09
<b>Audio – Visual</b>					
Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
BTHWYN	Motorola HTN9005 6 Station Charger	EBTHWYN 940262	940262	4	W/ E. Rovira
BTHWYN	Motorola HTN9005 6 Station Charger	EBTHWYN 940263	940263	4	@ COOP
BTHWYN	Motorola HTN9005 6 Station Charger	EBTHWYN 940264	940264	4	@ COOP
BTHWYN	Motorola HTN9005 6 Station Charger	EBTHWYN 940265	940265	4	@ COOP
BTHWYN	Motorola HTN9005 6 Station Charger	EBTHWYN 940598	940598	7	@ COOP
BTHWYN	Motorola HTN9005 6 Station Charger	BTHWYN 940664	940664	4	@ COOP
BTHWYN	Motorola HTN9005 6 Station Charger	BTHWYN 940665	940665	4	w/ E.Rovira
BTHWYN	Motorola Impres 6-Station Charger	BTHWYN 940975	940975	4	
BTHWYN	Motorola Impres 6-Station Charger	BTHWYN 940976	940976	4	
BTHWYN	Motorola Impres 6-Station Charger	BTHWYN 940977	940977	4	
BTHWYN	Motorola Impres 6-Station Charger	BTHWYN 940978	940978	4	
BTHWYN	Motorola Impres 6-Station Charger	BTHWYN 940981	940975	4	
BTHWYN	Motorola Impres 6-Station Charger	BTHWYN 940982	940976	4	
BTHWYN	Motorola Impres 6-Station Charger	BTHWYN 940983	940977	4	
BTHWYN	Motorola Impres 6-Station Charger	BTHWYN 940984	940978	4	

## Region III Equipment Condition Codes- Current as of 3/15/2013

BTHWYN	Motorola XTS3000 2 Way Radio	620AAA0290	A41020 0302	4	
BTHWYN	Motorola XTS3000 2 Way Radio	620AAA0291	A41019 0309	4	Cannot locate, no maint. file
BTHWYN	Motorola XTS3000 2 Way Radio	620AAA0292	A41026 0327	5	Entered in C.C. 5/16/08 @ COOP
BTHWYN	Motorola XTS3000 2 Way Radio	620AAA0294	A41023 0310	5	Entered in C.C. 5/16/08 w/ E. Rovira
BTHWYN	Motorola XTS3000 2 Way Radio	620AAA0295	A41025 0305	5	Entered in C.C. 5/16/08
BTHWYN	Motorola XTS3000 2 Way Radio	620AAA0296	A41028 0311	5	Entered in C.C. 5/16/08 Zennone
BTHWYN	Motorola XTS3000 2 Way Radio	620AAA0300	A41015 0303	5	Entered in C.C. 4/21/08 @ COOP
BTHWYN	Motorola XTS3000 2 Way Radio	620AAA0301	A41022 0326	4	Needs programming
BTHWYN	Motorola XTS3000 2 Way Radio	620AAA0302	A41021 0304	4	@ COOP
BTHWYN	Motorola XTS3000 2 Way Radio	620AAA0303	A41018 0330	9	
BTHWYN	Motorola XTS3000 2 Way Radio		A41043	9	
BTHWYN	Motorola XTS3000 2 Way Radio	620ABU4435	A41047 0301	4	Needs programming
BTHWYN	Motorola XTS3000 2 Way Radio	620ABU4436	A41046 0331	9	Unit removed for reBTHWYNir
BTHWYN	Motorola XTS3000 2 Way Radio	620AAA0297	A41027 0307	4	Reassigned To COOP
BTHWYN	Motorola XTS3000 2 Way Radio	620AAA0298	A41014 0308	4	Reassigned To COOP w/ E. Rovira
BTHWYN	Motorola XTS3000 2 Way Radio	620AAA0299	A41017 0306	4	Reassigned To COOP

### Audio – Visual

Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
BTHWYN	Motorola XTS3000 2 Way Radio	620AAA0293	A41016	4	
BTHWYN	Motorola XTS5000 2 Way Radio	320CEY5320	940487 0315	4	With M. Bartos 4/15/09
BTHWYN	Motorola XTS5000 2 Way Radio	320CEY5321	940493	4	
MD	Motorola XTS5000 2 Way Radio	320CEY5322	940484 0312	4	Radio & Acc. Kit to G.Ham 8/19/08
BTHWYN	Motorola XTS5000 2 Way Radio	320CEY5323	940501 0314	4	With M. Bartos 4/15/09
WV	Motorola XTS5000 2 Way Radio	320CEY5324	940498 0329	4	
WV	Motorola XTS5000 2 Way Radio	320CEY5325	940492 0328	4	Acc in WV
WV	Motorola XTS5000 2 Way Radio	320CEY5326	940465	4	
WV	Motorola XTS5000 2 Way Radio	320CEY5339	940467	4	Acc in WV
BTHWYN	Motorola XTS5000 2 Way Radio	320CEY5327	940494	4	
BTHWYN	Motorola XTS5000 2 Way Radio	320CEY5328	940499 0316	4	
MD	Motorola XTS5000 2 Way Radio	320CEY5329	940485	4	Accessories kit in MD
BTHWYN	Motorola XTS5000 2 Way Radio	320CEY5330	940500	4	
BTHWYN	Motorola XTS5000 2 Way Radio	320CEY5331	940482 0317	4	Assg. To D. Ventura
BTHWYN	Motorola XTS5000 2 Way Radio	320CEY5332	940490	4	

## Region III Equipment Condition Codes- Current as of 3/15/2013

BTHWYN	Motorola XTS5000 2 Way Radio	320CEY5333	940496 0313	4	
BTHWYN	Motorola XTS5000 2 Way Radio	320CEY5334	940464	4	
BTHWYN	Motorola XTS5000 2 Way Radio	320CEY5335	940495	4	
BTHWYN	Motorola XTS5000 2 Way Radio	320CEY5336	940462	4	
MD	Motorola XTS5000 2 Way Radio	320CEY5337	940489	4	In MD w/ 2 "Impress" Chargers
BTHWYN	Motorola XTS5000 2 Way Radio	320CEY5338	940491	4	
BTHWYN	Motorola XTS5000 UHF 2 Way Radio	320CKV0634	940878	4	New 11/09
BTHWYN	Motorola XTS5000 UHF 2 Way Radio	320CKV0636	940879	4	New 11/09
BTHWYN	Motorola XTS5000 UHF 2 Way Radio	320CKV0641	940880	4	New 11/09
BTHWYN	Motorola XTS5000 UHF 2 Way Radio	320CKV0637	940881	4	New 11/09
BTHWYN	Motorola XTS5000 UHF 2 Way Radio	320CKV0633	940882	4	New 11/09
BTHWYN	Motorola XTS5000 UHF 2 Way Radio	320CKV0639	940883	4	New 11/09
BTHWYN	Motorola XTS5000 UHF 2 Way Radio	320CKV0642	940884	4	New 11/09
BTHWYN	Motorola XTS5000 UHF 2 Way Radio	320CKV0638	940885	4	New 11/09
BTHWYN	Motorola XTS5000 UHF 2 Way Radio	320CKV0635	940886	4	New 11/09
BTHWYN	Motorola XTS5000 UHF 2 Way Radio	320CKV0640	940887	4	New 11/09

### Audio – Visual

Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
BTHWYN	Motorola XTL 5000 UHF	585CKV0245	940872	4	New 11/09
BTHWYN	Motorola XTL 5000 UHF	585CKV0246	940873	4	New 11/09
BTHWYN	Motorola XTL 5000 UHF	585CKV0244	940874	4	New 11/09
BTHWYN	Motorola XTL 5000 800 mhz	500CKV0716	940875	4	New 11/09
BTHWYN	Motorola XTL 5000 800 mhz	500CKV0715	940876	4	New 11/09
BTHWYN	Motorola XTL 5000 800 mhz	500CKV0714	940877	4	New 11/09
BTHWYN	XTVA VHF	15411	940888	4	New 11/09
BTHWYN	XTVA VHF	15408	940889	4	New 11/09
BTHWYN	XTVA UHF	7522	940890	4	New 11/09
BTHWYN	XTVA UHF	7521	940891	4	New 11/09
BTHWYN	Evolution 5000 Thermal Imaging Camera	A2-2074-K03	A73094	4	@ COOP
MD H&S	O'Conner 1030 Camera Ultimate Tripod	339005	339005	4	

## Region III Equipment Condition Codes- Current as of 3/15/2013

BTHWYN	Scott Envoy Radiocom System	007347	940363	4	
BTHWYN	Scott Envoy Radiocom System	007364	940366	1	
BTHWYN	Scott Envoy Radiocom System	007365	940367	1	
WV	Scott Envoy Radiocom System	007370	940475	4	M. Linden
BTHWYN	Scott Envoy Radiocom System	007339	007339	4	Sharma
BTHWYN	Scott Envoy Radiocom System	007340	007340	4	Towle
BTHWYN	Scott Envoy Radiocom System	007342	007342	4	J. Kelly
BTHWYN	Scott Envoy Radiocom System	007343	940478	4	
BTHWYN	Scott Envoy Radiocom System	007344	007344	4	Jarvela
BTHWYN	Scott Envoy Radiocom System	007345	940647	4	
BTHWYN	Scott Envoy Radiocom System	007346	940803	4	
BTHWYN	Scott Envoy Radiocom System	007349	940531	4	
BTHWYN	Scott Envoy Radiocom System	007350	007350	4	Bartos
BTHWYN	Scott Envoy Radiocom System	007351	940530	4	Lapsley
BTHWYN	Scott Envoy Radiocom System	007356	007356	4	
BTHWYN	Scott Envoy Radiocom System	007357	007357	4	
BTHWYN	Scott Envoy Radiocom System	007362	940964	4	

### Audio – Visual

Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
BTHWYN	Scott Envoy Radiocom System	007363	940365	4	
BTHWYN	Scott Envoy Radiocom System	007371	940529	4	
BTHWYN	Scott Envoy Radiocom System	007425	007425	4	Fetzer
BTHWYN	Scott Envoy Radiocom System	007368	940648	4	
WV	Scott Envoy Radiocom System	007359	940537	4	LINDSEY
WV	Scott Envoy Radiocom System	007360	940542	4	EASTON
WV	Scott Envoy Radiocom System	007366	940432	4	Matlock
WV	Scott Envoy Radiocom System	007367	007367	4	Missing?
MD	Scott Envoy Radiocom System	007358	940364	4	C. Fitzsimmons NI
BTHWYN	Scott E-Z Radiocom System	03238	940735	4	
WV	Scott E-Z Radiocom System	03239	940815	4	M. Linden
BTHWYN	Scott E-Z Radiocom System	03240	940738	4	

## Region III Equipment Condition Codes- Current as of 3/15/2013

BTHWYN	Scott E-Z Radiocom System	03241	940736	4	
BTHWYN	Scott E-Z Radiocom System	03242	940739	4	
BTHWYN	Scott E-Z Radiocom System	03243	03243	4	
BTHWYN	Scott E-Z Radiocom System	03244	03244	4	
BTHWYN	Scott E-Z Radiocom System	03245	940737	4	
BTHWYN	Scott E-Z Radiocom System	03246	940817	4	w/ OSC McLaughlin
WV	Scott E-Z Radiocom System	03247	940536	4	Changed to WV 11/28/07 Lindsey
BTHWYN	Scott E-Z Radiocom System	03248	940768	4	@ COOP Ret by G. Ham
BTHWYN	Scott E-Z Radiocom System	03249	940740	4	
BTHWYN	Scott E-Z Radiocom System	03250	940741	4	
BTHWYN	Scott E-Z Radiocom System	03251	940742	4	
WV	Scott E-Z Radiocom System	03252	940540	4	
WV	Scott E-Z Radiocom System	03253	940541	4	
BTHWYN	Scott E-Z Radiocom System	03254	03254	4	
BTHWYN	Scott E-Z Radiocom System	03255	940743	4	
BTHWYN	Scott E-Z Radiocom System	03256	03256	4	
BTHWYN	Scott E-Z Radiocom System	03257	940744	4	
BTHWYN	Scott E-Z Radiocom System	03258	03258	4	
BTHWYN	Scott E-Z Radiocom System	03259	940731	4	

### Audio – Visual

Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
BTHWYN	Scott E-Z Radiocom System	03260	940733	4	
BTHWYN	Scott E-Z Radiocom System	03261	940732	4	
WV	Scott E-Z Radiocom System	03262	940430	4	
BTHWYN	Scott E-Z Radiocom System	03263	940734	4	
BTHWYN	Scott E-Z Radiocom System	03264	03264	4	
BTHWYN	Scott E-Z Radiocom System	03265	940832	4	
BTHWYN	Scott Radiocom II System	07301	940632	4	
BTHWYN	Scott Radiocom II System	07302	940633	4	
BTHWYN	Scott Radiocom II System	07303	940634	4	
BTHWYN	Scott Radiocom II System	07304	940635	4	
BTHWYN	Scott Radiocom II System	07305	940636	4	
BTHWYN	Scott Radiocom II System	7542	940631	4	

## Region III Equipment Condition Codes- Current as of 3/15/2013

BTHWYN	Scott Radiocom II System	7543	940637	4	
BTHWYN	Scott Radiocom II System	7544	940638	4	
BTHWYN	Scott Radiocom II System	7545	940639	4	
BTHWYN	Scott Radiocom II System	7546	940640	4	
BTHWYN	Scott Voice Amplifier Face Pieces	52136R	940801	4	
BTHWYN	Scott Voice Amplifier Face Pieces	52139R	940476	4	
BTHWYN	Scott Voice Amplifier Face Pieces	52142R	940533	4	
BTHWYN	Scott Voice Amplifier Face Pieces	52149R	52149R	4	
BTHWYN	Scott Voice Amplifier Face Pieces	52391R	52391 R	4	
BTHWYN	Scott Voice Amplifier Face Pieces	52392R	940481	4	
BTHWYN	Scott Voice Amplifier Face Pieces	52393R	52393R	4	
BTHWYN	Scott Voice Amplifier Face Pieces	52395R	52395R	4	
BTHWYN	Scott Voice Amplifier Face Pieces	52396R	52396R	4	
BTHWYN	Scott Voice Amplifier Face Pieces	52397R	940831	4	Lapsley
BTHWYN	Scott Voice Amplifier Face Pieces	52404R	52404R	4	
BTHWYN	Scott Voice Amplifier Face Pieces	52405R	52405R	4	
BTHWYN	Scott Voice Amplifier Face Pieces	52411R	940532	4	
BTHWYN	Scott Voice Amplifier Face Pieces	52413A	940767	4	G. Ham

### Audio – Visual

Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
BTHWYN	Scott Voice Amplifier Face Pieces	52967R	52967R	4	
BTHWYN	Scott Voice Amplifier Face Pieces	52977R	940816	4	Returned from D. McLaughlin 2/2009
BTHWYN	Scott Voice Amplifier Face Pieces	52982R	940534	4	
MD	Scott Voice Amplifier Face Pieces	52984R	940416	4	NI added to Inv 1-15-09
BTHWYN	Scott Voice Amplifier Face Pieces	52994R	940965	4	
BTHWYN	Scott Voice Amplifier Face Pieces	52147R	A73033	4	D Ventura
WV	Scott Voice Amplifier Face Pieces	52390R	940538	4	Lindsey
WV	Scott Voice Amplifier Face Pieces	52398R	940431	4	Matlock
WV	Scott Voice Amplifier Face Pieces	52408R	940429	4	Easton
WV	Scott Voice Amplifier Face Pieces	52417R	940539	4	M. Linden
MD	Scott Voice Amplifier Face Pieces	940416	940416	4	NI
BTHWYN	Iridium Transceiver Indoor Sat Phone	10987	940953	4	

## Region III Equipment Condition Codes- Current as of 3/15/2013

WV	Command Board	940837	940837	4	Inserted in Inv. 1-15-09 in W.V.
BTHWYN	Command Board	940525	940525	4	Inserted in Inv. 1-15-09
BTHWYN	Command Board	940526	940526	4	Inserted in Inv. 1-15-09
BTHWYN	Command Board	940527	940527	4	Inserted in Inv. 1-15-09
BTHWYN	Command Board	940528	940528	4	Inserted in Inv. 1-15-09
WV	Command Board	940697	940697	4	
WV	Handy Cam, HDR-SR1, SONY	342302	TBD	4	w/ Matlock 6/28/07
BTHWYN	Sony DCR-TRV38 Handycam Digital	1382644	A73097	4	R. Kelly

### Calibration Kits

Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
BTHWYN	MSA SAFEMTX Calibration Kit	940568	940568	4	
BTHWYN	MSA SAFEMTX Calibration Kit	940569	940569	4	
BTHWYN	RAE Systems MultiRAE Cal Kit	PH0604-01	940935 @ Clearview LF	4	Multi-Gas expires: 9/12/2013 C <sub>4</sub> H <sub>8</sub> expire: 2/15/2014
BTHWYN	RAE Systems MultiRAE Cal Kit	PH0604-02	940754	4	Multi-Gas expires: 9/12/2013 C <sub>4</sub> H <sub>8</sub> expire: 9/12/2013
BTHWYN	RAE Systems MultiRAE Cal Kit	PH0604-03	940755	4	Multi-Gas expires: 9/12/2013 C <sub>4</sub> H <sub>8</sub> expire: 9/12/2013
BTHWYN	RAE Systems MultiRAE Cal Kit	PH0604-04	940756	4	Multi-Gas expires: 9/12/2013 C <sub>4</sub> H <sub>8</sub> expire: 5/8/2014
BTHWYN	RAE Systems MultiRAE Cal Kit	PH0604-05	PH0604-05	4	
BTHWYN	RAE Systems MultiRAE Cal Kit	PH0604-06	PH0604-06	4	Multi-Gas expires: 9/12/2013 C <sub>4</sub> H <sub>8</sub> expire: 5/8/2014
BTHWYN	RAE Systems MultiRAE Cal Kit	PH0604-1 1	940833	4	w/ OSC C. Wagner
BTHWYN	RAE Systems MultiRAE Cal Kit	PH0604-12	PH0604-12	4	
BTHWYN	RAE Systems MultiRAE Cal Kit	PH604-08	940518	4	
BTHWYN	RAE Systems MultiRAE Cal Kit	PH604-09	940519	4	Multi-Gas expires: 2/1/2014 C <sub>4</sub> H <sub>8</sub> expire: 2/9/2014
BTHWYN	RAE Systems MultiRAE Cal Kit	PH604-10	940520	4	Multi-Gas expires: 9/12/2013 C <sub>4</sub> H <sub>8</sub> expire: 9/12/2013
WV	RAE Systems MultiRAE Cal Kit	W0604-01	940543	4	Multi-Gas expires: 9/12/2013 C <sub>4</sub> H <sub>8</sub> expire: 5/18/2014
WV	RAE Systems MultiRAE Cal Kit	W0604-02	940544	4	Multi-Gas expires: 9/12/2013 C <sub>4</sub> H <sub>8</sub> expire: 5/8/2014
WV	RAE Systems MultiRAE Cal Kit	W0604-03	940545	4	Multi-Gas expires: 11/2012 C <sub>4</sub> H <sub>8</sub> expire: 5/8/2014
WV	RAE Systems MultiRAE Cal Kit	A41011	A41011	4	
MD	RAE Systems MultiRAE Cal Kit	PH0604-07	940668	4	Exp Iso 7/11 Mixed 10/10 needs iso knob

## Region III Equipment Condition Codes- Current as of 3/15/2013

BTHWYN	BGI hivolCAL, Model HVC-1	000082		4	Next Cal Due: 8/7/2013
BTHWYN	Thermo G28A Vari-Flo Cal Kit	1868	940399	4	Next Cal Due: 9/14/2013
BTHWYN	Thermo G28A Vari-Flo Cal Kit	1869	940396	4	Next Cal Due: 7/23/2013
WV	Thermo G28A Vari-Flo Cal Kit	1870	940397	4	Next Cal Due: 11/26/2013
BTHWYN	Thermo G28A Vari-Flo Cal Kit	1871	940394	4	Next Cal Due: 7/23/2013
BTHWYN	Thermo G28A Vari-Flo Cal Kit	1872	940395	4	Next Cal Due: 7/23/2013
BTHWYN	Thermo G28A Vari-Flo Cal Kit	1873	940398	4	Next Cal Due: 9/14/2013

### Chemical Warfare Agent ID & Biological Detection

Owning Location	Item Description	Serial Number	EPA TAG NUMBERD	Condition Code	Notes
BTHWYN	Haztech HazCat Basic KT1 009 Kit	A73060	A73060	4	
WV	Haztech HazCat Basic KT1 009 Kit	200978	A73059	4	
BTH	Smith's Detection, HazMat ID Ranger	100800234A/002324 (PDA)	940985 Boothwyn		New 10/13/10



# Region III Equipment Condition Codes- Current as of 3/15/2013

Chemical Warfare Agent ID & Biological Detection					
Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
BTHWYN	Proengin AP2Ce CW Detector	00256	A73012	4	
BTHWYN	Proengin AP2Ce CW Detector	00263	A73036	4	Next Cal Due: 3/20/2013
WV	Proengin AP2Ce CW Detector	00503	A73067	4	
BTHWYN	Proengin AP2Ce Hydrogen Storage Case	940811	940811	4	
BTHWYN	Smith APD 2000	0518	A41641	4	Next Wipe Test Due: 8/11/2013
BTHWYN	Smith APD 2000	3038	A73066	4	Next Wipe Test Due: 8/11/2013
WV	Smith APD 2000	0510	A73037	4	Next Wipe Test Due: 3/22/2012
MD H&S	MSA HAZMATCAD Chemical Agent Detector	903386	A61191	4	Next sensor change due: 3/5/2013 OOS for service

Field Support Equipment					
Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
BTHWYN	Lascar EL-USB-1 Temperature Data Logger	010017506 #01	940818	4	New January 2009
BTHWYN	Lascar EL-USB-1 Temperature Data Logger	010001680 #02	940819	4	New January 2009
BTHWYN	Lascar EL-USB-1 Temperature Data Logger	010001137 #03	940820	4	New January 2009
BTHWYN	Lascar EL-USB-1 Temperature Data Logger	010017399 #04	940821	4	New January 2009
BTHWYN	Lascar EL-USB-1 Temperature Data Logger	010017477 #05	940822	4	New January 2009
BTHWYN	Lascar EL-USB-1 Temperature Data Logger	010017090 #06	940823	4	New January 2009
BTHWYN	Lascar EL-USB-1 Temperature Data Logger	010017421 #07	940824	4	New January 2009
BTHWYN	Lascar EL-USB-1 Temperature Data Logger	010017432 #08	940825	4	New January 2009
BTHWYN	Lascar EL-USB-1 Temperature Data Logger	010018198 #09	940826	4	New January 2009
BTHWYN	Lascar EL-USB-1 Temperature Data Logger	010001694 #10	940827	4	New January 2009
BTHWYN	Air Compressor, Craftsman, 2 Gal. 1 HP	059573	940790	4	At Linwood Warehouse
WV	Geometrics G856-AX Magnetometer	6558	940428	4	Next Cal Due: 3/15/2013
WV	Soil Sampling Kit-Rotary Hammer/GVP AMS	0068885896	940764	4	
PHL H&S	Hanna Inst HI8564 Thermal Hygrometer	75225	944000	4	Do Not send for Calibration A.M
PHL H&S	Narda Model 8217 Microwave Oven Tester	210067	940795	4	Next Cal Due: 11/21/2013
PHL H&S	Extech Digital Multimeter Model 380280	B974838	A04539	4	
BTHWYN	Nilfisk GSJF HEBTHWYN Vacuum	10985	623229	4	
BTHWYN	Scale, Digital	940980	940980	4	For weighing packages

# Region III Equipment Condition Codes- Current as of 3/15/2013

Metals & Mercury Vapor Analyzers					
Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
BTHWYN	Arizona Instruments Jerome 431-X	431-3745	940314	4	Next Cal Due: 8/24/2013
WV	Arizona Instruments Jerome 431-X	431-3948	A73064	4	Next Cal Due: 4/16/2013
MD H&S	Arizona Instruments Jerome 431-X	431-3928	A34481	4	Next Cal Due: 6/19/2013
WV	Thermo XLt792Y Multi-Element XRF	12344	A61231 Bulk Mode	4	Next Certification Due: 6/19/2013
BTHWYN	Thermo XLt792Y Multi-Element XRF	12345	940847 Bulk Mode	4	Next Cal Due: 3/16/2013
BTHWYN	Thermo XLt792Y Multi-Element XRF	17286	940641	4	Next Cal Due: 9/22/2012
BTHWYN	Thermo XL3t Multi-Element XRF	39671	C14427	4	
WV	Thermo XLt 792Y Accessory Kit	940578	940578	4	@ Wheeling
BTHWYN	Thermo XLt 792Y Accessory Kit	940579	940579	4	
BTHWYN	Thermo XLt 792Y Accessory Kit	940643	940643	4	w/ J. Banks
BTHWYN	Thermo Portable Test Stand (XRF)	940765	940765	4	w/ J. Banks
BTHWYN	Thermo Portable Freestanding Test Stand (XRF)	940642	940642	4	w/ J. Banks
BTHWYN	OhioLumex RA-91 5+ Mercury Vapor Analyzer	260	940219	4	Next Cal Due: 6/15/2013
WV	OhioLumex RA-91 5+ Mercury Vapor Analyzer	494	A41642	4	Next Cal Due: 3/15/2013
BTHWYN	OhioLumex RP 91 Mercury Vaporizer	162	940694	4	
BTHWYN	OhioLumex RP 91C Mercury Pyrolysis (SOLID)	077	940692	4	
BTHWYN	OhioLumex RP 91C Mercury Pyrolysis (SOLID)	099	940693	4	

Meterological Equipment					
Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
BTHWYN	Coastal Env Systems Weather Pack-400	1420	A73034	4	@ COOP

Personal Protection & Monitoring					
Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
MD H&S	Casella CEL-231 Sound Level Meter W/Calibrator	2/05921099 2/05920631	940751 / 940752	4	Next Cal Due: 7/23/2013
PHL H&S	Casella CEL-254 Sound Level Meter w/Calibrator	4108110 / 4626100	S38193/ S38194	4	Next Cal Due: 8/23/2013
MD H&S	Casella CEL-420 Noise Dosimeter w/ Calibrator	2/011482 3/02123858	S38191 / S38192	4	Next Cal Due: 6/5/2013
MD H&S	DuPont Level A Suit, FE (LG)	711846	940776 DOM 9/14/01	4	Pressure Test Due 10/14/2011
MD H&S	DuPont Level A Suit, FE (LG)	738650	940777 DOM 9/19/01	4	Pressure Test Due: 12/13/2013 Pass @ 3.2" H <sub>2</sub> O
MD H&S	DuPont Level A Suit, FE (LG)	738652	940771 DOM 8/15/01	4	Pressure Test Due: 12/13/2013 Pass @ 3.6" H <sub>2</sub> O
MD H&S	DuPont Level A Suit, FE (LG)	738653	940774 DOM 8/9/01	4	Pressure Test Due: 12/13/2013 Pass @ 3.6" H <sub>2</sub> O
MD H&S	DuPont Level A Suit, FE (LG)	738654	940770 DOM 8/7/01	4	Pressure Test Due: 12/13/2013 Pass@ 3.6" H <sub>2</sub> O.

# Region III Equipment Condition Codes- Current as of 3/15/2013

Personal Protection & Monitoring					
Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
MD H&S	DuPont Level A Suit, FE (LG)	738655	940772 DOM 9/19/01	4	Pressure Test Due: 12/13/2013 Pass @ 3.2" H <sub>2</sub> O
MD H&S	DuPont Level A Suit, FE (LG)	738656	940773 DOM 8/9/01	4	Pressure Test Due: 12/13/2013 Pass @ 3.6" H <sub>2</sub> O
MD H&S	DuPont Level A Suit, FE (LG)	738657	940775 DOM 8/7/01	4	Next Pressure Test Due: 2/7/2014 Pass @ 3.6" H <sub>2</sub> O
MD H&S	DuPont Level A Suit, FE (XL)	5091416	940780 DOM 3/21/03	4	Next Pressure Test Due: 2/7/2014 Pass @ 3.4" H <sub>2</sub> O
MD H&S	DuPont Level A Suit, FE (2X)	147813	940779 DOM 10/10/00	4	Next Pressure Test Due: 2/7/2014 Pass @ 3.4" H <sub>2</sub> O
MD H&S	DuPont Level A Suit, FE (2X)	4959560	940782 DOM 2/20/03	4	Pressure Test Due: 12/13/2013 Pass @ 3.4" H <sub>2</sub> O
MD H&S	DuPont Level A Suit, FE (L)	M789MZQSF	940778 DOM 9/13/05	4	Pressure Test Due: 12/13/2013 Pass @ 3.2" H <sub>2</sub> O
MD H&S	DuPont Level A Suit, FE (L)	M361NO2H8	940785 DOM 12/16/05	4	Pressure Test Due: 12/13/2013 Pass @ 3.8" H <sub>2</sub> O.
MD H&S	DuPont Level A Suit, FE (XL)	M556NODIA	940784 DOM 7/12/06	4	Next Pressure Test Due: 2/7/2014 Pass @ 3.4" H <sub>2</sub> O
MD H&S	DuPont Universal Pressure Test Kit	000792	000792	4	8 each
BTHWYN	Suit, Fully Encapsulating, Level "A", DuPont Tychem TK Commander – Size L	MM055N1XNH EPA SA5707		1	Next Pressure Test Due: 9/20/2013 Change Outlet Valves: 9/20/2014
BTHWYN	Suit, Fully Encapsulating, Level "A", DuPont Tychem TK Commander – Size L	MM055N1XNI EPA SA 5708		1	Next Pressure Test Due: 9/20/2013 Change Outlet Valves: 9/20/2014
BTHWYN	Suit, Fully Encapsulating, Level "A", DuPont Tychem TK Commander – Size XL	MM872N1XCH EPA SA5709		1	Next Pressure Test Due: 9/10/2013 Change Outlet Valves: 9/10/2014
BTHWYN	Suit, Fully Encapsulating, Level "A", DuPont Tychem TK Commander – Size XL	MM872N1XCD EPA SA5710		1	Next Pressure Test Due: 9/7/2013 Change Outlet Valves: 9/7/2014
BTHWYN	Suit, Fully Encapsulating, Level "A", DuPont Tychem TK Commander – Size XL	MM056N1XN5 EPA SA5714		1	Next Pressure Test Due: 9/19/2013 Change Outlet Valves: 9/19/2014
BTHWYN	Suit, Fully Encapsulating, Level "A", DuPont Tychem TK Commander – Size XL	MM872N1XCL EPA SA5711		1	Next Pressure Test Due: 9/11/2013 Change Outlet Valves: 9/11/2014
BTHWYN	Suit, Fully Encapsulating, Level "A", DuPont Tychem TK Commander – Size XL	MM893N1UQ2 EPA SA5712		1	Next Pressure Test Due: 5/17/2013 Change Outlet Valves: 5/17/2014
BTHWYN	Suit, Fully Encapsulating, Level "A", DuPont Tychem TK Commander – Size XL	MM056N1XN4 EPA SA5713		1	Next Pressure Test Due: 9/19/2013 Change Outlet Valves: 9/19/2014
BTHWYN	Suit, Fully Encapsulating, Level "A", DuPont Tychem TK Commander – Size XL	MM825N1X3S EPA SA5715		1	Next Pressure Test Due: 8/30/2013 Change Outlet Valves: 8/30/2014
BTHWYN	Suit, Fully Encapsulating, Level "A", DuPont Tychem TK Commander – Size XL	MM825N1X3R EPA SA5716		1	Next Pressure Test Due: 8/29/2013 Change Outlet Valves: 8/29/2014
BTHWYN	Suit, Fully Encapsulating, Level "A", DuPont Tychem TK Commander – Size XL	MM872N1XCE EPA SA5717		1	Next Pressure Test Due: 9/10/2013 Change Outlet Valves: 9/10/2014
BTHWYN	Suit, Fully Encapsulating, Level "A", DuPont Tychem TK Commander – Size XL	MM825N1X3V EPA SA5718		1	Next Pressure Test Due: 8/31/2013 Change Outlet Valves: 8/31/2014
BTHWYN	Suit, Fully Encapsulating, Level "A", DuPont Tychem TK Commander – Size XL	MM825N1X3T EPA SA5719		1	Next Pressure Test Due: 8/30/2013 Change Outlet Valves: 8/30/2014
BTHWYN	Suit, Fully Encapsulating, Level "A", DuPont Tychem TK Commander – Size 2XL	MM053N1XMA EPA SA5720		1	Next Pressure Test Due: 9/20/2013 Change Outlet Valves: 9/20/2014
BTHWYN	Suit, Fully Encapsulating, Level "A", DuPont Tychem TK Commander – Size 2XL	MM053N1XMB EPA SA5721		1	Next Pressure Test Due: 9/9/2013 Change Outlet Valves: 9/9/2014
BTHWYN	Suit, Fully Encapsulating, Level "A", DuPont Tychem TK Commander – Size 2XL	MM053N1XMC EPA SA5722		1	Next Pressure Test Due: 9/20/2013 Change Outlet Valves: 9/20/2014

# Region III Equipment Condition Codes- Current as of 3/15/2013

## Personal Protection & Monitoring

Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
BTHWYN	Suit, Fully Encapsulating, Level "A", DuPont Tychem TK Commander – Size 2XL	MM053N1XMD EPA SA5723		1	Next Pressure Test Due: 9/18/2013 Change Outlet Valves: 9/18/2014
BTHWYN	Suit, Fully Encapsulating, Level "A", DuPont Tychem TK Commander – Size 2XL	MM053N1XME EPA SA5724		1	Next Pressure Test Due: 9/20/2013 Change Outlet Valves: 9/20/2014
WV	Suit, Fully Encapsulating, Level "A", DuPont Tychem TK Commander – Size 2XL	MM053N1XMF EPA SA5705	Change Outlet Valves: 9/20/2014	1	Next Pressure Test Due: 2/21/2014 Pass @ 3.5" H <sub>2</sub> O
BTHWYN	Suit, Fully Encapsulating, Level "A", DuPont Tychem TK Commander – Size 2XL	MM053N1XMG EPA SA5725		1	Next Pressure Test Due: 9/20/2013 Change Outlet Valves: 9/20/2014
BTHWYN	Suit, Fully Encapsulating, Level "A", DuPont Tychem TK Commander – Size 2XL	MM053N1XMH EPA SA5704		1	Next Pressure Test Due: 9/18/2013 Change Outlet Valves: 9/18/2014
BTHWYN	Suit, Fully Encapsulating, Level "A", DuPont Tychem TK Commander – Size 2XL	MM053N1XMI EPA SA5726		1	Next Pressure Test Due: 9/18/2013 Change Outlet Valves: 9/18/2014
BTHWYN	Suit, Fully Encapsulating, Level "A", DuPont Tychem TK Commander – Size 2XL	MM053N1XMJ EPA SA5727		1	Next Pressure Test Due: 9/20/2013 Change Outlet Valves: 9/20/2014
FM	Suit, Fully Encapsulating, Level "A", DuPont Tychem TK Commander – Size 2XL	MM053N1XMK EPA SA5728	Change Outlet Valves: 9/10/2014	4	Next Pressure Test Due: 2/7/2014 Pass @ 3.7" H <sub>2</sub> O
WV	Suit, Fully Encapsulating, Level "A", DuPont Tychem TK Commander – Size XL	MM482N1WVW EPA SA5706	Change Outlet Valves: 8/21/2014	1	Next Pressure Test Due: 2/21/2014 Pass @ 3.4" H <sub>2</sub> O
BTHWYN	Suit, Fully Encapsulating, Level "A", DuPont Tychem TK Commander – Size 2XL	MM869N1XDZ EPA SA5729	Change Outlet Valves: 9/10/2014	4	Next Pressure Test Due: 9/10/2013
WV	Suit, Fully Encapsulating, Level "A", DuPont Tychem TK Commander – Size MD	MM415N1Y0B EPA SA5730	Change Outlet Valves: 10/9/2014	4	Next Pressure Test Due: 2/21/2014 Pass @ 3.4" H <sub>2</sub> O
BTHWYN	Suit, Fully Encapsulating, Level "A", DuPont Tychem TK Commander – Size MD	MM415N1Y0C EPA SA5731	Change Outlet Valves: 10/9/2014	4	Next Pressure Test Due: 2/7/2014 Pass @ 3.6" H <sub>2</sub> O
MD	Kappler Encapsulated CSM Responder (XL)	263124	940789 DOM 3/3/03	4	PRESSURE TEST DUE 10/28/10
BTHWYN	Kappler Level A Pressure Tester	262421	940316	4	
BTHWYN	Kappler Level A Pressure Tester	262430	940317	4	
WV	Kappler Level A Pressure Tester	262424	940315	4	
BTHWYN	DuPont Level A Pressure Test Kit w/Pump	KNE1 MM0002507MKN E1	940812	4	
MD H&S	Metrosonics HS3700 Heat Stress Data Logger	1579	940684	4	Next Cal Due: 8/9/2013
MD H&S	Quest QT-1 1 Personal Heat Stress Monitor	JU1010003	940683	4	Next Cal Due: 8/9/2013

## Radiation Monitors

Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
MD	BNC SAM 940-2G (NaI) w/ Model 6997 Detector	40210 / 42145	940787 / 940788	4	Next Cal Due: 2/22/2013 OOS for cal
BTHWYN	BNC SAM 940(3G)w/ Model 6998 Detector w/ GPS	40208 / 44117 / 13006331	940798 / 940799 / 940805	4	Next Cal Due: 6/11/2013
BTHWYN	Bicron Micro Rem Radiation Survey Meter	B783W	940326	4	Next Cal Due: 7/11/2013
WV	Exploranium GR-1 30 miniSPEC	9679	940320	4	Next Cal Due: 5/10/2013
WV	Ludlum Model 3 w/43-90 Probe	151403 / PR154916	940338 / 940344	4	Next Cal Due: 10/15/2013 w/ Kit WVRAD001
WV	Ludlum Model 3 w/ 44-6/44-9 Probes	151793 / 154286 / 154495	940337 / 940343 / 940342	4	Next Cal Due: 10/12/2013 w/ Kit WVRAD001
BTHWYN	Ludlum Model 3 w/44-6 Probe	13100 / PRO29793	608310 / 940680	4	Next Cal Due: 7/11/2013
MD H&S	Ludlum Model 3 w/44-6 Probe	13147 / RN013165	S96060	4	Next Cal Due: 9/19/2013
MD H&S	Ludlum Model 3A w/44-6 Probe	134568 / PR135501	S96019	4	Next Cal Due: 9/24/2013

# Region III Equipment Condition Codes- Current as of 3/15/2013

Radiation Monitors					
Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
PHL H&S	Ludlum Model 4 w/44-9 Probe	172075 / PR206197	940423 / 940424	4	Next Cal Due: 12/5/2013
MD H&S	Ludlum Model 4 w/44-9 Probe	172081 / PR192762	S96010 / 940688	4	Next Cal Due: 7/18/2013
MD H&S	Ludlum Model 4 w/44-9 Probe	173210 / PR192761	S96064 / 940690	4	Next Cal Due: 8/6/2013
MD H&S	Ludlum Model 4 w/44-9 Probe	149290 / PR060412	S96020 / 940757	4	Next Cal Due: 9/26/2013
BTHWYN	Ludlum Model 9 ION Chamber	141210	940323	4	Next Cal Due: 11/5/2013
BTHWYN	Ludlum Model 12 w/43-2 Scintillator	46457 / PR032794	623375 / 940762	4	Next Cal Due: 7/10/2013
BTHWYN	Ludlum Model 12 w/ 44-9 G-M Pancake Probe	27198 / PR015861	940535 / 940425	4	Next Cal Due: 7/10/2013
WV	Ludlum Model 14C 44-2/44-9 Probes	152850 / 155298 / 155581	940339 / 940340 / 940341	4	Next Cal Due: 10/12/2013 w/ Kit WVRAD001
BTHWYN	Ludlum Model 14C w/ 44-6 Sidewall Beta / GM	30372 / PR032793	940322 / 940436	4	Next Cal Due: 5/10/2013
BTHWYN	Ludlum Model 14C w/ 44-6 Sidewall Beta / GM	45910 / PR015872	623376 / 940437	4	Next Cal Due: 7/11/2013
WV H&S	Ludlum Model 14C w/44-9 Probe	266611/PR287266	S38156	4	Next Cal Due: 7/11/2013
BTHWYN	Ludlum Model 15 w/ 44-7 GM probe	187682 / PR198139	940321 / 940837	4	Next Cal Due: 9/28/2013
BTHWYN	Ludlum Model 19 MicroR Meter	163284	940966	4	Next Cal Due: 5/16/2013
BTHWYN	Ludlum Model 19 MicroR Meter	45498	623374	4	Next Cal Due: 5/16/2013
BTHWYN	Ludlum Model 19 MicroR Meter	76258	898576	4	Next Cal Due: 8/9/2013
BTHWYN	Ludlum Model 19 MicroR Meter w/ 2241-3 Kit	98065	942714	4	Next Cal Due: 7/12/2013
WV	Ludlum Model 19 MicroR Meter	23073	940319	4	Next Cal Due: 5/16/2013
WV	Ludlum Model 19 MicroR Meter	148172	940336	4	Next Cal Due: 10/17/2013 w/ Kit WVRAD001
BTHWYN	Ludlum Model 192 MicroR Meter	191194	940220	4	Next Cal Due: 5/16/2013
MD	Ludlum Model 192 MicroR Meter w/2241-2 kit	191223	940332	4	Next Cal Due: 9/21/2013
BTHWYN	Ludlum Model 500 Pulser	152056	940420 Keep @ COOP	4	Next Cal Due: 9/26/2013
BTHWYN	Ludlum Model 2241-2 w/44-9 / 43-90 Probes	189792 / 199428 / 197296	940328 / 940330 / 940329	4	Next Cal Due: 5/21/2013
MD	Ludlum Model 2241-2 w/44-9 / 43-90 Probes	189742 / 199383 / 195065	940333 / 940334 / 940335	4	Next Cal Due: 5/21/2013
BTHWYN	Ludlum Model 2241-3 w/44-9 / 43-68 / 43-2 Probes	143886 / 32795 / 148803 / 15843	940324 / 940679 / 940630 / 940678	4	Next Cal Due: 7/12/2013
MD H&S	Ludlum Model 2360 w/ 43-93 Probe	268413 / PR286299	S96065	4	Next Cal Due: 7/18/2013
BTHWYN	Ludlum Model 3030 Alpha/Beta Scaler	187982	940318	4	Next Cal Due: 1/11/2014
BTHWYN	SE Int'l Inspector Exp+ w/GM Probe	21663 / 01927	940488 / 940676	4	Next Cal Due: 5/10/2013
WV	SE Int'l Inspector Exp+ w/GM Probe	21661 / 01925	940593 / 940438	4	Next Cal Due: 8/3/2013

# Region III Equipment Condition Codes- Current as of 3/15/2013

Radiation Monitors					
Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
MD	SE Int'l Inspector Exp+ w/GM Probe	21666 / 01930	940666 / 940667	4	Next Cal Due: 6/1/2013
BTHWYN	Teletrix SP418 Series Probe PAK w/4 remotes	P4-1147	940970	4	New 5/17/2010
BTHWYN	Teletrix SP418 Series Probe PAK w/4 remotes	P4-1148	940969	4	New 5/17/2010
BTHWYN	Teletrix SD418 Series Simulated Alarming Dosimeter w/ 1 remote	073	940973	4	New 5/17/2010
BTHWYN	Teletrix SD418 Series Simulated Alarming Dosimeter w/ 1 remote	081	940972	4	New 5/17/2010
BTHWYN	Siemens MK2 Electronic Personal Detector	42279	940558	4	Next Cal Due: 6/7/2013
BTHWYN	Siemens MK2 Electronic Personal Detector	42284	940452	7	Next Cal Due: 2/25/2014
BTHWYN	Siemens MK2 Electronic Personal Detector	42291	A41637 Towle	4	Next Cal Due: 2/21/2013
BTHWYN	Siemens MK2 Electronic Personal Detector	42297	A41636	7	Due for Calibration
BTHWYN	Siemens MK2 Electronic Personal Detector	42302	940453	7	Next Cal Due:
BTHWYN	Siemens MK2 Electronic Personal Detector	42308	940276 Unassigned	4	Next Cal Due: 8/3/2013
BTHWYN	Siemens MK2 Electronic Personal Detector	42337	940434 T. Richardson	4	Next Cal Due: 2/15/2012
BTHWYN	Siemens MK2 Electronic Personal Detector	42339	940435 Unassigned	4	Next Cal Due: 2/21/2013
BTHWYN	Siemens MK2 Electronic Personal Detector	42351	A41632 Jarvela	4	Next Cal Due: 2/15/2012
BTHWYN	Siemens MK2 Electronic Personal Detector	42384	940418 Fetzer	4	Cal Due 5/23/2011
BTHWYN	Siemens MK2 Electronic Personal Detector	42394	940406 DiDonato	4	Next Cal Due: 2/21/2013
BTHWYN	Siemens MK2 Electronic Personal Detector	42395	940455 J. Kelly	4	Next Cal Due: 2/15/2012
BTHWYN	Siemens MK2 Electronic Personal Detector	42406	A41640 R. Rupert	4	Next Cal Due: 2/21/2013
BTHWYN	Siemens MK2 Electronic Personal Detector	42409	940456	7	Due for Calibration
BTHWYN	Siemens MK2 Electronic Personal Detector	42412	940410 C. Wagner	4	Next Cal Due: 2/21/2013
BTHWYN	Siemens MK2 Electronic Personal Detector	42421	A41633 Unassigned	4	Next Cal Due: 2/17/2013
BTHWYN	Siemens MK2 Electronic Personal Detector	42669	940451 E. Rovira	4	Next Cal Due: 3/1/2012
BTHWYN	Siemens MK2 Electronic Personal Detector	42748	A41634 L. Casillas	4	Next Cal Due: 2/15/2012
BTHWYN	Siemens MK2 Electronic Personal Detector	42761	940408	7	Due for Calibration
BTHWYN	Siemens MK2 Electronic Personal Detector	42816	940419 D. Ventura	4	Next Cal Due: 6/16/2013
BTHWYN	Siemens MK2 Electronic Personal Detector	42836	940457 D. McLaughlin	4	Next Cal Due: 5/11/2013
BTHWYN	Siemens MK2 Electronic Personal Detector	42892	940412 R. Scharr	4	Next Cal Due: 5/11/2013
BTHWYN	Siemens MK2 Electronic Personal Detector	42901	940459 Unassigned	7	Next Cal Due: 2/21/2013
BTHWYN	Siemens MK2 Electronic Personal Detector	43673	940675 Zenone	4	Next Cal Due: 2/21/2013
PHL H&S	Siemens MK2 Electronic Personal Detector	00065032	940617	4	Next Cal Due: 3/26/2013 OOS for cal
PHL H&S	Siemens MK2 Electronic Personal Detector	00065367	940616	4	Next Cal Due: 3/26/2013 OOS for cal

# Region III Equipment Condition Codes- Current as of 3/15/2013

Radiation Monitors					
Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
PHL H&S	Siemens MK2 Electronic Personal Detector	00065483	940618	4	Next Cal Due: 3/26/2013 OOS for cal
PHL H&S	Siemens Mk2 Electronic Personal Detector	00066002	940753	4	Next Cal Due: 3/26/2013 OOS for cal
WV	Siemens Mk2 Electronic Personal Detector	42308	940276	4	Next Cal Due: 7/22/2011
WV	Siemens Mk2 Electronic Personal Detector	42404	940411 Matlock	4	Next Cal Due: 9/17/2013
WV	Siemens Mk2 Electronic Personal Detector	42430	940454 Lindsey	4	Next Cal Due: 7/6/2013
BTHWYN	Siemens Mk2 Electronic Personal Detector	42813	940407 Unassigned	4	Next Cal Due: 8/3/2013
MD H&S	Siemens Mk2 Electronic Personal Detector	66283	S96012	4	Next Cal Due: 7/23/2013
MD H&S	Siemens Mk2 Electronic Personal Detector	65434	S96015	4	Next Cal Due: 7/23/2013
MD H&S	Siemens Mk2 Electronic Personal Detector	209272	S96013	4	Next Cal Due: 7/23/2013
MD H&S	Siemens Mk2 Electronic Personal Detector	209378	S96014	4	Next Cal Due: 7/23/2013
PHL H&S	Victoreen Thyac III w/489-110C G-M Model 490	4849M / Probe S/N 511	313587 / 720208	4	Next Cal Due: 8/6/2013
PHL H&S	Victoreen Thyac III w/489-4/489-110 Model 490	4538M / Probes S/N 4116,856	197083 / 940421 / 940422	4	Next Cal Due: 10/24/2013

Respiratory Equipment					
Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
BTHWYN	Biomarine BioPak 240 Closed Circuit SCBA	3933	940603	4	Next service due: 1/19/2014
BTHWYN	Biomarine BioPak 240 Closed Circuit SCBA	3934	940605	4	Next service due: 2/1/2014
BTHWYN	Biomarine BioPak 240 Closed Circuit SCBA	3935	940602	4	Next service due: 12/15/2013
BTHWYN	Biomarine BioPak 240 Closed Circuit SCBA	3936	940601	4	Next service due: 12/15/2013
BTHWYN	Biomarine BioPak 240 Closed Circuit SCBA	3938	940600	4	Next service due: 2/1/2014
BTHWYN	Biomarine BioPak 240 Closed Circuit SCBA	3939	940604	4	Next service due: 1/19/2014
BTHWYN	Biomarine BioPak 240 Closed Circuit SCBA	3942	940609	4	Next service due: 12/15/2013
BTHWYN	Biomarine BioPak 240 Closed Circuit SCBA	3963	940608	4	Next service due: 1/19/2014
BTHWYN	Biomarine BioPak 240 Closed Circuit SCBA	3964	940607	4	Next service due: 12/15/2013
BTHWYN	Biomarine BioPak 240 Closed Circuit SCBA	4084	940606	4	Next service due: 2/1/2014
BTHWYN	Biomarine BioPak 240 Oxygen Cylinder	6163-1802	940907	4	H.T. Due 5/31/2013
BTHWYN	Biomarine BioPak 240 Oxygen Cylinder	6163-1814	940904	4	H.T. Due 5/31/2013
BTHWYN	Biomarine BioPak 240 Oxygen Cylinder	6163-1914	940909	4	H.T. Due 5/31/2013
BTHWYN	Biomarine BioPak 240 Oxygen Cylinder	6163-1918	940903	4	H.T. Due 5/31/2013
BTHWYN	Biomarine BioPak 240 Oxygen Cylinder	6163-1920	940905	4	Next Hydro Test Due: 11/2016.
BTHWYN	Biomarine BioPak 240 Oxygen Cylinder	6163-1940	940902	4	H.T. Due 5/31/2013



# Region III Equipment Condition Codes- Current as of 3/15/2013

BTHWYN	Biomarine BioPak 240 Oxygen Cylinder	6163-2327	940906	4	H.T. Due 5/31/2013
BTHWYN	Biomarine BioPak 240 Oxygen Cylinder	6163-2331	940915 in 3936	4	Next Hydro Test Due: 11/2016.
<b>Respiratory Equipment</b>					
Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
BTHWYN	Biomarine BioPak 240 Oxygen Cylinder	6163-2336	940916 in 3933	4	Next Hydro Test Due: 11/2016.
BTHWYN	Biomarine BioPak 240 Oxygen Cylinder	6163-2347	940910 in 3939	4	Next Hydro Test Due: 11/2016.
BTHWYN	Biomarine BioPak 240 Oxygen Cylinder	6163-2350	940912 in 3934	4	Next Hydro Test Due: 11/2016.
BTHWYN	Biomarine BioPak 240 Oxygen Cylinder	6163-2353	940917 in 3964	4	Next Hydro Test Due: 11/2016.
BTHWYN	Biomarine BioPak 240 Oxygen Cylinder	6163-2355	940919 in 3942	4	Next Hydro Test Due: 11/2016.
BTHWYN	Biomarine BioPak 240 Oxygen Cylinder	6163-2357	940914 in 3935	4	Next Hydro Test Due: 11/2016.
BTHWYN	Biomarine BioPak 240 Oxygen Cylinder	6163-2377	940918 in 3963	4	Next Hydro Test Due: 11/2016.
BTHWYN	Biomarine BioPak 240 Oxygen Cylinder	6163-2425	940908	4	Next Hydro Test Due: 11/2016.
BTHWYN	Biomarine BioPak 240 Oxygen Cylinder	6163-2446	940913 in 4084	4	Next Hydro Test Due: 11/2016.
BTHWYN	Biomarine BioPak 240 Oxygen Cylinder	6163-2454	940911 in 3938	4	Next Hydro Test Due: 11/2016.
BTHWYN	Mask, SCBA, AV-2000, (LG)			4	6 each
BTHWYN	Mask, SCBA, AV-3000, (MD))			4	0 each
BTHWYN	Facepiece, AV-3000, Scott (SM)			1	New 3/2011 - 6 each
BTHWYN	Facepiece, AV-3000, Scott (MD)			1	New 3/2011 - 16 each
BTHWYN	Facepiece, AV-3000, Scott (LG)			1	New 3/2011 - 5 each
BTHWYN	SCOTT FACEPIECE AV2000 Kevlar EPDM (MD)			4	6 each
BTHWYN	SCOTT FACEPIECE AV2000 Kevlar EPDM (LG)			4	1 each
BTHWYN	SCOTT FACEPIECE AV3000 Kevlar Poly (SM)			4	5 each
BTHWYN	SCOTT FACEPIECE AV3000 Kevlar Poly (MD)			4	5 each
BTHWYN	SCOTT FACEPIECE AV3000 Kevlar Poly (LG)			4	1 each
BTHWYN	SCOTT FACEPIECE AV3000, Sure Seal (M)		940994	4	Unassigned
BTHWYN	SCOTT FACEPIECE AV3000, Sure Seal (M)		940995	4	Unassigned
BTHWYN	SCOTT FACEPIECE AV3000, Sure Seal (M)		940996	4	Unassigned
BTHWYN	SCOTT FACEPIECE AV3000, Sure Seal (M)		940997	4	Unassigned
BTHWYN	SCOTT FACEPIECE AV3000, Sure Seal (L)			4	2 each
BTHWYN	SCOTT FACEPIECE AV3000, Sure Seal (L)			4	1 Assigned to C. Wagner
BTHWYN	SCOTT FACEPIECE AV3000, Sure Seal (L)			4	1 – Assigned to M. Towle
BTHWYN	Mask, UltraElite, SCBA, HyCar, (MD)			4	3 each



# Region III Equipment Condition Codes- Current as of 3/15/2013

Respiratory Equipment					
Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
BTHWYN	Mask, UltraElite, SCBA, HyCar, (SM)			4	1 each Assigned to A. DiDonato
BTHWYN	Mask, UltraElite, SCBA, CBRN (SM)			4	1 each Assigned to A. DiDonato
BTHWYN	Mask, UltraElite, APR, Silicone, (SM)			4	1 each
BTHWYN	Mask, UltraElite, APR, HyCar, (SM)			4	1 each Assigned to A. DiDonato
BTHWYN	Mask, UltraElite, APR, Silicone, (MD)			4	2 each
BTHWYN	Mask, Ultratwin, APR, (SM)			4	1 each
BTHWYN	Mask, Ultratwin, APR, (MD)			4	3 each
BTHWYN	Biosystems POSI-Chek III	LO4659	LO4659	4	
BTHWYN	ISI Emergency Escape Pack	37094	940595 @ COOP	5	Next Hydrotest Due: 8/2017
BTHWYN	ISI Emergency Escape Pack	35206	940597 @ COOP	5	Next Hydrotest Due: 8/2017
BTHWYN	ISI Emergency Escape Pack	2291	950594 @ COOP	5	Next Hydrotest Due: 8/2017
BTHWYN	ISI Emergency Escape Pack	36473	940761 @ COOP	5	Hydrotest Due 8/ 2013
BTHWYN	ISI Emergency Escape Pack	36290	940062	0	Next bottle hydro test due: 1/12/2017
WV	ISI Emergency Escape Pack	37329	940362	4	Next bottle hydro test due: 10/1//2014
MD H&S	MSA TransAire 5 Minute Escape Apparatus	J109855	940892	1	New 10/09 Rm E202
MD H&S	MSA TransAire 5 Minute Escape Apparatus	J109863	940893	1	New 10/09 Rm E202
MD H&S	MSA TransAire 5 Minute Escape Apparatus	J109820	940894	1	New 10/09 Rm E202
MD H&S	MSA TransAire 10 Minute Escape Apparatus	LT106515	940920	1	New 10/09 Rm E202
MD H&S	MSA TransAire 10 Minute Escape Apparatus	LT106436	940895	1	New 10/09 Rm E202
MD H&S	MSA TransAire 10 Minute Escape Apparatus	LT106468	940921	1	New 10/09 Rm E202
MD H&S	MSA Firehawk SCBA MMR w/ UltraElite (MD)	AOAB241192 / LZ233767	940929 CBRN 10/15/08	1	Flow Test Due 10/15/10
MD H&S	MSA Firehawk SCBA MMR w/ UltraElite (MD)	AOAB241197 / LZ233769	940927 CBRN 10/15/08	1	Flow Test Due 10/15/10
BTHWYN	MSA Firehawk SCBA M7 Air Mask	AOAD033880	940956	1	Next Flow Test Due: 3/13/2013 OOS for Flow Test
BTHWYN	MSA Firehawk SCBA M7 Air Mask	AOAD028669	940955	1	Next Flow Test Due: 4/4/2013
BTHWYN	MSA Firehawk SCBA M7 Air Mask	AOAD033884	940957 A. DiDonato	1	Next Flow Test Due: 3/13/2013 OOS for Flow Test
MD H&S	MSA Firehawk SCBA M7 Air Mask	AOAB276548	940931	1	
MD H&S	MSA Firehawk SCBA M7 Air Mask	AOAB276544	940933	1	
MD H&S	MSA 4500 SCBA MMR w/UltraElite (SM)	HT181176	HT181176	4	
BTHWYN	MSA 4500 SCBA MMR w/UltraElite (MD)	HS258043 URC OK	A34805	4	Next Flow Test Due: 4/4/2013
PHL H&S	MSA 4500 SCBA MMR w/UltraElite (MD)	ZU352148	940477	4	Next Flow Test Due: 8/17/2013

# Region III Equipment Condition Codes- Current as of 3/15/2013

Respiratory Equipment						
Owning Location	Item Description	Serial Number		EPA TAG NUMBER	Condition Code	Notes
BTHWYN	MSA 4500 SCBA MMR w/UltraElite (MD)	ZS243422	HS218019	A34869	4	Next Flow Test Due: 4/4/2013
PHL H&S	MSA 4500 SCBA MMR w/UltraElite (MD)	ZS255237	HS257007	A34808	4	Next Flow Test Due: 8/26/2013
PHL H&S	MSA 4500 SCBA MMR w/UltraElite (MD)	ZS259170	HS199088	A34807	4	Next Flow Test Due: 8/17/2013
BTHWYN	MSA 4500 SCBA MMR w/UltraElite (MD)	ZS259247	HX270191	A34800	4	Next Flow Test Due: 8/9/2012
MD H&S	MSA 4500 SCBA MMR w/UltraElite (MD)	HU352073		HU352073	4	
MD H&S	MSA 4500 SCBA MMR w/UltraElite (MD)	HU352074		940948	4	
MD H&S	MSA 4500 SCBA MMR w/UltraElite (MD)	HU352075		940950	4	
MD H&S	MSA 4500 SCBA MMR w/UltraElite (MD)	HU352084		940945	4	
MD H&S	MSA 4500 SCBA MMR w/UltraElite (MD)	HT177037		HT177037	4	
MD H&S	MSA 4500 SCBA MMR w/UltraElite (MD)	HT177047		940947	4	
PHL H&S	MSA OptimAir 6A PAPR w/Hycar Mask (MD)	BPC427242		940727	7	Needs MSA Cartridges @ COOP
PHL H&S	MSA OptimAir 6A PAPR w/Hycar Mask (MD)	BP6428110		940728	7	Needs MSA Cartridges @ COOP
PHL H&S	MSA OptimAir 6A PAPR w/Hycar Mask (MD)	BNJ426373		940730	7	Needs MSA Cartridges @ COOP
PHL H&S	MSA OptimAir 6A PAPR No Mask	BP6428217		940729	7	Needs MSA Cartridges @ COOP
MD H&S	MSA OptimAir 6A PAPR w/Hycar Mask (MD)	400536		400536	7	
MD H&S	MSA OptimAir 6A PAPR w/Hycar Mask (MD)	400537		400537	7	
MD H&S	MSA OptimAir 6A PAPR w/Hycar Mask (MD)	400538		400538	7	
MD H&S	MSA OptimAir 6A PAPR w/Hycar Mask (MD)	400539		400539	7	
MD H&S	MSA OptimAir 6A PAPR w/Hycar Mask (MD)	400540		400540	7	
MD H&S	MSA OptimAir 6A PAPR w/Hycar Mask (MD)	400541		400541	7	
MD H&S	MSA OptimAir MM2K PAPR w/Hycar (MD)	400542		400542	7	
MD H&S	MSA OptimAir MM2K PAPR w/Hycar (MD)	400543		400543	7	
BTHWYN	MSA OptimAir MM2K PAPR w/Silicone (MD)	N001010		940510	5	
BTHWYN	MSA OptimAir MM2K PAPR w/Silicone (MD)	N001009-1		940509	5	
BTHWYN	MSA OptimAir MM2K PAPR w/Silicone (MD)	N001009-2		940511	5	
BTHWYN	MSA OptimAir MM2K PAPR w/Silicone (MD)	N003066-1		940507	5	
BTHWYN	MSA OptimAir MM2K PAPR w/Silicone (MD)	K002023		940508	5	

# Region III Equipment Condition Codes- Current as of 3/15/2013

Respiratory Equipment					
Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
WV	MSA OptimAir MM2K PAPR w/Silicone (MD)	N001009-3	940513	7	Need Cartridge
WV	MSA OptimAir MM2K PAPR w/Silicone (MD)	K002024	940514	7	Need Cartridge
MD H&S	MSA OptimAir MM2K PAPR w/Silicone (MD)	400551	400551	7	
MD H&S	MSA OptimAir MM2K PAPR w/Silicone (MD)	400552	400552	7	
BTHWYN	MSA OptimAir MM2K PAPR w/Silicone (LG)	N003066-2	940512	5	
MD H&S	MSA OptimAir MM2K PAPR w/Silicone (LG)	400554	400554	7	
MD H&S	MSA OptimAir MM2K PAPR w/Silicone (LG)	400555	400555	7	
MD H&S	MSA RescueAire II Portable Air Supply	LU232046	LU232046	4	Flow Test Due 10/15/10
MD H&S	MSA RescueAire II Portable Air Supply	LU267068	LU267068	4	Flow Test Due 10/15/10
MD H&S	MSA RescueAire II Portable Air Supply	LU267069	LU267069	4	Flow Test Due 10/15/10
MD H&S	MSA RescueAire II Portable Air Supply	LU267072	LU267072	4	Flow Test Due 10/15/10
MD H&S	MSA SCBA WorkMask BMR w/ UltraElite Mask	NR090181	NR090181	4	
MD H&S	MSA SCBA WorkMask BMR w/ UltraElite Mask	NR091263	NR091263	4	
BTHWYN	MSA Stealth H-45 Cylinder 45 min 4500 psi	OM91778	940954	4	H.T. due 6/2014      New 3/10
BTHWYN	MSA Stealth H-45 Cylinder 45 min 4500 psi	OM91552	940958 G. Ham	4	H.T. due 6/2014      New 3/10
BTHWYN	MSA Stealth H-45 Cylinder 45 min 4500 psi	OM91117	940959	4	H.T. due 6/2014      New 3/10
PHL H&S	MSA Stealth H60 Cylinder 60 min 4500 psi	ON11240	940359	4	H.T. due 10/2014      w/A34808
PHL H&S	MSA Stealth H60 Cylinder 60 min 4500 psi	ON4683	940361	4	w/ pack A34800
PHL H&S	MSA Stealth H60 Cylinder 60 min 4500 psi	ON4879	940360	4	H.T. Due 4/2015
PHL H&S	MSA Stealth H60 Cylinder 60 min 4500 psi	ON4883	940523	4	Next HT Due: 3/2017
PHL H&S	MSA Stealth H60 Cylinder 60 min 4500 psi	ON4899	940417	4	H.T. due 10/2014 @ COOP
PHL H&S	MSA Stealth H60 Cylinder 60 min 4500 psi	ALT 604-8125	940358	4	H.T. due 10/2014 @ COOP
PHL H&S	MSA Stealth H60 Cylinder 60 min 4500 psi	ALT 604-8450	940357	4	Ham
PHL H&S	MSA Stealth H60 Cylinder 60 min 4500 psi	ALT 604-8455	940698	4	Returned from D.Fox H.T. Due 5-2013
PHL H&S	MSA Stealth H60 Cylinder 60 min 4500 psi	ALT 604-8469	940699	4	Next Hydro Due – 8/2015
PHL H&S	MSA Stealth H60 Cylinder 60 min 4500 psi	ALT 604-8471	940404	4	H.T. due 10/2014 @ COOP
MD H&S	MSA Stealth H60 Cylinder 60 min 4500 psi	ON10462	940924	4	H.T Due 12-2011
MD H&S	MSA Stealth H60 Cylinder 60 min 4500 psi	ON10475	ON10475	4	H.T Due 12-2011
MD H&S	MSA Stealth H60 Cylinder 60 min 4500 psi	ON10488	940925	4	H.T Due 12-2011
MD H&S	MSA Stealth H60 Cylinder 60 min 4500 psi	ON10500	940944	4	H.T Due 12-2011
MD H&S	MSA Stealth H60 Cylinder 60 min 4500 psi	ALT-604-10777	940943	4	HT Due 2-2013

# Region III Equipment Condition Codes- Current as of 3/15/2013

Respiratory Equipment					
Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
MD H&S	MSA Stealth H60 Cylinder 60 min 4500 psi	ON11287	ON1 1287	4	H.T Due 12-2011
MD H&S	MSA Stealth H60 Cylinder 60 min 4500 psi	ON11296	ON11296	4	H.T Due 12-2011
MD H&S	MSA Stealth H60 Cylinder 60 min 4500 psi	ON11314	ON11314	4	H.T Due 12-2011
MD H&S	MSA Stealth H60 Cylinder 60 min 4500 psi	ON11317	940928	4	H.T Due 12-2011
MD H&S	MSA Stealth H60 Cylinder 60 min 4500 psi	ON11342	940932	4	H.T Due 12-2011
MD H&S	MSA Stealth H60 Cylinder 60 min 4500 psi	ON11367	940940	4	H.T Due 12-2011
MD H&S	MSA Stealth H60 Cylinder 60 min 4500 psi	ON11381	940926	4	H.T Due 12-2011
MD H&S	MSA Stealth H60 Cylinder 60 min 4500 psi	ON11387	940942	4	H.T Due 12-2011
MD H&S	MSA Stealth H60 Cylinder 60 min 4500 psi	ON4811	940930	4	HT Due 2-2013
MD H&S	MSA Stealth H60 Cylinder 60 min 4500 psi	ON5784	940949	4	HT Due 2-2013
MD H&S	MSA Stealth H60 Cylinder 60 min 4500 psi	ON6555	940923	4	HT Due 2-2013
MD H&S	MSA Stealth H60 Cylinder 60 min 4500 psi	ON6598	ON6598	4	HT Due 2-2013
MD H&S	MSA Stealth H60 Cylinder 60 min 4500 psi	ON6819	940941	4	HT Due 8-2010
MD H&S	MSA Stealth H60 Cylinder 60 min 4500 psi	ON6832	940939	4	HT Due 4-2013
MD H&S	MSA Stealth H60 Cylinder 60 min 4500 psi	ON6900	940922	4	HT Due 4-2013
MD H&S	MSA Stealth H60 Cylinder 60 min 4500 psi	ALT-604-6920	ALT-604-6920	4	HT Due 2-2013
MD H&S	MSA Stealth H60 Cylinder 60 min 4500 psi	ALT-604-6948	940951	4	HT Due 2-2013
MD H&S	MSA Stealth H60 Cylinder 60 min 4500 psi	ON6953	940938	4	HT Due 42013
MD H&S	MSA Stealth H60 Cylinder 60 min 4500 psi	ON7054	940936	4	HT Due 4-2013
MD H&S	MSA Stealth H60 Cylinder 60 min 4500 psi	ON7111	940937	4	HT Due 4-2013
MD H&S	MSA Stealth H60 Cylinder 60 min 4500 psi	ALT-604-8462	940946	4	H.T. Due 2-2013
BTHWYN	Scott AirPak Fifty 4.5 MMR SCBA	0403000138AA NJ021803245SM2	A27983 D. Ventura	4	Next Flow Test Due: 4/4/2013
BTHWYN	Scott AirPak Fifty 4.5 MMR SCBA	0403000183AA NJ021804445SM1	A41719 Sharma	4	Next Flow Test Due: 4/4/2013
BTHWYN	Scott AirPak Fifty 4.5 MMR SCBA	0403000186AA NJ021800745SM1	A41729 McLaughlin	4	Next Flow Test Due: 4/4/2013
BTHWYN	Scott AirPak Fifty 4.5 MMR SCBA	0403000192AA NJ020406345SM1	A27967 Unassigned	4	Next Flow Test Due: 8/17/2013
BTHWYN	Scott AirPak Fifty 4.5 MMR SCBA	0403000200AA NJ021104045SM1	A41683 M. Bartos	4	Next Flow Test Due: 4/4/2013
BTHWYN	Scott AirPak Fifty 4.5 MMR SCBA	0403000201AA NJ021005345SM1	A41673 Unassigned	4	Next Flow Test Due: 4/4/2013
BTHWYN	Scott AirPak Fifty 4.5 MMR SCBA	0403000208AA NJ020404245SM2	A27980 G. Ham	4	Next Flow Test Due: 4/4/2013
BTHWYN	Scott AirPak Fifty 4.5 MMR SCBA	0403000235AA NJ021803045SM2	A41688 J. Kelly	4	Next Flow Test Due: 4/4/2013
BTHWYN	Scott AirPak Fifty 4.5 MMR SCBA	040300256AA	A41664 R. Fetzer	4	Flow Test Due: 4/6/2012
BTHWYN	Scott AirPak Fifty 4.5 MMR SCBA	115S0904000095 NJ02044645SM2	A41671 Zenone	4	Next Flow Test Due: 4/4/2013

# Region III Equipment Condition Codes- Current as of 3/15/2013

## Respiratory Equipment

Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
BTHWYN	Scott AirPak Fifty 4.5 MMR SCBA	0403000267AA	A41698 Jarvela	4	Next Flow Test Due: 2/8/2014
BTHWYN	Scott AirPak Fifty 4.5 MMR SCBA	NJ02180465SM2 0403000198AA	A27977	4	Next Flow Test Due: 2/8/2014
BTHWYN	Scott AirPak Fifty 4.5 MMR SCBA	NJ021801945SM1 0403000271	A73008 G. Lapsley	4	Flow Test Due 4/6/2012
BTHWYN	Scott AirPak Fifty 4.5 MMR SCBA	NJ021802445SM2 0403000285	A41724 T. Richardson	4	Next Flow Test Due: 4/4/2013
BTHWYN	Scott AirPak Fifty 4.5 MMR SCBA	NH020203745SM2 0403000282	A73084 C. Wagner	4	Next Flow Test Due: 4/4/2013
BTHWYN	Scott AirPak Fifty 4.5 MMR SCBA	NJ020404245SM1	A41703 M. Towle	4	Next Flow Test Due: 4/4/2013
BTHWYN	Scott AirPak Fifty 4.5 MMR SCBA	NJ020404945SM1	A41679 Unassigned	4	Next Flow Test Due: 2/8/2014
BTHWYN	Scott AirPak Fifty 4.5 MMR SCBA	NH020205445SM2	A73005 Rovira	4	Next Flow Test Due: 4/4/2013
BTHWYN	Scott AirPak Fifty 4.5 MMR SCBA	NJ020405745SM1 0403000209AA	A27972 R. Scharr	4	Next Flow Test Due: 1/9/2014
BTHWYN	Scott AirPak Fifty 4.5 MMR SCBA	NJ020906345SM2	A41693 L. Casillas	4	Next Flow Test Due: 4/12/2013
BTHWYN	Scott AirPak Fifty 4.5 MMR SCBA	NJ021103245SM1 0403000188AA	A27961 R. Rupert	4	Next Flow Test Due: 2/14/2012
WV	Scott AirPak Fifty 4.5 MMR SCBA	NH020204345SM2 0403000264AA	A41737 Matlock	4	Next Flow Test Due: 4/4/2013
BTHWYN	Scott AirPak Fifty 4.5 MMR SCBA	NJ021104545SM1 0403000265AA	A41709 Santiago	4	Next Flow Test Due: 4/4/2013
WV	Scott AirPak Fifty 4.5 MMR SCBA	NJ020404645SM1 0403000262AA	A41714 Lindsey	4	Next Flow Test Due: 4/4/2013
WV	Scott AirPak Fifty 4.5 MMR SCBA	NJ020406045SM2 0403000260AA	A41732 Linden	4	Next Flow Test Due: 4/4/2013
MD	Scott AirPak Fifty 4.5 MMR SCBA	0407011580AA 0407002379AA	940415 C. Fitzsimmons	4	Next Flow Test Due: 4/4/2013
BTHWYN	Scott C420 PAPR (MD)	02508	A73086	4	
BTHWYN	Scott C420 PAPR (MD)	02516	A27971	4	
BTHWYN	Scott C420 PAPR (MD)	02541	A41670	4	
BTHWYN	Scott C420 PAPR (MD)	02560	A41702	4	
BTHWYN	Scott C420 PAPR (MD)	02573	A41677	4	Ret to COOP by G. HAM 7/14/09
BTHWYN	Scott C420 PAPR (MD)	02575	A41687	4	
BTHWYN	Scott C420 PAPR (MD)	02583	A27976	4	
BTHWYN	Scott C420 PAPR (MD)	02586	A41723	4	
BTHWYN	Scott C420 PAPR (MD)	02590	A41707	4	
BTHWYN	Scott C420 PAPR (MD)	02633	A73003	4	
BTHWYN	Scott C420 PAPR (MD)	02634	A73009	4	
BTHWYN	Scott C420 PAPR (MD)	02671	A27979	4	Assigned to S. Wenning on 1/17/2012.

# Region III Equipment Condition Codes- Current as of 3/15/2013

Respiratory Equipment					
Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
BTHWYN	Scott C420 PAPR (MD)	02685	A27982	4	
BTHWYN	Scott C420 PAPR (MD)	02715	A41697	4	
BTHWYN	Scott C420 PAPR (MD)	02750	A73020	4	Returned to COOP 9/3/09
BTHWYN	Scott C420 PAPR (MD)	02753	A41692	4	
BTHWYN	Scott C420 PAPR (MD)	02798	A73006	4	
BTHWYN	Scott C420 PAPR (MD)	02803	A27957	4	
BTHWYN	Scott C420 PAPR (MD)	02806	A27966	4	
BTHWYN	Scott C420 PAPR (MD)	02807	A41371	4	
BTHWYN	Scott C420 PAPR (MD)	02811	A41728 McLaughlin	4	
BTHWYN	Scott C420 PAPR (MD)	08541	08541	1	
BTHWYN	Scott C420 PAPR (MD)	08544	940800	1	Zickler
WV	Scott C420 PAPR (MD)	02528	A41736	4	MATLOCK
WV	Scott C420 PAPR (MD)	02531	A41713	4	
WV	Scott C420 PAPR (MD)	02602	A41718	4	LINDSEY
WV	Scott C420 PAPR (MD)	02621	A27959	4	
MD	Scott C420 PAPR (MD)	02566	A41682	4	Fitzsimmons
BTHWYN	Scott C420 PLUS w/ M120 CBRN Mask (M/L)	01744	940898	4	New 11/2009
BTHWYN	Scott C420 PLUS w/ M120 CBRN Mask (M/L)	01695	940899	4	New 11/2009
BTHWYN	Scott C420 PLUS w/ M120 CBRN Mask (M/L)	01747	940900	4	New 11/2009
BTHWYN	PAPR, C420 Plus, Scott	3364	SA5512	1	New 3/2011
BTHWYN	PAPR, C420 Plus, Scott	3378	SA5518	1	New 3/2011
BTHWYN	PAPR, C420 Plus, Scott	3417	SA5520	1	New 3/2011
BTHWYN	PAPR, C420 Plus, Scott	3418	SA5519	1	New 3/2011
BTHWYN	PAPR, C420 Plus, Scott	3432	SA5511	1	New 3/2011
BTHWYN	PAPR, C420 Plus, Scott	3433	SA5516	1	New 3/2011
BTHWYN	PAPR, C420 Plus, Scott	3453	SA5524	1	New 3/2011
BTHWYN	PAPR, C420 Plus, Scott	3459	SA5521	1	New 3/2011
BTHWYN	PAPR, C420 Plus, Scott	3464	SA5510	1	New 3/2011
BTHWYN	PAPR, C420 Plus, Scott	3465	SA5509	1	New 3/2011
BTHWYN	PAPR, C420 Plus, Scott	3472	SA5506	1	New 3/2011
BTHWYN	PAPR, C420 Plus, Scott	3476	SA5505	1	New 3/2011

# Region III Equipment Condition Codes- Current as of 3/15/2013

Respiratory Equipment					
Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
BTHWYN	PAPR, C420 Plus, Scott	3490	SA5522	1	New 3/2011
BTHWYN	PAPR, C420 Plus, Scott	3491	SA5507	1	New 3/2011
BTHWYN	PAPR, C420 Plus, Scott	3493	SA5508	1	New 3/2011
BTHWYN	PAPR, C420 Plus, Scott	3494	SA5515	1	New 3/2011
BTHWYN	PAPR, C420 Plus, Scott	3495	SA5517	1	New 3/2011
BTHWYN	PAPR, C420 Plus, Scott	3496	SA5523	1	New 3/2011
BTHWYN	PAPR, C420 Plus, Scott	3497	SA5514	1	New 3/2011
BTHWYN	PAPR, C420 Plus, Scott	3691	SA5513	1	New 3/2011
BTHWYN	Scott Cylinder 60 min 4500 psi	OP93067	A27969	4	Next Hydro Test Due: 6/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP93069	A41720	4	Next Hydro Test Due: 3/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP93075	A41706	4	Next Hydro Test Due: 4/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP93076	A27978 w/G. Ham	4	Next Hydro Test Due: 3/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP93078	A73016	4	Next Hydro Test Due: 3/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP93099	A41726 w/C. Santiago	4	Next Hydro Test Due: 3/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP93100	A41668	4	Next Hydro Test Due: 8/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP93105	A41696	4	Next Hydro Test Due: 3/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP93108	A73032	4	Next Hydro Test Due: 3/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP93129	A41676	4	Next Hydro Test Due: 8/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP93130	A41675	4	Next Hydro Test Due: 3/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP93140	A41669	4	Next Hydro Test Due: 8/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP93143	A41685 w/ C. Wagner	4	Next Hydro Test Due: 3/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP93145	A41686	4	Next Hydro Test Due: 3/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP93146	A41691	4	Next Hydro Test Due: 6/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP93151	A41727	4	H.T. Due 4/30/2014 @ COOP
BTHWYN	Scott Cylinder 60 min 4500 psi	OP93156	A41690	4	Next Hydro Test Due: 3/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP93157	A73007	4	To R. Fetzer 11/12/09
WV	Scott Cylinder 60 min 4500 psi	OP93159	A41705 (Lindsey)	4	Next Hydro Test Due: 3/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP93172	A27968	4	Next Hydro Test Due: 3/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP93183	A41730	4	Next Hydro Test Due: 3/2017

# Region III Equipment Condition Codes- Current as of 3/15/2013

Respiratory Equipment					
Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
WV	Scott Cylinder 60 min 4500 psi	OP931 85	A73010	4	Next Hydro Test Due: 3/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP93188	A41725	4	
BTHWYN	Scott Cylinder 60 min 4500 psi	OP94307	A73026	4	H.T. Due 8/2012 McLaughlin
BTHWYN	Scott Cylinder 60 min 4500 psi	OP94320	A73024	4	Next Hydro Test Due: 8/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP94358	A27975	4	H.T. Due 12/2014 @ COOP J. Kelly
BTHWYN	Scott Cylinder 60 min 4500 psi	OP94473	A27970	4	Next Hydro Test Due: 8/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP94507	A73002	1	Next Hydro Test Due: 9/2015
BTHWYN	Scott Cylinder 60 min 4500 psi	OP94510	A73025	4	Next Hydro Test Due: 4/2014 @ COOP
BTHWYN	Scott Cylinder 60 min 4500 psi	OP94511	A73027	4	Next Hydro Test Due: 8/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP94512	A73001	4	Next Hydro Test Due: 8/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP94513	A73028	4	Next Hydro Test Due: 8/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP94514	A27965	4	Next Hydro Test Due: 8/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP94519	A27974	4	Next Hydro Test Due: 8/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP94531	A27964	7	Next Hydro Test Due: 8/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP94545	A73019	4	H.T. Due 1/28/2014 @ COOP
BTHWYN	Scott Cylinder 60 min 4500 psi	OP94579	A73017	1	Next Hydro Test Due: 8/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP94912	A27981 Rovira	4	H.T. Due 12/2014 @ COOP
BTHWYN	Scott Cylinder 60 min 4500 psi	OP94922	940897	4	Next Hydro Test Due: 12/2014
BTHWYN	Scott Cylinder 60 min 4500 psi	OP94941	A41694	4	H.T. Due 1/28/2014 @ COOP
BTHWYN	Scott Cylinder 60 min 4500 psi	OP94949	A41684	4	Next Hydro Test Due: 1/2018
BTHWYN	Scott Cylinder 60 min 4500 psi	OP94964	A73085	4	Next Hydro Test Due: 8/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP94992	A27963 w/ L. Casillas	4	H.T. Due 3-6-13 OOS for flow test
BTHWYN	Scott Cylinder 60 min 4500 psi	OP95007	A41699	4	Next Hydro Test Due: 8/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP95034	A41672	4	Next Hydro Test Due: 8/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP95040	A41674	4	H.T. due April 2014
BTHWYN	Scott Cylinder 60 min 4500 psi	OP95042	A27973	4	Next Hydro Test Due: 1/2018
BTHWYN	Scott Cylinder 60 min 4500 psi	OP95057	A41704	4	Next Hydro Test Due: 8/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP95119	A41678	4	Next Hydro Test: 1/2018
BTHWYN	Scott Cylinder 60 min 4500 psi	OP95194	A41689	4	H.T. Due 12/2014 @ COOP



# Region III Equipment Condition Codes- Current as of 3/15/2013

Respiratory Equipment					
Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
BTHWYN	Scott Cylinder 60 min 4500 psi	OP95483	A41721	4	Next Hydro Test Due: 8/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP95479	A41701	4	Next Hydro Test Due: 8/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP95573	A41695 w/A27977	4	H.T. Due 3-6-13 OOS for Hydro Test
BTHWYN	Scott Cylinder 60 min 4500 psi	OP79508	A73004	4	Next Hydro Test Due: 3/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP93072	A73021	4	Next Hydro Test Due: 3/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP93073	A41716	4	Next Hydro Test Due: 3/2017
WV	Scott Cylinder 60 min 4500 psi	OP93087	A73030	4	Next Hydro Test Due: 3/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP93092	A41734 Ventura	4	Next Hydro Test Due: 3/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP93098	A41667	1	Next Hydro Test Due: 11/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP93111	A41717	4	Next Hydro Test Due: 11/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP93119	A41666	4	Next Hydro Test Due: 11/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP93 120	A73023 Jarvela	4	Next Hydro Test Due: 3/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP93127	A73022	1	Next Hydro Test Due: 11/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP93147	A41711	4	Next Hydro Test Due: 11/2017
WV	Scott Cylinder 60 min 4500 psi	OP93149	A41680	4	Next Hydro Test Due: 3/2017
WV	Scott Cylinder 60 min 4500 psi	OP93150	A41722	4	Next Hydro Test Due: 3/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP94316	A41739	4	Next Hydro Test Due: 11/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP94515	A41735	4	Next Hydro Test Due: 11/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP94532	A27958	4	Next Hydro Test Due: 11/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP94931	A41715	4	Next Hydro Test Due: 11/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP94951	A41738	4	Next Hydro Test Due: 11/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP95003	A41710	4	Next Hydro Test Due: 11/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP95050	A41712	4	Next Hydro Test Due: 11/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP79461	A41733	4	Next Hydro Test Due: 3/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP93118	A41681	4	Next Hydro Test Due: 8/2017
BTHWYN	Scott Cylinder 60 min 4500 psi	OP94586	A73018	4	Next Hydro Test Due: 1/2018
BTHWYN	Scott Cylinder 60 min 4500 psi	OP129131	940414	4	H.T. Due 2/2015
PHL H&S	TSI Particle Generator 8026	80260248	940793	4	
MD H&S	TSI Particle Generator 8026	11067	940896	4	
BTHWYN	TSI PortaCount Plus 8020A	15033	944003 Due for Exchange	4	Next Cal Due: 10/18/2013 Manuf. won't repair
WV H&S	TSI Porta Count Pro	8030101202	940962	4	Next Cal Due: 8/9/2013

## Region III Equipment Condition Codes- Current as of 3/15/2013

Respiratory Equipment					
Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
PHL H&S	TSI Porta Count Pro	80300093701	940963	4	Next Cal Due: 1/27/2013
MD H&S	TSI Porta Count Pro	8030101203	A27639	4	Next Cal Due: 6/22/2013
BTHWYN	TSI Porta Count Pro – Calibration Back-Up	8030124126		4	Next Cal Due: 10/12/2013
BTHWYN	TSI Porta Count Pro – Calibration Back-Up	8030124126		4	Next Cal Due: 10/12/2013

Surveying Equipment					
Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
BTHWYN	Garmin GPS V Personal Navigator GPS	73072850	940974	4	
BTHWYN	Garmin eTrex Personal Navigator GPS	73072851	73072851	4	
BTHWYN	Garmin eTrex Personal Navigator GPS	73072852	73072852	4	
BTHWYN	Garmin eTrex Personal Navigator GPS	73072854	73072854	4	
BTHWYN	Garmin eTrex Personal Navigator GPS	73072856	73072856	4	
BTHWYN	Garmin eTrex Personal Navigator GPS	73072857	73072857	4	
BTHWYN	Garmin eTrex Personal Navigator GPS	73072860	73072860	4	
BTHWYN	Garmin GPS V Personal Navigator GPS	73072861	940952	4	
BTHWYN	Garmin eTrex Personal Navigator GPS	73076242	73076242	4	
MD	Garmin eTrex Personal Navigator GPS	73072853	73072853	4	
BTHWYN	Garmin GPS NUVI 360	10368526	S37841	4	@ COOP from M. Aquino
BTHWYN	Garmin GPS Map 76	80415912	A41619	4	@ COOP from M. Aquino
WV	Lietz Model 115 Transit w/Tripod	658095	940546 Tripod 658096	4	CAL DUE 6/ 17/06 Calibrate???
WV	Trimble Ranger Handheld Pathfinder Pro XRS	SS59C17622 / 0224099642	940439 / 940440	4	w/ M. Easton
BTHWYN	Trimble Ranger Handheld/Pathfinder Pro XRS	SS59C17624 / 0224099644	940644 / 940645	4	@ COOP
WV	Trimble Recon Handheld/Pro XH Receiver Kit #2	CS42A26894 / 4737488682	940613 / 940612	4	w/ M. Easton
BTHWYN	Trimble Recon Handheld/Pro XH Receiver Kit #1	CS42A26893 / 4732480098	940614 / 940611	4	

Toxic & Specific Gas Meters					
Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
BTHWYN	Draeger CMS Analyser w/ Remote System	ARZF-0089	940791	1	New 12/2008
MD H&S	MSA Sirius Multi-Gas Detector	A3-4222- J06	S96007	4	Next Cal Due: 8/24/2013
MD H&S	MSA Sirius Multi-Gas Detector	A3-7288	S96009	4	Next Cal Due: 7/20/2013

# Region III Equipment Condition Codes- Current as of 3/15/2013

Toxic & Specific Gas Meters					
Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
MD H&S	Industrial Sci TMX 412 CL2/CLO2/LEL/O2	0108013-179	0108013-179	4	
MD H&S	Industrial Sci TMX 412 LEL/NO2/O2/SO2	0108013-113	0108013-113	4	
BTHWYN	MDA Scientific Single Point Monitor	5077	940355	4	
BTHWYN	MDA Scientific Single Point Monitor	5076	940356	4	
BTHWYN	MSA SAFECOM Command Center	J06-2241288-10-001	940563	4	
BTHWYN	MSA SAFEMTX CW/CI2/HCN/H2S/NH3/VOC/RAD	J06-2241288-20-001	940559	4	OOS for repair
BTHWYN	MSA SAFEMTX CW/CI2/HCN/H2S/NH3/VOC/RAD	J06-2241288-20-002	940560	4	
BTHWYN	MSA SAFEMTX CW/CI2/HCN/H2S/NH3/VOC/RAD	J06-2241288-20-003	940561	4	
BTHWYN	MSA SAFEMTX CW/CI2/HCN/H2S/NH3/VOC/RAD	J06-2241288-20-004	940562	4	OOS for repair
BTHWYN	Thermo TVA 1000B FID/PID	0330202743	940313	4	
BTHWYN	RAE Systems PPB RAE 3000 Model PGM-7340	594-900951	940836	4	
BTHWYN	RAE Systems MultiRAE Plus CO/H2S/LEL/O2	095-510848	A73061 G.Ham	4	Sensor Expirations - LEL, O <sub>2</sub> & H <sub>2</sub> S – 10/2014, CO – 9/2014
BTHWYN	RAE Systems MultiRAE Plus CO/H2S/LEL/O2	095-510852	A73063 D. McLaughlin	4	Sensor Expirations - LEL – 10/2014, O <sub>2</sub> – 1/2015, H <sub>2</sub> S – 12/2014 CO – 9/2014
BTHWYN	RAE Systems MultiRAE Plus CO/H2S/LEL/O2	095-511308	940275 T. Richardson	4	CO, H2S, O2, and LEL Exp 10/12
BTHWYN	RAE Systems MultiRAE Plus CO/H2S/LEL/O2	095-511686	940270 @ COOP	4	Sensor Expirations - LEL & O <sub>2</sub> – 8/2014, H <sub>2</sub> S – 4/2014 CO – 7/2014
BTHWYN	RAE Systems MultiRAE Plus CO/H2S/LEL/O2	095-511696	940272 @ COOP	4	CO, H2S, CO, and O2 EXP 10/12
BTHWYN	RAE Systems MultiRAE Plus CO/H2S/LEL/O2	095-511721	940269 Bartos	4	Sensor Expirations - CO, H2S, & LEL – 7/ 2014 O <sub>2</sub> - 8/2014
BTHWYN	RAE Systems MultiRAE Plus CO/H2S/LEL/O2	095-512340	940273 R. Scharr	4	Sensor Expirations - LEL – 8/2014, O <sub>2</sub> – 12/2014, H <sub>2</sub> S – 10/2014 CO – 8/2014
BTHWYN	RAE Systems MultiRAE Plus CO/H2S/LEL/O2	503608	A41004 M. Towle	4	Sensor Expirations - LEL – 9/2014, O <sub>2</sub> – 12/2014, H <sub>2</sub> S – 10/2014 CO – 8/2014
BTHWYN	RAE Systems MultiRAE Plus CO/H2S/LEL/O2	503610	A41008 R Fetzer	4	Exp. CO 5/12, LEL 1/12, H2S 4/12, O2 6/12
BTHWYN	RAE Systems MultiRAE Plus CO/H2S/LEL/O2	503612	A41006 R. Kelly	4	Exp CO/O2 – 2/13, LEL – 2/13, H2S – 12/12
BTHWYN	RAE Systems MultiRAE Plus CO/H2S/LEL/O2	503613	A41005 C. Wagner	4	Exp CO 5/12,LEL 4/12, H2S 3/12, O2 7/13
BTHWYN	RAE Systems MultiRAE Plus CO/H2S/LEL/O2	095-503616	A41007 Unassigned	4	Sensor Expirations - LEL – 8/2014, O <sub>2</sub> – 11/2015, H <sub>2</sub> S – 12/2014 CO – 8/2014
BTHWYN	RAE Systems MultiRAE Plus CO/H2S/LEL/O2	503744 v. 3.21C	A41009 Unassigned	4	Sensor Expirations - LEL – 6/2014, O <sub>2</sub> – 2/2015, H <sub>2</sub> S – 12/2014 CO – 6/2014
WV	RAE Systems MultiRAE Plus CO/H2S/LEL/O2	095-517070	940516 M. Linden	4	Sensor Expirations - CO/LEL-9/14, H <sub>2</sub> S - 3/14, O2- 10/14
BTHWYN	RAE Systems MultiRAE Plus CO/H2S/LEL/O2	095-517075	940515 J.Kelly	4	Exp. CO, H2S - 01/13, LEL, O2 2/13
BTHWYN	RAE Systems MultiRAE Plus CO/H2S/LEL/O2	095-517087	940517 L. Casillas	4	O2 9/12, LEL 10/12, CO 10/12, H2S 11/12
BTHWYN	RAE Systems MultiRAE Plus CO/H2S/LEL/O2	095-108214	940979 Ventura	4	Sensor Expirations - LEL – 9/2014, O <sub>2</sub> – 10/2014, H <sub>2</sub> S – 8/2014 CO – 5/2014
PHL H&S	RAE Systems MultiRAE Plus CO/H2S/LEL/O2	095-511407	940599	4	Sensor Expirations - LEL & O <sub>2</sub> – 8/2014, H <sub>2</sub> S – 4/2014 CO – 7/2014
WV	RAE Systems MultiRAE Plus CO/H2S/LEL/O2	095-511687	940266 Lindsey	4	Sensor Expirations - CO, H2S & LEL - 7/2014, O2 8/2014
WV	RAE Systems MultiRAE Plus CO/H2S/LEL/O2	095-512337	940268 Matlock	4	Sensor Expirations - CO - 7/2014, H <sub>2</sub> S – 6/2014 O <sub>2</sub> & LEL – 8/2014

## Region III Equipment Condition Codes- Current as of 3/15/2013

Toxic & Specific Gas Meters					
Owning Location	Item Description	Serial Number	EPA TAG NUMBER	Condition Code	Notes
WV	RAE Systems MultiRAE Plus CO/H2S/LEL/O2	095-512338	940274 Floater	4	Sensor Expirations - CO & LEL – 7/2014, H <sub>2</sub> S – 4/2014 O <sub>2</sub> – 8/2014
BTHWYN	RAE Systems MultiRAE Plus CO/H2S/LEL/O2	095-503607	A41003 Unassigned	4	Sensor Expirations - CO - 7/2014, H <sub>2</sub> S – 6/2014 O <sub>2</sub> & LEL – 8/2014
MD	RAE Systems MultiRAE Plus CO/H2S/LEL/O2	095-511708	940267 Fitzsimmons	4	Sensor Expirations – CO & H <sub>2</sub> S, – 7/ 2014 LEL & O <sub>2</sub> - 8/2014
WV H&S	MSA Orion Multi-Gas Sensor	A5-6366KO1		4	Factory Calibration 2/14/2012

Water Quality & Sampling					
Owning Location	Item Description	Serial Number	EPA Tag Number	Condition Code	Notes
BTHWYN	Hach 2100P Turbidity Meter	04070036492	940804	4	
WV	Hach 2100P Turbidity Meter	40700036510	940553	4	
MD	Hach 2100P Turbidity Meter	40700036501	940721	4	
BTHWYN	Hach Digital Titrator	A4188	940802	4	
WV	Hach Digital Titrator	A4190-1	940554	4	
MD	Hach Digital Titrator	A4190-2	940720	4	
BTHWYN	Hach DR 2400 Portable Spectrophotometer	040300003845	940619	4	Next Certification Due: 3/23/2013 OOS for cal
MD	Hach DR 2400 Portable Spectrophotometer	040200003638	940672	4	Next Certification Due: 3/23/2013 OOS for cal
BTHWYN	Hach DR 2700 Portable Spectrophotometer	1425353	S38137	1	New 2/2012 OOS for cal
WV	Hach DR 2700 Portable Spectrophotometer	1427837	S38186	1	New 2/2012
BTHWYN	Hach DR/890 Colorimeter	120190C87870	S38185	1	New 2/2012
BTHWYN	Hach Reagent Kit	940503	940503	4	
WV	Hach Reagent Kit	940504	940504	4	
MD	Hach Reagent Kit	940505	940505	4	
BTHWYN	LaMotte Smart 2 Colorimeter	02069-2704	940629	4	
WV	LaMotte Smart 2 Colorimeter	02070	940427	4	
MD	LaMotte Smart 2 Colorimeter	2048-2604	940669	4	
WV	MMC Oil Water Interface Probe (100')	1256	940433	5	
BTHWYN	Severn Trent Eclox Water Test Kit	6520202	940695	4	Reagents Exp 10/3/12
WV	Severn Trent Eclox Water Test Kit	6540202	940696	4	Reagents Exp 10/3/12
MD	Severn Trent Eclox Water Test Kit	6590202	940687	4	Reagents Exp 10/3/12
BTHWYN	YSI Professional Plus pH/Temp/Do/Cond/ORP	11L100604	S38161	S	New 11/2011. 10 m probe.
BTHWYN	YSI 556 MPS pH/Temp/Do/Cond 4 Meter probe	04G13008AB	940592	4	Replaced pH/ORP Probe 11/12/10

## Region III Equipment Condition Codes- Current as of 3/15/2013

Water Quality & Sampling					
Owning Location	Item Description	Serial Number	EPA Tag Number	Condition Code	Notes
WV	YSI 556 MPS pH/Temp/Do/Cond (5564 pH only)	04G13008AA	940556	4	
MD	YSI 556 MPS pH/Temp/Do/Cond (5564 pH only)	04G12881AN	940671	4	OOS for sensor replacement/repair

Removed From Service					
	Item Description	Serial Number	EPA TAG NUMBER	C.C.	Notes
BTHWYN	RAE Systems MultiRAE Plus CO/H2S/LEL/O2	095-503749 OOS	A41012 Unassigned	S	Unit OOS - Unrepairable
BTHWYN	RAE Systems MultiRAE Plus CO/H2S/LEL/O2	095-503609 V. 3.21C	A41002 S. JARVELA	S	Unit non-repairable
BTHWYN	ISI Emergency Escape Pack	36854	940596 @ COOP	S	Cylinder condemned during Hydro Test
BTHWYN	Draeger Sensor Bias Box	ERTK-0007	940353	S	Obsolete
BTHWYN	Draeger Sensor Bias Box	ERTK-0009	940354	S	Obsolete